

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112123\
 Data File : PO099976.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Nov 2023 10:06
 Operator : YP/AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 10:54:56 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 10:47:35 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.479	3.705	127.7E6	110.1E6	72.499	72.624
2) SA Decachlor...	10.359	8.801	53625522	82806939	74.543	74.321
Target Compounds						
16) L4 AR-1242-1	5.666	4.811	24806394	17821190	725.571	726.404
17) L4 AR-1242-2	5.689	4.831	38296455	24991409	732.978	731.727
18) L4 AR-1242-3	5.752	5.010	25660627	13636189	735.864	735.036
19) L4 AR-1242-4	5.852	5.094	19402789	14376219	733.990	737.494
20) L4 AR-1242-5	6.599	5.624	16501921	16836480	736.125	730.490



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112123\
Data File : PO099976.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Nov 2023 10:06
Operator : YP/AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 10:54:56 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 10:47:35 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.479 min
Delta R.T.: 0.000 min
Response: 127692855
Conc: 72.50 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.705 min
Delta R.T.: 0.000 min
Response: 110119760
Conc: 72.62 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.359 min
Delta R.T.: 0.000 min
Response: 53625522
Conc: 74.54 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.801 min
Delta R.T.: 0.000 min
Response: 82806939
Conc: 74.32 ng/ml



#16 AR-1242-1


R.T.: 5.666 min
Delta R.T.: 0.000 min
Response: 24806394
Conc: 725.57 ng/ml

Instrument :
ECD_O
ClientSampleId :




#16 AR-1242-1

R.T.: 4.811 min
Delta R.T.: 0.000 min
Response: 17821190
Conc: 726.40 ng/ml




#17 AR-1242-2

R.T.: 5.689 min
Delta R.T.: 0.000 min
Response: 38296455
Conc: 732.98 ng/ml



#17 AR-1242-2


R.T.: 4.831 min
Delta R.T.: 0.000 min
Response: 24991409
Conc: 731.73 ng/ml



#18 AR-1242-3


R.T.: 5.752 min
Delta R.T.: 0.000 min
Response: 25660627
Conc: 735.86 ng/ml

Instrument :
ECD_O
ClientSampleId :




#18 AR-1242-3

R.T.: 5.010 min
Delta R.T.: 0.000 min
Response: 13636189
Conc: 735.04 ng/ml




#19 AR-1242-4

R.T.: 5.852 min
Delta R.T.: 0.000 min
Response: 19402789
Conc: 733.99 ng/ml



#19 AR-1242-4


R.T.: 5.094 min
Delta R.T.: 0.000 min
Response: 14376219
Conc: 737.49 ng/ml



#20 AR-1242-5

R.T.: 6.599 min
Delta R.T.: 0.000 min
Response: 16501921
Conc: 736.13 ng/ml

Instrument :
ECD_O
ClientSampleId :



#20 AR-1242-5

R.T.: 5.624 min
Delta R.T.: 0.000 min
Response: 16836480
Conc: 730.49 ng/ml