

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0112123\
 Data File : P0099978.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Nov 2023 10:39
 Operator : YP/AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 10:58:23 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0112123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 10:58:09 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.480	3.706	44434268	38382540	25.171	25.234
2) SA Decachlor...	10.357	8.801	19351294	29448952	26.398	26.058
Target Compounds						
16) L4 AR-1242-1	5.667	4.812	8755606	6736932	254.544	268.008
17) L4 AR-1242-2	5.689	4.831	13359834	9147411	254.252	263.137
18) L4 AR-1242-3	5.753	5.010	8836623	4883862	252.546	259.812
19) L4 AR-1242-4	5.852	5.095	6720835	5231489	253.169	263.531
20) L4 AR-1242-5	6.599	5.624	5581396	6016884	249.232	258.202



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112123\
Data File : PO099978.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Nov 2023 10:39
Operator : YP/AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 10:58:23 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 10:58:09 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.480 min
Delta R.T.: 0.000 min
Response: 44434268
Conc: 25.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.706 min
Delta R.T.: 0.000 min
Response: 38382540
Conc: 25.23 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.357 min
Delta R.T.: 0.000 min
Response: 19351294
Conc: 26.40 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.801 min
Delta R.T.: 0.000 min
Response: 29448952
Conc: 26.06 ng/ml



#16 AR-1242-1


R.T.: 5.667 min
Delta R.T.: 0.000 min
Response: 8755606
Conc: 254.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250




#16 AR-1242-1

R.T.: 4.812 min
Delta R.T.: 0.000 min
Response: 6736932
Conc: 268.01 ng/ml




#17 AR-1242-2

R.T.: 5.689 min
Delta R.T.: 0.000 min
Response: 13359834
Conc: 254.25 ng/ml



#17 AR-1242-2


R.T.: 4.831 min
Delta R.T.: 0.000 min
Response: 9147411
Conc: 263.14 ng/ml



#18 AR-1242-3


R.T.: 5.753 min
Delta R.T.: 0.000 min
Response: 8836623
Conc: 252.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250




#18 AR-1242-3

R.T.: 5.010 min
Delta R.T.: 0.000 min
Response: 4883862
Conc: 259.81 ng/ml




#19 AR-1242-4

R.T.: 5.852 min
Delta R.T.: 0.000 min
Response: 6720835
Conc: 253.17 ng/ml



#19 AR-1242-4


R.T.: 5.095 min
Delta R.T.: 0.000 min
Response: 5231489
Conc: 263.53 ng/ml



#20 AR-1242-5

R.T.: 6.599 min
Delta R.T.: 0.000 min
Response: 5581396
Conc: 249.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250



#20 AR-1242-5

R.T.: 5.624 min
Delta R.T.: 0.000 min
Response: 6016884
Conc: 258.20 ng/ml