

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112123\
 Data File : PO099989.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Nov 2023 13:45
 Operator : YP/AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 14:03:37 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 14:03:25 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.479	3.704	8448985	8020149	4.562	4.918
2) SA Decachlor...	10.357	8.800	3006748	5063533	4.159	4.473
Target Compounds						
26) L6 AR-1254-1	6.534	5.625	2304284	2554490	47.853	49.279
27) L6 AR-1254-2	6.755	5.774	3255940	2218020	47.476	47.630
28) L6 AR-1254-3	7.125	6.182	2962396	3677207	46.500	47.149
29) L6 AR-1254-4	7.414	6.413	1706386	1997352	45.138	46.802
30) L6 AR-1254-5	7.839	6.835	1849401	3354444	43.208	46.356



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112123\
Data File : PO099989.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Nov 2023 13:45
Operator : YP/AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 14:03:37 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 14:03:25 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.479 min
Delta R.T.: 0.000 min
Response: 8448985
Conc: 4.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.704 min
Delta R.T.: 0.000 min
Response: 8020149
Conc: 4.92 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.357 min
Delta R.T.: 0.000 min
Response: 3006748
Conc: 4.16 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.800 min
Delta R.T.: 0.000 min
Response: 5063533
Conc: 4.47 ng/ml



#26 AR-1254-1


R.T.: 6.534 min
Delta R.T.: 0.000 min
Response: 2304284
Conc: 47.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050




#26 AR-1254-1

R.T.: 5.625 min
Delta R.T.: 0.000 min
Response: 2554490
Conc: 49.28 ng/ml




#27 AR-1254-2

R.T.: 6.755 min
Delta R.T.: 0.000 min
Response: 3255940
Conc: 47.48 ng/ml



#27 AR-1254-2


R.T.: 5.774 min
Delta R.T.: 0.000 min
Response: 2218020
Conc: 47.63 ng/ml



#28 AR-1254-3


R.T.: 7.125 min
Delta R.T.: 0.000 min
Response: 2962396
Conc: 46.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050




#28 AR-1254-3

R.T.: 6.182 min
Delta R.T.: 0.000 min
Response: 3677207
Conc: 47.15 ng/ml




#29 AR-1254-4

R.T.: 7.414 min
Delta R.T.: 0.000 min
Response: 1706386
Conc: 45.14 ng/ml



#29 AR-1254-4


R.T.: 6.413 min
Delta R.T.: 0.000 min
Response: 1997352
Conc: 46.80 ng/ml



#30 AR-1254-5

R.T.: 7.839 min
Delta R.T.: 0.000 min
Response: 1849401
Conc: 43.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050



#30 AR-1254-5

R.T.: 6.835 min
Delta R.T.: 0.000 min
Response: 3354444
Conc: 46.36 ng/ml