

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112123\
 Data File : PO099994.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Nov 2023 15:09
 Operator : YP/AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/22/2023
 Supervised By :Ankita Jodhani 11/22/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 15:46:12 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 15:35:12 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.480	3.705	49852810	43705436	25.941	26.277
2) SA Decachlor...	10.360	8.802	34557146	52025019	25.675	25.527
Target Compounds						
41) L9 AR-1268-1	8.795	7.664	29046661	42711748	256.713m	253.074
42) L9 AR-1268-2	8.891	7.729	25464935	39201396	257.957	251.656
43) L9 AR-1268-3	9.129	7.939	23935832	37162888	258.300	255.069
44) L9 AR-1268-4	9.571	8.231	7302032	11975116	244.667	248.958
45) L9 AR-1268-5	10.005	8.534	65594452	101.3E6	254.288	247.276

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112123\
Data File : PO099994.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Nov 2023 15:09
Operator : YP/AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1268ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/22/2023
Supervised By :Ankita Jodhani 11/22/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 15:46:12 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 15:35:12 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.480 min
Delta R.T.: 0.000 min
Response: 49852810
Conc: 25.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/22/2023
Supervised By :Ankita Jodhani 11/22/2023




#1 Tetrachloro-m-xylene

R.T.: 3.705 min
Delta R.T.: 0.000 min
Response: 43705436
Conc: 26.28 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.360 min
Delta R.T.: 0.000 min
Response: 34557146
Conc: 25.68 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.802 min
Delta R.T.: 0.000 min
Response: 52025019
Conc: 25.53 ng/ml




#41 AR-1268-1

R.T.: 8.795 min
Delta R.T.: 0.000 min
Response: 29046661
Conc: 256.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/22/2023
Supervised By :Ankita Jodhani 11/22/2023




#41 AR-1268-1

R.T.: 7.664 min
Delta R.T.: 0.000 min
Response: 42711748
Conc: 253.07 ng/ml




#42 AR-1268-2

R.T.: 8.891 min
Delta R.T.: 0.000 min
Response: 25464935
Conc: 257.96 ng/ml



#42 AR-1268-2

R.T.: 7.729 min
Delta R.T.: 0.000 min
Response: 39201396
Conc: 251.66 ng/ml




#43 AR-1268-3

R.T.: 9.129 min
Delta R.T.: 0.000 min
Response: 23935832
Conc: 258.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/22/2023
Supervised By :Ankita Jodhani 11/22/2023




#43 AR-1268-3

R.T.: 7.939 min
Delta R.T.: 0.000 min
Response: 37162888
Conc: 255.07 ng/ml




#44 AR-1268-4

R.T.: 9.571 min
Delta R.T.: 0.000 min
Response: 7302032
Conc: 244.67 ng/ml



#44 AR-1268-4

R.T.: 8.231 min
Delta R.T.: 0.000 min
Response: 11975116
Conc: 248.96 ng/ml




#45 AR-1268-5

R.T.: 10.005 min
Delta R.T.: 0.000 min
Response: 65594452
Conc: 254.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/22/2023
Supervised By :Ankita Jodhani 11/22/2023



#45 AR-1268-5

R.T.: 8.534 min
Delta R.T.: 0.000 min
Response: 101323737
Conc: 247.28 ng/ml