

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112123\
 Data File : PO099996.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Nov 2023 15:43
 Operator : YP/AJ
 Sample : PO112123ICV500
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO112123

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 16:00:14 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 15:58:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.480	3.705	95284447	82754946	55.825	55.174
2) SA Decachlor...	10.358	8.800	38416643	57867138	52.890	51.713
Target Compounds						
3) L1 AR-1016-1	5.667	4.812	22747641	16591266	528.691	525.435
4) L1 AR-1016-2	5.690	4.831	35246487	22962049	544.413	531.059
5) L1 AR-1016-3	5.753	5.011	23524908	12515619	539.735	539.675
6) L1 AR-1016-4	5.852	5.052	17752813	10784245	537.810	526.030
7) L1 AR-1016-5	6.152	5.269	17271901	13431251	550.088	526.670
31) L7 AR-1260-1	7.292	6.314	28165475	30091613	528.985	520.758
32) L7 AR-1260-2	7.552	6.502	28065208	34839575	530.561	524.560
33) L7 AR-1260-3	7.916	6.659	19214345	34319267	538.669	536.343
34) L7 AR-1260-4	8.145	7.135	21937328	26721747	530.039	529.916
35) L7 AR-1260-5	8.472	7.377	35112220	59038803	525.634	526.648



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112123\
Data File : PO099996.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Nov 2023 15:43
Operator : YP/AJ
Sample : PO112123ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO112123

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 16:00:14 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 15:58:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.480 min
Delta R.T.: 0.000 min
Response: 95284447
Conc: 55.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO112123




#1 Tetrachloro-m-xylene

R.T.: 3.705 min
Delta R.T.: 0.000 min
Response: 82754946
Conc: 55.17 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.358 min
Delta R.T.: 0.001 min
Response: 38416643
Conc: 52.89 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.800 min
Delta R.T.: -0.001 min
Response: 57867138
Conc: 51.71 ng/ml



#3 AR-1016-1


R.T.: 5.667 min
Delta R.T.: 0.000 min
Response: 22747641
Conc: 528.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO112123




#3 AR-1016-1

R.T.: 4.812 min
Delta R.T.: 0.000 min
Response: 16591266
Conc: 525.44 ng/ml




#4 AR-1016-2

R.T.: 5.690 min
Delta R.T.: 0.000 min
Response: 35246487
Conc: 544.41 ng/ml



#4 AR-1016-2


R.T.: 4.831 min
Delta R.T.: 0.000 min
Response: 22962049
Conc: 531.06 ng/ml



#5 AR-1016-3


R.T.: 5.753 min
Delta R.T.: 0.000 min
Response: 23524908
Conc: 539.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO112123




#5 AR-1016-3

R.T.: 5.011 min
Delta R.T.: 0.000 min
Response: 12515619
Conc: 539.67 ng/ml




#6 AR-1016-4

R.T.: 5.852 min
Delta R.T.: 0.000 min
Response: 17752813
Conc: 537.81 ng/ml



#6 AR-1016-4


R.T.: 5.052 min
Delta R.T.: 0.000 min
Response: 10784245
Conc: 526.03 ng/ml



#7 AR-1016-5


R.T.: 6.152 min
Delta R.T.: 0.000 min
Response: 17271901
Conc: 550.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO112123




#7 AR-1016-5

R.T.: 5.269 min
Delta R.T.: 0.000 min
Response: 13431251
Conc: 526.67 ng/ml




#31 AR-1260-1

R.T.: 7.292 min
Delta R.T.: 0.000 min
Response: 28165475
Conc: 528.98 ng/ml



#31 AR-1260-1


R.T.: 6.314 min
Delta R.T.: 0.000 min
Response: 30091613
Conc: 520.76 ng/ml



#32 AR-1260-2


R.T.: 7.552 min
Delta R.T.: 0.000 min
Response: 28065208
Conc: 530.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO112123




#32 AR-1260-2

R.T.: 6.502 min
Delta R.T.: 0.000 min
Response: 34839575
Conc: 524.56 ng/ml




#33 AR-1260-3

R.T.: 7.916 min
Delta R.T.: 0.000 min
Response: 19214345
Conc: 538.67 ng/ml



#33 AR-1260-3


R.T.: 6.659 min
Delta R.T.: 0.000 min
Response: 34319267
Conc: 536.34 ng/ml



#34 AR-1260-4


R.T.: 8.145 min
Delta R.T.: 0.001 min
Response: 21937328
Conc: 530.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO112123




#34 AR-1260-4

R.T.: 7.135 min
Delta R.T.: 0.000 min
Response: 26721747
Conc: 529.92 ng/ml



#35 AR-1260-5

R.T.: 8.472 min
Delta R.T.: 0.000 min
Response: 35112220
Conc: 525.63 ng/ml



#35 AR-1260-5

R.T.: 7.377 min
Delta R.T.: -0.001 min
Response: 59038803
Conc: 526.65 ng/ml