

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112123\
 Data File : PO099997.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Nov 2023 16:00
 Operator : YP/AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO112123AR1242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 16:12:44 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 15:58:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.480	3.705	96608812	83665492	56.601	55.781
2)	SA Decachlor...	10.360	8.802	37977645	57199823	52.286	51.117
Target Compounds							
16)	L4 AR-1242-1	5.667	4.812	18542731	13765979	535.252	523.799
17)	L4 AR-1242-2	5.690	4.831	28759733	18740650	548.056	532.975
18)	L4 AR-1242-3	5.753	5.010	18974973	10245529	545.193	535.014
19)	L4 AR-1242-4	5.852	5.095	14317833	10813811	544.603	536.876
20)	L4 AR-1242-5	6.599	5.625	12046452	12539772	538.776	536.672



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112123\
Data File : PO099997.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Nov 2023 16:00
Operator : YP/AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO112123AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 16:12:44 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 15:58:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.480 min
Delta R.T.: 0.000 min
Response: 96608812
Conc: 56.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO112123AR1242




#1 Tetrachloro-m-xylene

R.T.: 3.705 min
Delta R.T.: 0.000 min
Response: 83665492
Conc: 55.78 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.360 min
Delta R.T.: 0.003 min
Response: 37977645
Conc: 52.29 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.802 min
Delta R.T.: 0.000 min
Response: 57199823
Conc: 51.12 ng/ml



#16 AR-1242-1


R.T.: 5.667 min
Delta R.T.: 0.000 min
Response: 18542731
Conc: 535.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO112123AR1242




#16 AR-1242-1

R.T.: 4.812 min
Delta R.T.: 0.000 min
Response: 13765979
Conc: 523.80 ng/ml




#17 AR-1242-2

R.T.: 5.690 min
Delta R.T.: 0.000 min
Response: 28759733
Conc: 548.06 ng/ml



#17 AR-1242-2


R.T.: 4.831 min
Delta R.T.: 0.000 min
Response: 18740650
Conc: 532.98 ng/ml



#18 AR-1242-3


R.T.: 5.753 min
Delta R.T.: 0.000 min
Response: 18974973
Conc: 545.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO112123AR1242




#18 AR-1242-3

R.T.: 5.010 min
Delta R.T.: 0.000 min
Response: 10245529
Conc: 535.01 ng/ml




#19 AR-1242-4

R.T.: 5.852 min
Delta R.T.: 0.000 min
Response: 14317833
Conc: 544.60 ng/ml



#19 AR-1242-4


R.T.: 5.095 min
Delta R.T.: 0.000 min
Response: 10813811
Conc: 536.88 ng/ml



#20 AR-1242-5

R.T.: 6.599 min
Delta R.T.: 0.000 min
Response: 12046452
Conc: 538.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO112123AR1242



#20 AR-1242-5

R.T.: 5.625 min
Delta R.T.: 0.000 min
Response: 12539772
Conc: 536.67 ng/ml