

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112125\
 Data File : PO115334.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Nov 2025 17:36
 Operator : YP/AJ
 Sample : Q3693-13
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 C0AL6

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 25 10:46:23 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 06:00:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.341	3.644	164.7E6	104.7E6	23.403	23.434
2)	SA Decachlor...	9.919	8.630	62433619	55437546	14.124	14.201
Target Compounds							
31)	L7 AR-1260-1	7.063	6.197	15157575	16057334	49.594	54.228
32)	L7 AR-1260-2	7.317	6.385	28732041	29773198	72.578	69.235
33)	L7 AR-1260-3	7.674	6.537	20946951	17022617	63.128	48.477
34)	L7 AR-1260-4	7.893	7.007	27097905	14983670	88.840	49.446 #
35)	L7 AR-1260-5	8.202	7.249	40642515	46460286	56.985	56.547



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO112125\
Data File : PO115334.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Nov 2025 17:36
Operator : YP/AJ
Sample : Q3693-13
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
COAL6

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 25 10:46:23 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.341 min
Delta R.T.: -0.007 min
Response: 164710560
Conc: 23.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
C0AL6




#1 Tetrachloro-m-xylene

R.T.: 3.644 min
Delta R.T.: -0.004 min
Response: 104673970
Conc: 23.43 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.919 min
Delta R.T.: -0.011 min
Response: 62433619
Conc: 14.12 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.630 min
Delta R.T.: -0.011 min
Response: 55437546
Conc: 14.20 ng/ml



#31 AR-1260-1


R.T.: 7.063 min
Delta R.T.: -0.007 min
Response: 15157575
Conc: 49.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
C0AL6




#31 AR-1260-1

R.T.: 6.197 min
Delta R.T.: -0.009 min
Response: 16057334
Conc: 54.23 ng/ml




#32 AR-1260-2

R.T.: 7.317 min
Delta R.T.: -0.008 min
Response: 28732041
Conc: 72.58 ng/ml



#32 AR-1260-2


R.T.: 6.385 min
Delta R.T.: -0.009 min
Response: 29773198
Conc: 69.23 ng/ml



#33 AR-1260-3


R.T.: 7.674 min
Delta R.T.: -0.008 min
Response: 20946951
Conc: 63.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
C0AL6




#33 AR-1260-3

R.T.: 6.537 min
Delta R.T.: -0.009 min
Response: 17022617
Conc: 48.48 ng/ml




#34 AR-1260-4

R.T.: 7.893 min
Delta R.T.: -0.010 min
Response: 27097905
Conc: 88.84 ng/ml



#34 AR-1260-4


R.T.: 7.007 min
Delta R.T.: -0.010 min
Response: 14983670
Conc: 49.45 ng/ml



#35 AR-1260-5

R.T.: 8.202 min
Delta R.T.: -0.007 min
Response: 40642515
Conc: 56.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
C0AL6



#35 AR-1260-5

R.T.: 7.249 min
Delta R.T.: -0.010 min
Response: 46460286
Conc: 56.55 ng/ml