

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO112219\
 Data File : PO064016.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Nov 2019 15:03
 Operator : SM/AJ
 Sample : K5966-07
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 PCB-110719-07

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 22 15:52:11 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO111419.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 15 01:29:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.315	3.566	903788	640298	29.560	28.828
2)	SA Decachlor...	9.942	8.535	864373	691326	21.272	23.638
Target Compounds							
3)	L1 AR-1016-1	5.475	4.643	93402	138792	67.253	135.437 #
4)	L1 AR-1016-2	5.519	4.686	185005	128296	94.258	94.165
5)	L1 AR-1016-3	5.519	4.838	185005	139178	148.102	186.910 #
6)	L1 AR-1016-4	5.694	4.919	603735	153803	586.997	246.871 #
7)	L1 AR-1016-5	5.949	5.087	555350	34445	521.905	43.369 #
8)	L2 AR-1221-1	4.512	3.808	39826	100267	98.839	369.079 #
9)	L2 AR-1221-2	4.619	3.870	128375	123906	409.285	589.278 #
10)	L2 AR-1221-3	4.700	3.981	160324	265510	169.035	398.305 #
11)	L3 AR-1232-1	4.700	3.981	160324	265510	120.537	281.670 #
12)	L3 AR-1232-2	5.213	4.686	204494	128296	286.760	127.318 #
13)	L3 AR-1232-3	5.519	4.838	185005	139178	129.304	262.950 #
14)	L3 AR-1232-4	5.694	4.919	603735	153803	813.258	327.590 #
15)	L3 AR-1232-5	5.749	5.087	552818	34445	943.075	66.819 #
16)	L4 AR-1242-1	5.475	4.643	93402	138792	88.280	175.947 #
17)	L4 AR-1242-2	5.519	4.686	185005	128296	123.342	121.192
18)	L4 AR-1242-3	5.519	4.838	185005	139178	193.591	242.922 #
19)	L4 AR-1242-4	5.694	4.919	603735	153803	768.719	269.381 #
20)	L4 AR-1242-5	6.392	5.439	411355	152670	395.859	196.146 #
21)	L5 AR-1248-1	5.475	4.643	93402	138792	106.949	215.384 #
22)	L5 AR-1248-2	5.749	4.919	552818	153803	436.303	178.449 #
23)	L5 AR-1248-3	5.949	4.919	555350	153803	389.296	173.838 #
24)	L5 AR-1248-4	6.352	5.087	224578	34445	129.747	31.996 #
25)	L5 AR-1248-5	6.392	5.481	411355	43814	243.889	39.283 #
26)	L6 AR-1254-1	6.325	5.439	368208	152670	210.762	86.168 #
27)	L6 AR-1254-2	6.544	5.586	427847	160273	153.476	101.647 #
28)	L6 AR-1254-3	6.911	5.988	417411	229625	132.958	85.637 #
29)	L6 AR-1254-4	7.194	6.215	226611	135630	91.648	76.483
30)	L6 AR-1254-5	7.611	6.632	211614	277837	78.069	111.191 #
31)	L7 AR-1260-1	7.071	6.117	192331	112630	95.303	71.720
32)	L7 AR-1260-2	7.323	6.304	308991	143345	122.646	72.870 #
33)	L7 AR-1260-3	7.683	6.457	39030	97500	19.263	54.668 #
34)	L7 AR-1260-4	7.900	6.934	89566	80279	37.610	50.988 #
35)	L7 AR-1260-5	8.225	7.168	121659	93333	25.916	24.074
36)	L8 AR-1262-1	7.683	6.701	39030	36641	12.123	35.046 #
37)	L8 AR-1262-2	8.225	7.168	121659	93333	20.650	19.677
38)	L8 AR-1262-3	8.534	7.453	76344	74257	18.632	39.620 #
39)	L8 AR-1262-4	8.610	7.535	28124	626243	8.993	177.384 #
40)	L8 AR-1262-5	9.234	8.007	281489	18353	129.220	12.020 #
41)	L9 AR-1268-1	8.534	7.453	76344	74257	11.294	13.910

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0112219\
 Data File : P0064016.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Nov 2019 15:03
 Operator : SM/AJ
 Sample : K5966-07
 Misc :
 ALS Vial : 22 Sample Multiplier: 1


Instrument :
 ECD_O
 ClientSampleId :
 PCB-110719-07

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 22 15:52:11 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0111419.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 15 01:29:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.610	7.535	28124	626243	4.482	125.665 #
43) L9	AR-1268-3	8.821	7.771	58479	116482	11.265	29.344 #
44) L9	AR-1268-4	9.234	8.007	281489	18353	115.801	10.835 #
45) L9	AR-1268-5	9.621	0.000	26264	0	1.708	N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.



#1 Tetrachloro-m-xylene


R.T.: 4.315 min
Delta R.T.: 0.001 min
Response: 903788
Conc: 29.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-110719-07




#1 Tetrachloro-m-xylene

R.T.: 3.566 min
Delta R.T.: -0.003 min
Response: 640298
Conc: 28.83 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.942 min
Delta R.T.: 0.007 min
Response: 864373
Conc: 21.27 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.535 min
Delta R.T.: -0.005 min
Response: 691326
Conc: 23.64 ng/ml



#3 AR-1016-1


R.T.: 5.475 min
Delta R.T.: 0.000 min
Response: 93402
Conc: 67.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-110719-07




#3 AR-1016-1

R.T.: 4.643 min
Delta R.T.: -0.005 min
Response: 138792
Conc: 135.44 ng/ml




#4 AR-1016-2

R.T.: 5.519 min
Delta R.T.: 0.022 min
Response: 185005
Conc: 94.26 ng/ml



#4 AR-1016-2


R.T.: 4.686 min
Delta R.T.: 0.020 min
Response: 128296
Conc: 94.16 ng/ml



#5 AR-1016-3


R.T.: 5.519 min
Delta R.T.: -0.040 min
Response: 185005
Conc: 148.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-110719-07




#5 AR-1016-3

R.T.: 4.838 min
Delta R.T.: -0.003 min
Response: 139178
Conc: 186.91 ng/ml




#6 AR-1016-4

R.T.: 5.694 min
Delta R.T.: 0.038 min
Response: 603735
Conc: 587.00 ng/ml



#6 AR-1016-4


R.T.: 4.919 min
Delta R.T.: 0.037 min
Response: 153803
Conc: 246.87 ng/ml



#7 AR-1016-5


R.T.: 5.949 min
Delta R.T.: 0.000 min
Response: 555350
Conc: 521.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-110719-07




#7 AR-1016-5

R.T.: 5.087 min
Delta R.T.: -0.007 min
Response: 34445
Conc: 43.37 ng/ml




#8 AR-1221-1

R.T.: 4.512 min
Delta R.T.: -0.012 min
Response: 39826
Conc: 98.84 ng/ml



#8 AR-1221-1


R.T.: 3.808 min
Delta R.T.: 0.025 min
Response: 100267
Conc: 369.08 ng/ml



#9 AR-1221-2


R.T.: 4.619 min
Delta R.T.: 0.009 min
Response: 128375
Conc: 409.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-110719-07




#9 AR-1221-2

R.T.: 3.870 min
Delta R.T.: 0.002 min
Response: 123906
Conc: 589.28 ng/ml




#10 AR-1221-3

R.T.: 4.700 min
Delta R.T.: 0.015 min
Response: 160324
Conc: 169.03 ng/ml



#10 AR-1221-3


R.T.: 3.981 min
Delta R.T.: 0.037 min
Response: 265510
Conc: 398.30 ng/ml



#11 AR-1232-1


R.T.: 4.700 min
Delta R.T.: 0.015 min
Response: 160324
Conc: 120.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-110719-07




#11 AR-1232-1

R.T.: 3.981 min
Delta R.T.: 0.038 min
Response: 265510
Conc: 281.67 ng/ml




#12 AR-1232-2

R.T.: 5.213 min
Delta R.T.: 0.004 min
Response: 204494
Conc: 286.76 ng/ml



#12 AR-1232-2


R.T.: 4.686 min
Delta R.T.: 0.019 min
Response: 128296
Conc: 127.32 ng/ml



#13 AR-1232-3


R.T.: 5.519 min
Delta R.T.: 0.020 min
Response: 185005
Conc: 129.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-110719-07




#13 AR-1232-3

R.T.: 4.838 min
Delta R.T.: -0.003 min
Response: 139178
Conc: 262.95 ng/ml




#14 AR-1232-4

R.T.: 5.694 min
Delta R.T.: 0.036 min
Response: 603735
Conc: 813.26 ng/ml



#14 AR-1232-4


R.T.: 4.919 min
Delta R.T.: -0.005 min
Response: 153803
Conc: 327.59 ng/ml



#15 AR-1232-5


R.T.: 5.749 min
Delta R.T.: 0.001 min
Response: 552818
Conc: 943.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-110719-07




#15 AR-1232-5

R.T.: 5.087 min
Delta R.T.: -0.007 min
Response: 34445
Conc: 66.82 ng/ml




#16 AR-1242-1

R.T.: 5.475 min
Delta R.T.: -0.001 min
Response: 93402
Conc: 88.28 ng/ml



#16 AR-1242-1


R.T.: 4.643 min
Delta R.T.: -0.005 min
Response: 138792
Conc: 175.95 ng/ml



#17 AR-1242-2


R.T.: 5.519 min
Delta R.T.: 0.021 min
Response: 185005
Conc: 123.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-110719-07




#17 AR-1242-2

R.T.: 4.686 min
Delta R.T.: 0.020 min
Response: 128296
Conc: 121.19 ng/ml




#18 AR-1242-3

R.T.: 5.519 min
Delta R.T.: -0.041 min
Response: 185005
Conc: 193.59 ng/ml



#18 AR-1242-3


R.T.: 4.838 min
Delta R.T.: -0.003 min
Response: 139178
Conc: 242.92 ng/ml



#19 AR-1242-4


R.T.: 5.694 min
Delta R.T.: 0.037 min
Response: 603735
Conc: 768.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-110719-07




#19 AR-1242-4

R.T.: 4.919 min
Delta R.T.: -0.005 min
Response: 153803
Conc: 269.38 ng/ml




#20 AR-1242-5

R.T.: 6.392 min
Delta R.T.: 0.003 min
Response: 411355
Conc: 395.86 ng/ml



#20 AR-1242-5


R.T.: 5.439 min
Delta R.T.: -0.004 min
Response: 152670
Conc: 196.15 ng/ml



#21 AR-1248-1


R.T.: 5.475 min
Delta R.T.: 0.000 min
Response: 93402
Conc: 106.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-110719-07




#21 AR-1248-1

R.T.: 4.643 min
Delta R.T.: -0.005 min
Response: 138792
Conc: 215.38 ng/ml




#22 AR-1248-2

R.T.: 5.749 min
Delta R.T.: 0.002 min
Response: 552818
Conc: 436.30 ng/ml



#22 AR-1248-2


R.T.: 4.919 min
Delta R.T.: 0.036 min
Response: 153803
Conc: 178.45 ng/ml



#23 AR-1248-3


R.T.: 5.949 min
Delta R.T.: 0.000 min
Response: 555350
Conc: 389.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-110719-07




#23 AR-1248-3

R.T.: 4.919 min
Delta R.T.: -0.006 min
Response: 153803
Conc: 173.84 ng/ml




#24 AR-1248-4

R.T.: 6.352 min
Delta R.T.: 0.000 min
Response: 224578
Conc: 129.75 ng/ml



#24 AR-1248-4


R.T.: 5.087 min
Delta R.T.: -0.008 min
Response: 34445
Conc: 32.00 ng/ml



#25 AR-1248-5


R.T.: 6.392 min
Delta R.T.: 0.003 min
Response: 411355
Conc: 243.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-110719-07




#25 AR-1248-5

R.T.: 5.481 min
Delta R.T.: -0.004 min
Response: 43814
Conc: 39.28 ng/ml




#26 AR-1254-1

R.T.: 6.325 min
Delta R.T.: 0.001 min
Response: 368208
Conc: 210.76 ng/ml



#26 AR-1254-1


R.T.: 5.439 min
Delta R.T.: -0.006 min
Response: 152670
Conc: 86.17 ng/ml



#27 AR-1254-2


R.T.: 6.544 min
Delta R.T.: 0.001 min
Response: 427847
Conc: 153.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-110719-07




#27 AR-1254-2

R.T.: 5.586 min
Delta R.T.: -0.005 min
Response: 160273
Conc: 101.65 ng/ml




#28 AR-1254-3

R.T.: 6.911 min
Delta R.T.: 0.004 min
Response: 417411
Conc: 132.96 ng/ml



#28 AR-1254-3


R.T.: 5.988 min
Delta R.T.: -0.005 min
Response: 229625
Conc: 85.64 ng/ml



#29 AR-1254-4


R.T.: 7.194 min
Delta R.T.: 0.003 min
Response: 226611
Conc: 91.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-110719-07




#29 AR-1254-4

R.T.: 6.215 min
Delta R.T.: -0.005 min
Response: 135630
Conc: 76.48 ng/ml




#30 AR-1254-5

R.T.: 7.611 min
Delta R.T.: 0.003 min
Response: 211614
Conc: 78.07 ng/ml



#30 AR-1254-5


R.T.: 6.632 min
Delta R.T.: -0.005 min
Response: 277837
Conc: 111.19 ng/ml



#31 AR-1260-1


R.T.: 7.071 min
Delta R.T.: 0.003 min
Response: 192331
Conc: 95.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-110719-07




#31 AR-1260-1

R.T.: 6.117 min
Delta R.T.: -0.005 min
Response: 112630
Conc: 71.72 ng/ml




#32 AR-1260-2

R.T.: 7.323 min
Delta R.T.: -0.001 min
Response: 308991
Conc: 122.65 ng/ml



#32 AR-1260-2


R.T.: 6.304 min
Delta R.T.: -0.004 min
Response: 143345
Conc: 72.87 ng/ml



#33 AR-1260-3


R.T.: 7.683 min
Delta R.T.: 0.000 min
Response: 39030
Conc: 19.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-110719-07




#33 AR-1260-3

R.T.: 6.457 min
Delta R.T.: -0.005 min
Response: 97500
Conc: 54.67 ng/ml




#34 AR-1260-4

R.T.: 7.900 min
Delta R.T.: -0.005 min
Response: 89566
Conc: 37.61 ng/ml



#34 AR-1260-4


R.T.: 6.934 min
Delta R.T.: 0.002 min
Response: 80279
Conc: 50.99 ng/ml



#35 AR-1260-5


R.T.: 8.225 min
Delta R.T.: 0.010 min
Response: 121659
Conc: 25.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-110719-07




#35 AR-1260-5

R.T.: 7.168 min
Delta R.T.: -0.004 min
Response: 93333
Conc: 24.07 ng/ml




#36 AR-1262-1

R.T.: 7.683 min
Delta R.T.: 0.001 min
Response: 39030
Conc: 12.12 ng/ml



#36 AR-1262-1


R.T.: 6.701 min
Delta R.T.: -0.026 min
Response: 36641
Conc: 35.05 ng/ml



#37 AR-1262-2


R.T.: 8.225 min
Delta R.T.: 0.008 min
Response: 121659
Conc: 20.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-110719-07




#37 AR-1262-2

R.T.: 7.168 min
Delta R.T.: -0.005 min
Response: 93333
Conc: 19.68 ng/ml




#38 AR-1262-3

R.T.: 8.534 min
Delta R.T.: 0.018 min
Response: 76344
Conc: 18.63 ng/ml



#38 AR-1262-3


R.T.: 7.453 min
Delta R.T.: -0.003 min
Response: 74257
Conc: 39.62 ng/ml



#39 AR-1262-4


R.T.: 8.610 min
Delta R.T.: 0.007 min
Response: 28124
Conc: 8.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-110719-07




#39 AR-1262-4

R.T.: 7.535 min
Delta R.T.: 0.016 min
Response: 626243
Conc: 177.38 ng/ml




#40 AR-1262-5

R.T.: 9.234 min
Delta R.T.: 0.007 min
Response: 281489
Conc: 129.22 ng/ml



#40 AR-1262-5


R.T.: 8.007 min
Delta R.T.: -0.005 min
Response: 18353
Conc: 12.02 ng/ml



#41 AR-1268-1


R.T.: 8.534 min
Delta R.T.: 0.021 min
Response: 76344
Conc: 11.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-110719-07




#41 AR-1268-1

R.T.: 7.453 min
Delta R.T.: -0.003 min
Response: 74257
Conc: 13.91 ng/ml




#42 AR-1268-2

R.T.: 8.610 min
Delta R.T.: 0.006 min
Response: 28124
Conc: 4.48 ng/ml



#42 AR-1268-2


R.T.: 7.535 min
Delta R.T.: 0.016 min
Response: 626243
Conc: 125.67 ng/ml



#43 AR-1268-3


R.T.: 8.821 min
Delta R.T.: 0.000 min
Response: 58479
Conc: 11.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-110719-07




#43 AR-1268-3

R.T.: 7.771 min
Delta R.T.: 0.044 min
Response: 116482
Conc: 29.34 ng/ml




#44 AR-1268-4

R.T.: 9.234 min
Delta R.T.: 0.006 min
Response: 281489
Conc: 115.80 ng/ml



#44 AR-1268-4


R.T.: 8.007 min
Delta R.T.: -0.004 min
Response: 18353
Conc: 10.83 ng/ml



#45 AR-1268-5

R.T.: 9.621 min
Delta R.T.: 0.002 min
Response: 26264
Conc: 1.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
PCB-110719-07



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T. : 8.293 min
Response: 0
Conc: N.D.