

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0112222\
 Data File : P0090885.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Nov 2022 20:54
 Operator : YP/AJ
 Sample : N5734-01 10X
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 5683

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/23/2022
 Supervised By :Ankita Jodhani 11/23/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 23 04:47:10 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0102722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 28 05:05:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.295	3.477	4811755	1195204	1.497m	1.125
2) SA Decachlor...	10.049	8.483	10276402	1683750	4.193	1.635 #

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112222\
Data File : PO090885.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Nov 2022 20:54
Operator : YP/AJ
Sample : N5734-01 10X
Misc :
ALS Vial : 17 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
5683


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/23/2022
Supervised By :Ankita Jodhani 11/23/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 23 04:47:10 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO102722.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 28 05:05:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.295 min
Delta R.T.: -0.012 min
Response: 4811755
Conc: 1.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
5683


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/23/2022
Supervised By :Ankita Jodhani 11/23/2022




#1 Tetrachloro-m-xylene

R.T.: 3.477 min
Delta R.T.: -0.010 min
Response: 1195204
Conc: 1.12 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.049 min
Delta R.T.: -0.005 min
Response: 10276402
Conc: 4.19 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.483 min
Delta R.T.: -0.014 min
Response: 1683750
Conc: 1.63 ng/ml