

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112223\
 Data File : PO100020.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Nov 2023 17:26
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 22 22:55:19 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 15:58:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.483	3.706	77648870	67679638	45.493	45.123
2) SA Decachlor...	10.371	8.805	37584440	52218830	51.744	46.666
Target Compounds						
26) L6 AR-1254-1	6.538	5.628	21495669	23163567	446.404	446.853
27) L6 AR-1254-2	6.760	5.776	31161287	21070929	454.376	452.479
28) L6 AR-1254-3	7.130	6.185	30104892	35339302	472.549	453.122
29) L6 AR-1254-4	7.419	6.415	19177608	20309210	507.292	475.884
30) L6 AR-1254-5	7.844	6.837	21622556	34098570	505.173	471.215



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112223\
Data File : PO100020.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Nov 2023 17:26
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 22 22:55:19 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 15:58:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.483 min
Delta R.T.: 0.003 min
Response: 77648870
Conc: 45.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.706 min
Delta R.T.: 0.000 min
Response: 67679638
Conc: 45.12 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.371 min
Delta R.T.: 0.014 min
Response: 37584440
Conc: 51.74 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.805 min
Delta R.T.: 0.004 min
Response: 52218830
Conc: 46.67 ng/ml



#26 AR-1254-1


R.T.: 6.538 min
Delta R.T.: 0.004 min
Response: 21495669
Conc: 446.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.628 min
Delta R.T.: 0.002 min
Response: 23163567
Conc: 446.85 ng/ml




#27 AR-1254-2

R.T.: 6.760 min
Delta R.T.: 0.005 min
Response: 31161287
Conc: 454.38 ng/ml



#27 AR-1254-2


R.T.: 5.776 min
Delta R.T.: 0.002 min
Response: 21070929
Conc: 452.48 ng/ml



#28 AR-1254-3


R.T.: 7.130 min
Delta R.T.: 0.005 min
Response: 30104892
Conc: 472.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.185 min
Delta R.T.: 0.002 min
Response: 35339302
Conc: 453.12 ng/ml




#29 AR-1254-4

R.T.: 7.419 min
Delta R.T.: 0.005 min
Response: 19177608
Conc: 507.29 ng/ml



#29 AR-1254-4


R.T.: 6.415 min
Delta R.T.: 0.003 min
Response: 20309210
Conc: 475.88 ng/ml



#30 AR-1254-5

R.T.: 7.844 min
Delta R.T.: 0.005 min
Response: 21622556
Conc: 505.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.837 min
Delta R.T.: 0.002 min
Response: 34098570
Conc: 471.22 ng/ml