

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112223\
 Data File : PO100027.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Nov 2023 19:24
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 22 22:57:19 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 15:58:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.482	3.706	73626121	64217605	43.136	42.815
2) SA Decachlor...	10.369	8.806	34529641	49181417	47.539	43.951
Target Compounds						
21) L5 AR-1248-1	5.670	4.812	11606308	8817909	448.941	441.077
22) L5 AR-1248-2	5.948	5.053	17851094	12140610	436.491	433.891
23) L5 AR-1248-3	6.155	5.096	18883581	12728844	440.865	440.781
24) L5 AR-1248-4	6.564	5.271	17802654	14699243	472.327	448.289
25) L5 AR-1248-5	6.603	5.668	17320660	13691505	468.977	458.344


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112223\
Data File : PO100027.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Nov 2023 19:24
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 22 22:57:19 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 15:58:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.482 min
Delta R.T.: 0.002 min
Response: 73626121
Conc: 43.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.706 min
Delta R.T.: 0.000 min
Response: 64217605
Conc: 42.81 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.369 min
Delta R.T.: 0.012 min
Response: 34529641
Conc: 47.54 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.806 min
Delta R.T.: 0.005 min
Response: 49181417
Conc: 43.95 ng/ml



#21 AR-1248-1


R.T.: 5.670 min
Delta R.T.: 0.004 min
Response: 11606308
Conc: 448.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.812 min
Delta R.T.: 0.002 min
Response: 8817909
Conc: 441.08 ng/ml




#22 AR-1248-2

R.T.: 5.948 min
Delta R.T.: 0.004 min
Response: 17851094
Conc: 436.49 ng/ml



#22 AR-1248-2


R.T.: 5.053 min
Delta R.T.: 0.002 min
Response: 12140610
Conc: 433.89 ng/ml



#23 AR-1248-3


R.T.: 6.155 min
Delta R.T.: 0.004 min
Response: 18883581
Conc: 440.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.096 min
Delta R.T.: 0.002 min
Response: 12728844
Conc: 440.78 ng/ml




#24 AR-1248-4

R.T.: 6.564 min
Delta R.T.: 0.004 min
Response: 17802654
Conc: 472.33 ng/ml



#24 AR-1248-4


R.T.: 5.271 min
Delta R.T.: 0.002 min
Response: 14699243
Conc: 448.29 ng/ml



#25 AR-1248-5

R.T.: 6.603 min
Delta R.T.: 0.004 min
Response: 17320660
Conc: 468.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.668 min
Delta R.T.: 0.002 min
Response: 13691505
Conc: 458.34 ng/ml