

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112224\
 Data File : PO108175.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Nov 2024 09:34
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 22 12:46:46 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 19 03:13:56 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.372	3.639	478.6E6	191.5E6	54.968	56.480
2)	SA Decachlor...	10.104	8.638	174.0E6	176.5E6	56.959	51.878
Target Compounds							
16)	L4 AR-1242-1	5.528	4.720	118.5E6	49606391	535.126	560.562
17)	L4 AR-1242-2	5.550	4.739	170.7E6	69464096	542.746	561.935
18)	L4 AR-1242-3	5.613	4.915	107.6E6	37563332	531.398	566.706
19)	L4 AR-1242-4	5.711	4.999	86082473	35442478	551.414	566.366
20)	L4 AR-1242-5	6.450	5.520	85365364	45283609	530.734	541.795



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112224\
Data File : PO108175.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Nov 2024 09:34
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 22 12:46:46 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111824.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 19 03:13:56 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.372 min
Delta R.T.: 0.002 min
Response: 478557277
Conc: 54.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.639 min
Delta R.T.: 0.000 min
Response: 191490101
Conc: 56.48 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.104 min
Delta R.T.: 0.013 min
Response: 173963603
Conc: 56.96 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.638 min
Delta R.T.: 0.002 min
Response: 176475628
Conc: 51.88 ng/ml



#16 AR-1242-1


R.T.: 5.528 min
Delta R.T.: 0.003 min
Response: 118459429
Conc: 535.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.720 min
Delta R.T.: 0.000 min
Response: 49606391
Conc: 560.56 ng/ml




#17 AR-1242-2

R.T.: 5.550 min
Delta R.T.: 0.003 min
Response: 170694555
Conc: 542.75 ng/ml



#17 AR-1242-2


R.T.: 4.739 min
Delta R.T.: 0.000 min
Response: 69464096
Conc: 561.93 ng/ml



#18 AR-1242-3


R.T.: 5.613 min
Delta R.T.: 0.004 min
Response: 107591247
Conc: 531.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.915 min
Delta R.T.: 0.000 min
Response: 37563332
Conc: 566.71 ng/ml




#19 AR-1242-4

R.T.: 5.711 min
Delta R.T.: 0.003 min
Response: 86082473
Conc: 551.41 ng/ml



#19 AR-1242-4


R.T.: 4.999 min
Delta R.T.: 0.000 min
Response: 35442478
Conc: 566.37 ng/ml



#20 AR-1242-5

R.T.: 6.450 min
Delta R.T.: 0.005 min
Response: 85365364
Conc: 530.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.520 min
Delta R.T.: 0.000 min
Response: 45283609
Conc: 541.79 ng/ml