

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0112320.M
 Title : GC EXTRACTABLES
 Last Update : Wed Nov 25 09:42:52 2020
 Response Via : Initial Calibration

Calibration Files

500 =P0073444.D 1000 =P0073442.D 750 =P0073443.D
 250 =P0073445.D 50 =P0073446.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	8.697	7.786	8.720	8.696	8.810	8.542 E4	4.98
2) SA	Decachlorobiphenyl	8.651	7.731	8.313	9.004	9.357	8.611 E4	7.29
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	3.451	2.998	3.340	3.664	3.839	3.458 E3	9.30
27) L6	AR-1254-2	5.410	4.686	5.232	5.686	6.227	5.448 E3	10.43
28) L6	AR-1254-3	5.618	4.960	5.453	5.835	6.470	5.667 E3	9.75
29) L6	AR-1254-4	4.772	4.220	4.603	5.016	5.579	4.838 E3	10.45
30) L6	AR-1254-5	4.759	4.212	4.613	4.938	5.280	4.760 E3	8.30
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

Signal #2 Calibration Files

500 =P0073447.D 1000 =P0073442.D 750 =P0073443.D
 250 =P0073445.D 50 =P0073446.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	5.137	4.619	5.140	5.253	5.575	5.145 E4	6.68
2) SA	Decachlorobiphenyl	4.755	4.134	4.578	5.067	5.839	4.875 E4	13.05
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\
 Method File : P0112320.M
 Title : GC EXTRACTABLES
 Last Update : Wed Nov 25 09:42:52 2020
 Response Via : Initial Calibration

Calibration Files

500 =P0073444.D 1000 =P0073442.D 750 =P0073443.D
 250 =P0073445.D 50 =P0073446.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	3.078	2.541	2.971	3.324	3.656	3.114 E3	13.32
27) L6	AR-1254-2	2.692	2.255	2.573	2.912	3.217	2.730 E3	13.23
28) L6	AR-1254-3	4.171	3.519	4.034	4.432	4.990	4.229 E3	12.77
29) L6	AR-1254-4	2.842	2.382	2.743	3.061	3.408	2.887 E3	13.19
30) L6	AR-1254-5	3.732	3.155	3.583	3.775	4.221	3.693 E3	10.39
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range