

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112423\
 Data File : PO100055.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Nov 2023 19:30
 Operator : YP/AJ
 Sample : 05525-02
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TWP13

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 25 00:05:56 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 15:58:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.479	3.703	45770212	38371035	26.816	25.582m
2) SA Decachlor...	10.358	8.800	23553634	33969688	32.427	30.357

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112423\
Data File : PO100055.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Nov 2023 19:30
Operator : YP/AJ
Sample : 05525-02
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
TWP13

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 25 00:05:56 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 15:58:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.479 min
Delta R.T.: -0.001 min
Response: 45770212
Conc: 26.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
TWP13




#1 Tetrachloro-m-xylene

R.T.: 3.703 min
Delta R.T.: -0.002 min
Response: 38371035
Conc: 25.58 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.358 min
Delta R.T.: 0.001 min
Response: 23553634
Conc: 32.43 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.800 min
Delta R.T.: -0.001 min
Response: 33969688
Conc: 30.36 ng/ml