

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112423\
 Data File : PO100063.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Nov 2023 22:33
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 25 00:10:32 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 15:58:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.479	3.704	72481324	63762295	42.465	42.511
2) SA Decachlor...	10.358	8.798	34722600	50640845	47.804	45.255
Target Compounds						
21) L5 AR-1248-1	5.666	4.810	11492157	8963329	444.526	448.351
22) L5 AR-1248-2	5.944	5.050	17643318	12173380	431.410	435.062
23) L5 AR-1248-3	6.150	5.093	18446662	12787890	430.664	442.825
24) L5 AR-1248-4	6.559	5.268	17809116	14586959	472.498	444.865
25) L5 AR-1248-5	6.598	5.665	17450859	13590695	472.502	454.969



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112423\
Data File : PO100063.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Nov 2023 22:33
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 25 00:10:32 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 15:58:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.479 min
Delta R.T.: -0.001 min
Response: 72481324
Conc: 42.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.704 min
Delta R.T.: -0.002 min
Response: 63762295
Conc: 42.51 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.358 min
Delta R.T.: 0.001 min
Response: 34722600
Conc: 47.80 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.798 min
Delta R.T.: -0.003 min
Response: 50640845
Conc: 45.26 ng/ml



#21 AR-1248-1


R.T.: 5.666 min
Delta R.T.: 0.000 min
Response: 11492157
Conc: 444.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.810 min
Delta R.T.: 0.000 min
Response: 8963329
Conc: 448.35 ng/ml




#22 AR-1248-2

R.T.: 5.944 min
Delta R.T.: 0.000 min
Response: 17643318
Conc: 431.41 ng/ml



#22 AR-1248-2


R.T.: 5.050 min
Delta R.T.: -0.001 min
Response: 12173380
Conc: 435.06 ng/ml



#23 AR-1248-3


R.T.: 6.150 min
Delta R.T.: 0.000 min
Response: 18446662
Conc: 430.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.093 min
Delta R.T.: -0.001 min
Response: 12787890
Conc: 442.83 ng/ml




#24 AR-1248-4

R.T.: 6.559 min
Delta R.T.: 0.000 min
Response: 17809116
Conc: 472.50 ng/ml



#24 AR-1248-4


R.T.: 5.268 min
Delta R.T.: 0.000 min
Response: 14586959
Conc: 444.86 ng/ml



#25 AR-1248-5

R.T.: 6.598 min
Delta R.T.: 0.000 min
Response: 17450859
Conc: 472.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.665 min
Delta R.T.: 0.000 min
Response: 13590695
Conc: 454.97 ng/ml