

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0112520\
 Data File : P0073545.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 Nov 2020 9:36
 Operator : DD\AJ
 Sample : PB133240BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB133240BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 25 11:42:27 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0112320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 25 10:04:26 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.940	4.004	1667419	928324	19.933	18.935
2) SA Decachlor...	10.963	9.271	1557892	877649	18.895	19.141

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112520\
Data File : PO073545.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 Nov 2020 9:36
Operator : DD\AJ
Sample : PB133240BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB133240BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 25 11:42:27 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112320.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 25 10:04:26 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.940 min
Delta R.T.: -0.004 min
Response: 1667419
Conc: 19.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB133240BL




#1 Tetrachloro-m-xylene

R.T.: 4.004 min
Delta R.T.: -0.007 min
Response: 928324
Conc: 18.93 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.963 min
Delta R.T.: -0.009 min
Response: 1557892
Conc: 18.89 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.271 min
Delta R.T.: -0.011 min
Response: 877649
Conc: 19.14 ng/ml