

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112522\
 Data File : PO090964.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Nov 2022 05:12
 Operator : YP/AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 26 05:54:20 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Nov 26 05:48:52 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.298	3.475	184.1E6	58517412	50.416	50.213
2) SA Decachlor...	10.048	8.480	144.3E6	52380994	50.434	50.245
Target Compounds						
21) L5 AR-1248-1	5.474	4.561	37022759	12207999	506.973	502.602
22) L5 AR-1248-2	5.750	4.799	53669249	17199736	499.602	509.690
23) L5 AR-1248-3	5.955	4.840	57644965	17654881	497.339	507.325
24) L5 AR-1248-4	6.362	5.012	60841107	20680398	497.805	503.172
25) L5 AR-1248-5	6.400	5.406	58831066	19000622	502.029	502.620



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112522\
Data File : PO090964.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Nov 2022 05:12
Operator : YP/AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 26 05:54:20 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Nov 26 05:48:52 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.298 min
Delta R.T.: 0.000 min
Response: 184144492
Conc: 50.42 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.475 min
Delta R.T.: 0.000 min
Response: 58517412
Conc: 50.21 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.048 min
Delta R.T.: -0.003 min
Response: 144253525
Conc: 50.43 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.480 min
Delta R.T.: -0.002 min
Response: 52380994
Conc: 50.24 ng/ml



#21 AR-1248-1


R.T.: 5.474 min
Delta R.T.: 0.000 min
Response: 37022759
Conc: 506.97 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.561 min
Delta R.T.: 0.000 min
Response: 12207999
Conc: 502.60 ng/ml




#22 AR-1248-2

R.T.: 5.750 min
Delta R.T.: -0.002 min
Response: 53669249
Conc: 499.60 ng/ml



#22 AR-1248-2


R.T.: 4.799 min
Delta R.T.: 0.000 min
Response: 17199736
Conc: 509.69 ng/ml



#23 AR-1248-3


R.T.: 5.955 min
Delta R.T.: -0.001 min
Response: 57644965
Conc: 497.34 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 4.840 min
Delta R.T.: 0.000 min
Response: 17654881
Conc: 507.33 ng/ml




#24 AR-1248-4

R.T.: 6.362 min
Delta R.T.: -0.001 min
Response: 60841107
Conc: 497.81 ng/ml



#24 AR-1248-4


R.T.: 5.012 min
Delta R.T.: 0.000 min
Response: 20680398
Conc: 503.17 ng/ml



#25 AR-1248-5

R.T.: 6.400 min
Delta R.T.: -0.002 min
Response: 58831066
Conc: 502.03 ng/ml

Instrument :
ECD_O
ClientSampleId :



#25 AR-1248-5

R.T.: 5.406 min
Delta R.T.: 0.000 min
Response: 19000622
Conc: 502.62 ng/ml