

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0112619.M
 Title : GC EXTRACTABLES
 Last Update : Wed Nov 27 03:34:17 2019
 Response Via : Initial Calibration

Calibration Files

500 =P0064098.D 1000 =P0064096.D 750 =P0064097.D
 250 =P0064099.D 50 =P0064100.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.672	2.615	2.533	2.637	2.813	2.654 E4	3.87
2) SA	Decachlorobiphenyl	3.518	3.186	3.333	3.690	3.751	3.496 E4	6.79
3) L1	AR-1016-1	1.191	1.104	1.123	1.257	1.451	1.225 E3	11.42
4) L1	AR-1016-2	1.690	1.587	1.605	1.772	2.026	1.736 E3	10.26
5) L1	AR-1016-3	1.073	1.002	1.019	1.111	1.261	1.093 E3	9.44
6) L1	AR-1016-4	0.884	0.833	0.849	0.900	1.016	0.896 E3	8.02
7) L1	AR-1016-5	0.914	0.833	0.859	0.933	1.070	0.922 E3	9.99
8) L2	AR-1221-1	3.396					3.396 E2	0.00
9) L2	AR-1221-2	2.670					2.670 E2	0.00
10) L2	AR-1221-3	8.062					8.062 E2	0.00
11) L3	AR-1232-1	1.238					1.238 E3	0.00
12) L3	AR-1232-2	6.616					6.616 E2	0.00
13) L3	AR-1232-3	1.359					1.359 E3	0.00
14) L3	AR-1232-4	7.023					7.023 E2	0.00
15) L3	AR-1232-5	5.475					5.475 E2	0.00
16) L4	AR-1242-1	9.619					9.619 E2	0.00
17) L4	AR-1242-2	1.354					1.354 E3	0.00
18) L4	AR-1242-3	8.763					8.763 E2	0.00
19) L4	AR-1242-4	7.162					7.162 E2	0.00
20) L4	AR-1242-5	8.789					8.789 E2	0.00
21) L5	AR-1248-1	7.388					7.388 E2	0.00
22) L5	AR-1248-2	1.075					1.075 E3	0.00
23) L5	AR-1248-3	1.173					1.173 E3	0.00
24) L5	AR-1248-4	1.403					1.403 E3	0.00
25) L5	AR-1248-5	1.350					1.350 E3	0.00
26) L6	AR-1254-1	1.457					1.457 E3	0.00
27) L6	AR-1254-2	2.336					2.336 E3	0.00
28) L6	AR-1254-3	2.643					2.643 E3	0.00
29) L6	AR-1254-4	2.042					2.042 E3	0.00
30) L6	AR-1254-5	2.140					2.140 E3	0.00
31) L7	AR-1260-1	1.758	1.585	1.669	1.878	1.925	1.763 E3	8.02
32) L7	AR-1260-2	2.195	1.973	2.058	2.307	2.485	2.204 E3	9.20
33) L7	AR-1260-3	1.746	1.553	1.625	1.794	1.784	1.700 E3	6.25
34) L7	AR-1260-4	2.108	1.887	1.947	2.227	2.090	2.052 E3	6.60
35) L7	AR-1260-5	4.160	3.854	3.987	4.337	4.842	4.236 E3	9.07
36) L8	AR-1262-1	2.574					2.574 E3	0.00
37) L8	AR-1262-2	4.856					4.856 E3	0.00
38) L8	AR-1262-3	3.401					3.401 E3	0.00
39) L8	AR-1262-4	2.591					2.591 E3	0.00
40) L8	AR-1262-5	1.905					1.905 E3	0.00
41) L9	AR-1268-1	5.608					5.608 E3	0.00
42) L9	AR-1268-2	5.209					5.209 E3	0.00
43) L9	AR-1268-3	4.323					4.323 E3	0.00
44) L9	AR-1268-4	2.033					2.033 E3	0.00
45) L9	AR-1268-5	1.323					1.323 E4	0.00

Signal #2 Calibration Files

500 =P0064098.D 1000 =P0064096.D 750 =P0064097.D
 250 =P0064099.D 50 =P0064100.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.175	2.043	2.114	2.150	2.174	2.131 E4	2.60
2) SA	Decachlorobiphenyl	3.001	2.773	2.847	3.098	3.065	2.957 E4	4.76
3) L1	AR-1016-1	0.970	0.879	0.915	1.010	1.093	0.974 E3	8.60
4) L1	AR-1016-2	1.306	1.198	1.239	1.332	1.441	1.303 E3	7.18
5) L1	AR-1016-3	7.149	6.476	6.789	7.255	8.097	7.153 E2	8.54
6) L1	AR-1016-4	6.013	5.377	5.643	6.242	6.835	6.022 E2	9.35

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\
 Method File : P0112619.M
 Title : GC EXTRACTABLES
 Last Update : Wed Nov 27 03:34:17 2019
 Response Via : Initial Calibration

Calibration Files

500 =P0064098.D 1000 =P0064096.D 750 =P0064097.D
 250 =P0064099.D 50 =P0064100.D

Compound			500	1000	750	250	50	Avg	%RSD	
7)	L1	AR-1016-5	7.826	6.999	7.356	8.079	9.506	7.953	E2	12.11
8)	L2	AR-1221-1	2.537					2.537	E2	0.00
9)	L2	AR-1221-2	1.959					1.959	E2	0.00
10)	L2	AR-1221-3	6.207					6.207	E2	0.00
11)	L3	AR-1232-1	9.815					9.815	E2	0.00
12)	L3	AR-1232-2	1.043					1.043	E3	0.00
13)	L3	AR-1232-3	5.605					5.605	E2	0.00
14)	L3	AR-1232-4	5.026					5.026	E2	0.00
15)	L3	AR-1232-5	5.658					5.658	E2	0.00
16)	L4	AR-1242-1	7.802					7.802	E2	0.00
17)	L4	AR-1242-2	1.039					1.039	E3	0.00
18)	L4	AR-1242-3	5.709					5.709	E2	0.00
19)	L4	AR-1242-4	5.766					5.766	E2	0.00
20)	L4	AR-1242-5	7.935					7.935	E2	0.00
21)	L5	AR-1248-1	5.987					5.987	E2	0.00
22)	L5	AR-1248-2	7.963					7.963	E2	0.00
23)	L5	AR-1248-3	8.122					8.122	E2	0.00
24)	L5	AR-1248-4	9.940					9.940	E2	0.00
25)	L5	AR-1248-5	1.038					1.038	E3	0.00
26)	L6	AR-1254-1	1.668					1.668	E3	0.00
27)	L6	AR-1254-2	1.504					1.504	E3	0.00
28)	L6	AR-1254-3	2.548					2.548	E3	0.00
29)	L6	AR-1254-4	1.698					1.698	E3	0.00
30)	L6	AR-1254-5	2.431					2.431	E3	0.00
31)	L7	AR-1260-1	1.596	1.419	1.499	1.697	1.877	1.618	E3	11.04
32)	L7	AR-1260-2	2.018	1.800	1.896	2.152	2.478	2.069	E3	12.77
33)	L7	AR-1260-3	1.863	1.678	1.762	1.947	2.108	1.872	E3	8.90
34)	L7	AR-1260-4	1.635	1.463	1.542	1.728	1.829	1.639	E3	8.86
35)	L7	AR-1260-5	4.008	3.659	3.786	4.113	4.185	3.950	E3	5.62
36)	L8	AR-1262-1	1.034					1.034	E3	0.00
37)	L8	AR-1262-2	4.815					4.815	E3	0.00
38)	L8	AR-1262-3	1.866					1.866	E3	0.00
39)	L8	AR-1262-4	3.527					3.527	E3	0.00
40)	L8	AR-1262-5	1.600					1.600	E3	0.00
41)	L9	AR-1268-1	5.404					5.404	E3	0.00
42)	L9	AR-1268-2	4.924					4.924	E3	0.00
43)	L9	AR-1268-3	3.970					3.970	E3	0.00
44)	L9	AR-1268-4	1.748					1.748	E3	0.00
45)	L9	AR-1268-5	1.209					1.209	E4	0.00

(#) = Out of Range