

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0112619\
 Data File : P0064140.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Nov 2019 22:18
 Operator : SM/AJ
 Sample : PB125131BSD
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB125131BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 27 03:54:16 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0112619.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 27 03:34:17 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.315	3.564	568926	453284	21.438	21.270
2)	SA Decachlor...	9.937	8.526	598948	533106	17.134	18.030
Target Compounds							
3)	L1 AR-1016-1	5.475	4.639	269101	212717	219.687	218.499
4)	L1 AR-1016-2	5.497	4.659	378208	286031	217.851	219.510
5)	L1 AR-1016-3	5.559	4.833	244355	153577	223.538	214.706
6)	L1 AR-1016-4	5.656	4.873	203402	127859	226.939	212.323
7)	L1 AR-1016-5	5.949	5.085	200024	165416	217.031	207.984
8)	L2 AR-1221-1	4.519	3.775	20254	17479	59.644	68.890
9)	L2 AR-1221-2	4.610	3.861	35686	25111	133.672	128.171
10)	L2 AR-1221-3	4.685	3.937	131302	98340	162.857	158.430
11)	L3 AR-1232-1	4.685	3.937	131302	98340	106.047	100.195
12)	L3 AR-1232-2	5.209	4.659	180418	286031	272.714	274.266
13)	L3 AR-1232-3	5.497	4.833	378208	153577	278.366	273.976
14)	L3 AR-1232-4	5.656	4.915	203402	155771	289.612	309.952
15)	L3 AR-1232-5	5.746	5.085	184961	165416	337.834	292.364
16)	L4 AR-1242-1	5.475	4.639	269101	212717	279.772	272.635
17)	L4 AR-1242-2	5.497	4.659	378208	286031	279.378	275.336
18)	L4 AR-1242-3	5.559	4.833	244355	153577	278.840	268.992
19)	L4 AR-1242-4	5.656	4.915	203402	155771	284.015	270.174
20)	L4 AR-1242-5	6.389	5.435	30074	137907	34.219	173.805 #
21)	L5 AR-1248-1	5.475	4.639	269101	212717	364.217	355.301
22)	L5 AR-1248-2	5.746	4.873	184961	127859	172.068	160.575
23)	L5 AR-1248-3	5.949	4.915	200024	155771	170.544	191.789
24)	L5 AR-1248-4	6.350	5.085	34932	165416	24.895	166.411 #
25)	L5 AR-1248-5	6.389	5.475	30074	21748	22.270	20.945
26)	L6 AR-1254-1	6.323	5.435	159029	137907	109.111	82.685
27)	L6 AR-1254-2	6.540	5.581	176938	140123	75.739	93.183
28)	L6 AR-1254-3	6.906	5.997	119862	248195	45.351	97.390 #
29)	L6 AR-1254-4	7.190	6.210	68306	28884	33.446	17.014 #
30)	L6 AR-1254-5	7.607	6.625	601185	532098	280.966	218.900
31)	L7 AR-1260-1	7.068	6.112	423719	367186	240.355	226.993
32)	L7 AR-1260-2	7.325	6.298	502517	461849	228.033	223.229
33)	L7 AR-1260-3	7.681	6.452	317957	413815	186.993	221.085
34)	L7 AR-1260-4	7.906	6.922	393049	296001	191.567	180.574
35)	L7 AR-1260-5	8.216	7.162	712664	731301	168.242	185.122
36)	L8 AR-1262-1	7.681	6.716	317957	195205	123.521	188.834 #
37)	L8 AR-1262-2	8.216	7.162	712664	731301	146.746	151.889
38)	L8 AR-1262-3	8.530	7.445	450035	134807	132.309	72.236 #
39)	L8 AR-1262-4	8.582	7.508	244343	519105	94.300	147.199 #
40)	L8 AR-1262-5	9.229	8.000	176564	141875	92.667	88.683
41)	L9 AR-1268-1	8.530	7.445	450035	134807	80.251	24.945 #
42)	L9 AR-1268-2	8.582	7.508	244343	519105	46.904	105.421 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0112619\
 Data File : P0064140.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Nov 2019 22:18
 Operator : SM/AJ
 Sample : PB125131BSD
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB125131BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 27 03:54:16 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0112619.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 27 03:34:17 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
43) L9	AR-1268-3	8.817	7.715	15490	12887	3.583	3.247
44) L9	AR-1268-4	9.229	8.000	176564	141875	86.833	81.185
45) L9	AR-1268-5	9.619	8.280	52882	57202	3.997	4.729



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112619\
Data File : PO064140.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Nov 2019 22:18
Operator : SM/AJ
Sample : PB125131BSD
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB125131BSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 27 03:54:16 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112619.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 27 03:34:17 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.315 min
Delta R.T.: -0.001 min
Response: 568926
Conc: 21.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD




#1 Tetrachloro-m-xylene

R.T.: 3.564 min
Delta R.T.: 0.000 min
Response: 453284
Conc: 21.27 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.937 min
Delta R.T.: -0.004 min
Response: 598948
Conc: 17.13 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.526 min
Delta R.T.: -0.005 min
Response: 533106
Conc: 18.03 ng/ml



#3 AR-1016-1


R.T.: 5.475 min
Delta R.T.: 0.000 min
Response: 269101
Conc: 219.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD




#3 AR-1016-1

R.T.: 4.639 min
Delta R.T.: -0.002 min
Response: 212717
Conc: 218.50 ng/ml




#4 AR-1016-2

R.T.: 5.497 min
Delta R.T.: -0.001 min
Response: 378208
Conc: 217.85 ng/ml



#4 AR-1016-2


R.T.: 4.659 min
Delta R.T.: -0.002 min
Response: 286031
Conc: 219.51 ng/ml



#5 AR-1016-3


R.T.: 5.559 min
Delta R.T.: -0.001 min
Response: 244355
Conc: 223.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD




#5 AR-1016-3

R.T.: 4.833 min
Delta R.T.: -0.002 min
Response: 153577
Conc: 214.71 ng/ml




#6 AR-1016-4

R.T.: 5.656 min
Delta R.T.: -0.002 min
Response: 203402
Conc: 226.94 ng/ml



#6 AR-1016-4


R.T.: 4.873 min
Delta R.T.: -0.002 min
Response: 127859
Conc: 212.32 ng/ml



#7 AR-1016-5


R.T.: 5.949 min
Delta R.T.: -0.002 min
Response: 200024
Conc: 217.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD




#7 AR-1016-5

R.T.: 5.085 min
Delta R.T.: -0.002 min
Response: 165416
Conc: 207.98 ng/ml




#8 AR-1221-1

R.T.: 4.519 min
Delta R.T.: -0.005 min
Response: 20254
Conc: 59.64 ng/ml



#8 AR-1221-1


R.T.: 3.775 min
Delta R.T.: -0.002 min
Response: 17479
Conc: 68.89 ng/ml



#9 AR-1221-2


R.T.: 4.610 min
Delta R.T.: 0.000 min
Response: 35686
Conc: 133.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD




#9 AR-1221-2

R.T.: 3.861 min
Delta R.T.: 0.000 min
Response: 25111
Conc: 128.17 ng/ml




#10 AR-1221-3

R.T.: 4.685 min
Delta R.T.: 0.000 min
Response: 131302
Conc: 162.86 ng/ml



#10 AR-1221-3


R.T.: 3.937 min
Delta R.T.: 0.000 min
Response: 98340
Conc: 158.43 ng/ml



#11 AR-1232-1


R.T.: 4.685 min
Delta R.T.: 0.000 min
Response: 131302
Conc: 106.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD




#11 AR-1232-1

R.T.: 3.937 min
Delta R.T.: -0.001 min
Response: 98340
Conc: 100.19 ng/ml




#12 AR-1232-2

R.T.: 5.209 min
Delta R.T.: -0.001 min
Response: 180418
Conc: 272.71 ng/ml



#12 AR-1232-2


R.T.: 4.659 min
Delta R.T.: -0.001 min
Response: 286031
Conc: 274.27 ng/ml



#13 AR-1232-3


R.T.: 5.497 min
Delta R.T.: -0.002 min
Response: 378208
Conc: 278.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD




#13 AR-1232-3

R.T.: 4.833 min
Delta R.T.: -0.002 min
Response: 153577
Conc: 273.98 ng/ml




#14 AR-1232-4

R.T.: 5.656 min
Delta R.T.: -0.003 min
Response: 203402
Conc: 289.61 ng/ml



#14 AR-1232-4


R.T.: 4.915 min
Delta R.T.: -0.002 min
Response: 155771
Conc: 309.95 ng/ml



#15 AR-1232-5


R.T.: 5.746 min
Delta R.T.: -0.001 min
Response: 184961
Conc: 337.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD




#15 AR-1232-5

R.T.: 5.085 min
Delta R.T.: -0.003 min
Response: 165416
Conc: 292.36 ng/ml




#16 AR-1242-1

R.T.: 5.475 min
Delta R.T.: 0.000 min
Response: 269101
Conc: 279.77 ng/ml



#16 AR-1242-1


R.T.: 4.639 min
Delta R.T.: -0.002 min
Response: 212717
Conc: 272.64 ng/ml



#17 AR-1242-2


R.T.: 5.497 min
Delta R.T.: 0.000 min
Response: 378208
Conc: 279.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD




#17 AR-1242-2

R.T.: 4.659 min
Delta R.T.: -0.001 min
Response: 286031
Conc: 275.34 ng/ml




#18 AR-1242-3

R.T.: 5.559 min
Delta R.T.: 0.000 min
Response: 244355
Conc: 278.84 ng/ml



#18 AR-1242-3


R.T.: 4.833 min
Delta R.T.: 0.000 min
Response: 153577
Conc: 268.99 ng/ml



#19 AR-1242-4


R.T.: 5.656 min
Delta R.T.: 0.000 min
Response: 203402
Conc: 284.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD




#19 AR-1242-4

R.T.: 4.915 min
Delta R.T.: -0.001 min
Response: 155771
Conc: 270.17 ng/ml




#20 AR-1242-5

R.T.: 6.389 min
Delta R.T.: 0.000 min
Response: 30074
Conc: 34.22 ng/ml



#20 AR-1242-5


R.T.: 5.435 min
Delta R.T.: -0.001 min
Response: 137907
Conc: 173.80 ng/ml



#21 AR-1248-1


R.T.: 5.475 min
Delta R.T.: 0.000 min
Response: 269101
Conc: 364.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD




#21 AR-1248-1

R.T.: 4.639 min
Delta R.T.: 0.000 min
Response: 212717
Conc: 355.30 ng/ml




#22 AR-1248-2

R.T.: 5.746 min
Delta R.T.: 0.000 min
Response: 184961
Conc: 172.07 ng/ml



#22 AR-1248-2


R.T.: 4.873 min
Delta R.T.: 0.000 min
Response: 127859
Conc: 160.57 ng/ml



#23 AR-1248-3


R.T.: 5.949 min
Delta R.T.: -0.001 min
Response: 200024
Conc: 170.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD




#23 AR-1248-3

R.T.: 4.915 min
Delta R.T.: 0.000 min
Response: 155771
Conc: 191.79 ng/ml




#24 AR-1248-4

R.T.: 6.350 min
Delta R.T.: -0.002 min
Response: 34932
Conc: 24.89 ng/ml



#24 AR-1248-4


R.T.: 5.085 min
Delta R.T.: -0.001 min
Response: 165416
Conc: 166.41 ng/ml



#25 AR-1248-5


R.T.: 6.389 min
Delta R.T.: -0.001 min
Response: 30074
Conc: 22.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD




#25 AR-1248-5

R.T.: 5.475 min
Delta R.T.: -0.002 min
Response: 21748
Conc: 20.94 ng/ml




#26 AR-1254-1

R.T.: 6.323 min
Delta R.T.: -0.002 min
Response: 159029
Conc: 109.11 ng/ml



#26 AR-1254-1


R.T.: 5.435 min
Delta R.T.: -0.002 min
Response: 137907
Conc: 82.69 ng/ml



#27 AR-1254-2


R.T.: 6.540 min
Delta R.T.: -0.003 min
Response: 176938
Conc: 75.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD




#27 AR-1254-2

R.T.: 5.581 min
Delta R.T.: -0.002 min
Response: 140123
Conc: 93.18 ng/ml




#28 AR-1254-3

R.T.: 6.906 min
Delta R.T.: -0.002 min
Response: 119862
Conc: 45.35 ng/ml



#28 AR-1254-3


R.T.: 5.997 min
Delta R.T.: 0.012 min
Response: 248195
Conc: 97.39 ng/ml



#29 AR-1254-4


R.T.: 7.190 min
Delta R.T.: -0.004 min
Response: 68306
Conc: 33.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD




#29 AR-1254-4

R.T.: 6.210 min
Delta R.T.: -0.002 min
Response: 28884
Conc: 17.01 ng/ml




#30 AR-1254-5

R.T.: 7.607 min
Delta R.T.: -0.003 min
Response: 601185
Conc: 280.97 ng/ml



#30 AR-1254-5


R.T.: 6.625 min
Delta R.T.: -0.003 min
Response: 532098
Conc: 218.90 ng/ml



#31 AR-1260-1


R.T.: 7.068 min
Delta R.T.: -0.002 min
Response: 423719
Conc: 240.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD




#31 AR-1260-1

R.T.: 6.112 min
Delta R.T.: -0.003 min
Response: 367186
Conc: 226.99 ng/ml




#32 AR-1260-2

R.T.: 7.325 min
Delta R.T.: -0.002 min
Response: 502517
Conc: 228.03 ng/ml



#32 AR-1260-2


R.T.: 6.298 min
Delta R.T.: -0.003 min
Response: 461849
Conc: 223.23 ng/ml



#33 AR-1260-3


R.T.: 7.681 min
Delta R.T.: -0.003 min
Response: 317957
Conc: 186.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD




#33 AR-1260-3

R.T.: 6.452 min
Delta R.T.: -0.003 min
Response: 413815
Conc: 221.08 ng/ml




#34 AR-1260-4

R.T.: 7.906 min
Delta R.T.: -0.004 min
Response: 393049
Conc: 191.57 ng/ml



#34 AR-1260-4


R.T.: 6.922 min
Delta R.T.: -0.003 min
Response: 296001
Conc: 180.57 ng/ml



#35 AR-1260-5


R.T.: 8.216 min
Delta R.T.: -0.003 min
Response: 712664
Conc: 168.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD




#35 AR-1260-5

R.T.: 7.162 min
Delta R.T.: -0.004 min
Response: 731301
Conc: 185.12 ng/ml




#36 AR-1262-1

R.T.: 7.681 min
Delta R.T.: -0.003 min
Response: 317957
Conc: 123.52 ng/ml



#36 AR-1262-1


R.T.: 6.716 min
Delta R.T.: -0.002 min
Response: 195205
Conc: 188.83 ng/ml



#37 AR-1262-2


R.T.: 8.216 min
Delta R.T.: -0.003 min
Response: 712664
Conc: 146.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD




#37 AR-1262-2

R.T.: 7.162 min
Delta R.T.: -0.003 min
Response: 731301
Conc: 151.89 ng/ml




#38 AR-1262-3

R.T.: 8.530 min
Delta R.T.: 0.013 min
Response: 450035
Conc: 132.31 ng/ml



#38 AR-1262-3


R.T.: 7.445 min
Delta R.T.: -0.002 min
Response: 134807
Conc: 72.24 ng/ml



#39 AR-1262-4


R.T.: 8.582 min
Delta R.T.: -0.022 min
Response: 244343
Conc: 94.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD




#39 AR-1262-4

R.T.: 7.508 min
Delta R.T.: -0.003 min
Response: 519105
Conc: 147.20 ng/ml




#40 AR-1262-5

R.T.: 9.229 min
Delta R.T.: -0.003 min
Response: 176564
Conc: 92.67 ng/ml



#40 AR-1262-5


R.T.: 8.000 min
Delta R.T.: -0.003 min
Response: 141875
Conc: 88.68 ng/ml



#41 AR-1268-1


R.T.: 8.530 min
Delta R.T.: 0.014 min
Response: 450035
Conc: 80.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD




#41 AR-1268-1

R.T.: 7.445 min
Delta R.T.: -0.002 min
Response: 134807
Conc: 24.94 ng/ml




#42 AR-1268-2

R.T.: 8.582 min
Delta R.T.: -0.024 min
Response: 244343
Conc: 46.90 ng/ml



#42 AR-1268-2


R.T.: 7.508 min
Delta R.T.: -0.004 min
Response: 519105
Conc: 105.42 ng/ml



#43 AR-1268-3


R.T.: 8.817 min
Delta R.T.: -0.007 min
Response: 15490
Conc: 3.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD




#43 AR-1268-3

R.T.: 7.715 min
Delta R.T.: -0.003 min
Response: 12887
Conc: 3.25 ng/ml




#44 AR-1268-4

R.T.: 9.229 min
Delta R.T.: -0.003 min
Response: 176564
Conc: 86.83 ng/ml



#44 AR-1268-4


R.T.: 8.000 min
Delta R.T.: -0.003 min
Response: 141875
Conc: 81.19 ng/ml



#45 AR-1268-5

R.T.: 9.619 min
Delta R.T.: -0.005 min
Response: 52882
Conc: 4.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD



#45 AR-1268-5

R.T.: 8.280 min
Delta R.T.: -0.004 min
Response: 57202
Conc: 4.73 ng/ml