

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0112619\
 Data File : P0064146.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Nov 2019 00:49
 Operator : SM/AJ
 Sample : K6014-01
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CS-01

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 27 03:57:13 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0112619.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 27 03:34:17 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.313	3.563	729009	562122	27.470	26.377
2)	SA Decachlor...	9.936	8.525	851622	779243	24.362	26.355
Target Compounds							
9)	L2 AR-1221-2	4.617	0.000	10610	0	39.744	N.D. #
10)	L2 AR-1221-3	4.617	0.000	10610	0	13.160	N.D. #
11)	L3 AR-1232-1	4.617	0.000	10610	0	8.570	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0112619\
Data File : P0064146.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 Nov 2019 00:49
Operator : SM/AJ
Sample : K6014-01
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
CS-01

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 27 03:57:13 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0112619.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 27 03:34:17 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.313 min
Delta R.T.: -0.003 min
Response: 729009
Conc: 27.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-01




#1 Tetrachloro-m-xylene

R.T.: 3.563 min
Delta R.T.: -0.002 min
Response: 562122
Conc: 26.38 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.936 min
Delta R.T.: -0.005 min
Response: 851622
Conc: 24.36 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.525 min
Delta R.T.: -0.005 min
Response: 779243
Conc: 26.35 ng/ml



#9 AR-1221-2


R.T.: 4.617 min
Delta R.T.: 0.007 min
Response: 10610
Conc: 39.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-01




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 3.862 min
Response: 0
Conc: N.D.




#10 AR-1221-3

R.T.: 4.617 min
Delta R.T.: -0.068 min
Response: 10610
Conc: 13.16 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 3.937 min
Response: 0
Conc: N.D.



#11 AR-1232-1

R.T.: 4.617 min
Delta R.T.: -0.068 min
Response: 10610
Conc: 8.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
CS-01



#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 3.938 min
Response: 0
Conc: N.D.