

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112619\
 Data File : P0064182.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Nov 2019 13:50
 Operator : SM/AJ
 Sample : PB125131BSD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB125131BSD

Manual Integrations
 APPROVED

Ankita
 11/27/2019 4:17:16 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 27 14:10:04 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112619.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 27 03:34:17 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.315	3.560	663316	537644	24.994	25.228
2)	SA Decachlor...	9.942	8.526	817358	795838	23.382	26.916
Target Compounds							
3)	L1 AR-1016-1	5.476	4.638	319972	258480	261.216	265.505
4)	L1 AR-1016-2	5.499	4.657	446136	345830	256.978	265.402
5)	L1 AR-1016-3	5.561	4.831	281651	177112	257.657	247.609m
6)	L1 AR-1016-4	5.658	4.873	228304	136185	254.723	226.148
7)	L1 AR-1016-5	5.951	5.084	224283	198644	243.353	249.762
31)	L7 AR-1260-1	7.070	6.111	417882	398413	237.044	246.298
32)	L7 AR-1260-2	7.327	6.298	540495	526610	245.267	254.530
33)	L7 AR-1260-3	7.684	6.452	418605	456255	246.185	243.759
34)	L7 AR-1260-4	7.909	6.922	491348	403007	239.477	245.853
35)	L7 AR-1260-5	8.219	7.163	956236	998970	225.743	252.880

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112619\
Data File : PO064182.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 Nov 2019 13:50
Operator : SM/AJ
Sample : PB125131BSD
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
PB125131BSD


Manual Integrations
APPROVED

Ankita
11/27/2019 4:17:16 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 27 14:10:04 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112619.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 27 03:34:17 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.315 min
Delta R.T.: -0.001 min
Response: 663316
Conc: 24.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD


Manual Integrations
APPROVED

Ankita
11/27/2019 4:17:16 PM




#1 Tetrachloro-m-xylene

R.T.: 3.560 min
Delta R.T.: -0.004 min
Response: 537644
Conc: 25.23 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.942 min
Delta R.T.: 0.001 min
Response: 817358
Conc: 23.38 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.526 min
Delta R.T.: -0.004 min
Response: 795838
Conc: 26.92 ng/ml




#3 AR-1016-1

R.T.: 5.476 min
Delta R.T.: 0.000 min
Response: 319972
Conc: 261.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD


Manual Integrations
APPROVED

Ankita
11/27/2019 4:17:16 PM




#3 AR-1016-1

R.T.: 4.638 min
Delta R.T.: -0.003 min
Response: 258480
Conc: 265.50 ng/ml




#4 AR-1016-2

R.T.: 5.499 min
Delta R.T.: 0.000 min
Response: 446136
Conc: 256.98 ng/ml



#4 AR-1016-2

R.T.: 4.657 min
Delta R.T.: -0.003 min
Response: 345830
Conc: 265.40 ng/ml




#5 AR-1016-3

R.T.: 5.561 min
Delta R.T.: 0.000 min
Response: 281651
Conc: 257.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD


Manual Integrations
APPROVED

Ankita
11/27/2019 4:17:16 PM




#5 AR-1016-3

R.T.: 4.831 min
Delta R.T.: -0.004 min
Response: 177112
Conc: 247.61 ng/ml m




#6 AR-1016-4

R.T.: 5.658 min
Delta R.T.: 0.000 min
Response: 228304
Conc: 254.72 ng/ml



#6 AR-1016-4

R.T.: 4.873 min
Delta R.T.: -0.002 min
Response: 136185
Conc: 226.15 ng/ml




#7 AR-1016-5

R.T.: 5.951 min
Delta R.T.: 0.000 min
Response: 224283
Conc: 243.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD


Manual Integrations
APPROVED

Ankita
11/27/2019 4:17:16 PM




#7 AR-1016-5

R.T.: 5.084 min
Delta R.T.: -0.003 min
Response: 198644
Conc: 249.76 ng/ml




#31 AR-1260-1

R.T.: 7.070 min
Delta R.T.: 0.000 min
Response: 417882
Conc: 237.04 ng/ml



#31 AR-1260-1

R.T.: 6.111 min
Delta R.T.: -0.003 min
Response: 398413
Conc: 246.30 ng/ml




#32 AR-1260-2

R.T.: 7.327 min
Delta R.T.: 0.000 min
Response: 540495
Conc: 245.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD


Manual Integrations
APPROVED

Ankita
11/27/2019 4:17:16 PM




#32 AR-1260-2

R.T.: 6.298 min
Delta R.T.: -0.003 min
Response: 526610
Conc: 254.53 ng/ml




#33 AR-1260-3

R.T.: 7.684 min
Delta R.T.: 0.000 min
Response: 418605
Conc: 246.19 ng/ml



#33 AR-1260-3

R.T.: 6.452 min
Delta R.T.: -0.003 min
Response: 456255
Conc: 243.76 ng/ml




#34 AR-1260-4

R.T.: 7.909 min
Delta R.T.: 0.000 min
Response: 491348
Conc: 239.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB125131BSD


Manual Integrations
APPROVED

Ankita
11/27/2019 4:17:16 PM




#34 AR-1260-4

R.T.: 6.922 min
Delta R.T.: -0.003 min
Response: 403007
Conc: 245.85 ng/ml



#35 AR-1260-5

R.T.: 8.219 min
Delta R.T.: 0.000 min
Response: 956236
Conc: 225.74 ng/ml



#35 AR-1260-5

R.T.: 7.163 min
Delta R.T.: -0.002 min
Response: 998970
Conc: 252.88 ng/ml