

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0112624\
 Data File : P0108219.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Nov 2024 10:17
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 26 14:21:29 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0111824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 19 03:13:56 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.373	3.638	444.8E6	187.0E6	51.095	55.166
2) SA Decachlor...	10.097	8.633	154.0E6	161.8E6	50.426	47.579
Target Compounds						
16) L4 AR-1242-1	5.528	4.719	110.0E6	48066062	496.779	543.156
17) L4 AR-1242-2	5.550	4.738	158.6E6	67280407	504.193	544.270
18) L4 AR-1242-3	5.613	4.914	99551856	36221880	491.691	546.468
19) L4 AR-1242-4	5.710	4.997	79227369	34167178	507.503	545.987
20) L4 AR-1242-5	6.449	5.519	76600912	43140656	476.243	516.155



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112624\
Data File : PO108219.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Nov 2024 10:17
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 26 14:21:29 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111824.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 19 03:13:56 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.373 min
Delta R.T.: 0.003 min
Response: 444839952
Conc: 51.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.638 min
Delta R.T.: 0.000 min
Response: 187034273
Conc: 55.17 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.097 min
Delta R.T.: 0.006 min
Response: 154010416
Conc: 50.43 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.633 min
Delta R.T.: -0.002 min
Response: 161849163
Conc: 47.58 ng/ml



#16 AR-1242-1


R.T.: 5.528 min
Delta R.T.: 0.003 min
Response: 109970756
Conc: 496.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.719 min
Delta R.T.: 0.000 min
Response: 48066062
Conc: 543.16 ng/ml




#17 AR-1242-2

R.T.: 5.550 min
Delta R.T.: 0.002 min
Response: 158569345
Conc: 504.19 ng/ml



#17 AR-1242-2


R.T.: 4.738 min
Delta R.T.: 0.000 min
Response: 67280407
Conc: 544.27 ng/ml



#18 AR-1242-3


R.T.: 5.613 min
Delta R.T.: 0.003 min
Response: 99551856
Conc: 491.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.914 min
Delta R.T.: -0.001 min
Response: 36221880
Conc: 546.47 ng/ml




#19 AR-1242-4

R.T.: 5.710 min
Delta R.T.: 0.002 min
Response: 79227369
Conc: 507.50 ng/ml



#19 AR-1242-4


R.T.: 4.997 min
Delta R.T.: 0.000 min
Response: 34167178
Conc: 545.99 ng/ml



#20 AR-1242-5

R.T.: 6.449 min
Delta R.T.: 0.004 min
Response: 76600912
Conc: 476.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.519 min
Delta R.T.: 0.000 min
Response: 43140656
Conc: 516.16 ng/ml