

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0112624\
 Data File : P0108234.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Nov 2024 15:56
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 26 16:06:25 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0111824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 19 03:13:56 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.370	3.637	475.4E6	194.4E6	54.611	57.352
2)	SA Decachlor...	10.094	8.632	168.5E6	169.3E6	55.154	49.768
Target Compounds							
16)	L4 AR-1242-1	5.525	4.718	118.3E6	50041566	534.453	565.479
17)	L4 AR-1242-2	5.548	4.737	167.0E6	69993378	530.895	566.216
18)	L4 AR-1242-3	5.610	4.913	106.3E6	37861297	524.866	571.201
19)	L4 AR-1242-4	5.708	4.997	84746723	35779803	542.858	571.757
20)	L4 AR-1242-5	6.446	5.518	83862337	44956812	521.389	537.885



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112624\
Data File : PO108234.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Nov 2024 15:56
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 26 16:06:25 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111824.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 19 03:13:56 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.370 min
Delta R.T.: 0.000 min
Response: 475445188
Conc: 54.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.637 min
Delta R.T.: -0.002 min
Response: 194448696
Conc: 57.35 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.094 min
Delta R.T.: 0.003 min
Response: 168451788
Conc: 55.15 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.632 min
Delta R.T.: -0.003 min
Response: 169295812
Conc: 49.77 ng/ml



#16 AR-1242-1


R.T.: 5.525 min
Delta R.T.: 0.000 min
Response: 118310512
Conc: 534.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.718 min
Delta R.T.: -0.001 min
Response: 50041566
Conc: 565.48 ng/ml




#17 AR-1242-2

R.T.: 5.548 min
Delta R.T.: 0.000 min
Response: 166967276
Conc: 530.90 ng/ml



#17 AR-1242-2


R.T.: 4.737 min
Delta R.T.: -0.001 min
Response: 69993378
Conc: 566.22 ng/ml



#18 AR-1242-3


R.T.: 5.610 min
Delta R.T.: 0.000 min
Response: 106268663
Conc: 524.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.913 min
Delta R.T.: -0.002 min
Response: 37861297
Conc: 571.20 ng/ml




#19 AR-1242-4

R.T.: 5.708 min
Delta R.T.: 0.000 min
Response: 84746723
Conc: 542.86 ng/ml



#19 AR-1242-4


R.T.: 4.997 min
Delta R.T.: -0.001 min
Response: 35779803
Conc: 571.76 ng/ml



#20 AR-1242-5

R.T.: 6.446 min
Delta R.T.: 0.002 min
Response: 83862337
Conc: 521.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.518 min
Delta R.T.: -0.002 min
Response: 44956812
Conc: 537.88 ng/ml