

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO112625\
 Data File : PO1153461.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Nov 2025 02:04
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 12/01/2025
 Supervised By :mohammad ahmed 12/02/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 28 20:36:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 06:00:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.339	3.643	401.2E6	249.1E6	57.002	55.769
2) SA Decachlor...	9.916	8.626	247.7E6	202.4E6	56.040m	51.840
Target Compounds						
21) L5 AR-1248-1	5.477	4.717	74461960	59508759	532.458	544.405m
22) L5 AR-1248-2	5.748	4.953	104.7E6	73829625	538.181	495.832
23) L5 AR-1248-3	5.948	4.995	114.7E6	80133262	543.997	515.632
24) L5 AR-1248-4	6.344	5.165	135.0E6	95640413	534.336	521.906
25) L5 AR-1248-5	6.383	5.556	127.0E6	96566509	538.837	523.882

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112625\
Data File : PO1153461.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Nov 2025 02:04
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 12/01/2025
Supervised By :mohammad ahmed 12/02/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 28 20:36:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

