

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112625\
 Data File : PO1153426.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Nov 2025 10:10
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 26 11:54:19 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 05 06:00:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.339	3.641	417.6E6	256.8E6	59.331	57.482
2)	SA Decachlor...	9.918	8.628	255.9E6	219.9E6	57.898	56.336
Target Compounds							
21)	L5 AR-1248-1	5.478	4.718	76585678	59590779	547.644	545.155m
22)	L5 AR-1248-2	5.749	4.954	108.0E6	75349618	555.512	506.040
23)	L5 AR-1248-3	5.949	4.996	116.8E6	79622536	554.111	512.346
24)	L5 AR-1248-4	6.345	5.166	136.1E6	94125782	538.912	513.640
25)	L5 AR-1248-5	6.383	5.557	128.8E6	96460436	546.092	523.307



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112625\
Data File : PO1153426.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Nov 2025 10:10
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 26 11:54:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO110425.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 05 06:00:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.339 min
Delta R.T.: -0.009 min
Response: 417577378
Conc: 59.33 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.641 min
Delta R.T.: -0.007 min
Response: 256757777
Conc: 57.48 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.918 min
Delta R.T.: -0.012 min
Response: 255921865
Conc: 57.90 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.628 min
Delta R.T.: -0.013 min
Response: 219924012
Conc: 56.34 ng/ml



#21 AR-1248-1


R.T.: 5.478 min
Delta R.T.: -0.009 min
Response: 76585678
Conc: 547.64 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.718 min
Delta R.T.: -0.010 min
Response: 59590779
Conc: 545.16 ng/ml m




#22 AR-1248-2

R.T.: 5.749 min
Delta R.T.: -0.010 min
Response: 108034274
Conc: 555.51 ng/ml



#22 AR-1248-2


R.T.: 4.954 min
Delta R.T.: -0.010 min
Response: 75349618
Conc: 506.04 ng/ml



#23 AR-1248-3


R.T.: 5.949 min
Delta R.T.: -0.010 min
Response: 116825825
Conc: 554.11 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 4.996 min
Delta R.T.: -0.011 min
Response: 79622536
Conc: 512.35 ng/ml




#24 AR-1248-4

R.T.: 6.345 min
Delta R.T.: -0.009 min
Response: 136126152
Conc: 538.91 ng/ml



#24 AR-1248-4


R.T.: 5.166 min
Delta R.T.: -0.011 min
Response: 94125782
Conc: 513.64 ng/ml



#25 AR-1248-5

R.T.: 6.383 min
Delta R.T.: -0.010 min
Response: 128752034
Conc: 546.09 ng/ml

Instrument :
ECD_O
ClientSampleId :



#25 AR-1248-5

R.T.: 5.557 min
Delta R.T.: -0.011 min
Response: 96460436
Conc: 523.31 ng/ml