

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112823\
 Data File : PO100122.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Nov 2023 14:53
 Operator : YP/AJ
 Sample : 05553-01MS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 HD-01-112723MS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/29/2023
 Supervised By :Ankita Jodhani 11/29/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 28 23:08:03 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 15:58:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.480	3.704	57572392	46128796	33.731	30.755
2)	SA Decachlor...	10.365	8.800	29385556	35418585	40.456	31.652
Target Compounds							
3)	L1 AR-1016-1	5.668	4.810	37536547	26072781	872.409	825.709m
4)	L1 AR-1016-2	5.691	4.830	55141233	34174283	851.704	790.373m
5)	L1 AR-1016-3	5.753	5.009	35541912	19231071	815.443	829.246
6)	L1 AR-1016-4	5.853	5.051	28551570	15895529	864.952	775.347
7)	L1 AR-1016-5	6.152	5.268	26912642	19703119	857.134	772.605
31)	L7 AR-1260-1	7.293	6.313	46252927	42350910	868.691	732.914
32)	L7 AR-1260-2	7.554	6.502	50095754	51941331	947.039	782.052
33)	L7 AR-1260-3	7.917	6.659	31107408	49089344	872.088	767.171
34)	L7 AR-1260-4	8.147	7.135	38955197	37359759	941.217	740.877
35)	L7 AR-1260-5	8.476	7.377	67628335	89000766	1012.404	793.920

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112823\
Data File : PO100122.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Nov 2023 14:53
Operator : YP/AJ
Sample : 05553-01MS
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
HD-01-112723MS

Manual Integrations
APPROVED

Reviewed By : Yogesh Patel 11/29/2023
Supervised By : Ankita Jodhani 11/29/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 28 23:08:03 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 15:58:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

