

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112923\
 Data File : PO100130.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Nov 2023 12:31
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 29 15:30:59 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 15:58:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.481	3.704	73405188	63611182	43.007	42.410
2) SA Decachlor...	10.361	8.797	38845966	54572004	53.481	48.769
Target Compounds						
26) L6 AR-1254-1	6.535	5.623	20578317	22116877	427.353	426.661
27) L6 AR-1254-2	6.756	5.772	29734008	20505560	433.564	440.338
28) L6 AR-1254-3	7.127	6.182	28666696	34833077	449.974	446.631
29) L6 AR-1254-4	7.415	6.411	18411391	19911436	487.024	466.564
30) L6 AR-1254-5	7.839	6.833	21926996	34705500	512.285	479.603



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO112923\
Data File : PO100130.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Nov 2023 12:31
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 29 15:30:59 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 15:58:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.481 min
Delta R.T.: 0.000 min
Response: 73405188
Conc: 43.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.704 min
Delta R.T.: -0.002 min
Response: 63611182
Conc: 42.41 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.361 min
Delta R.T.: 0.004 min
Response: 38845966
Conc: 53.48 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.797 min
Delta R.T.: -0.004 min
Response: 54572004
Conc: 48.77 ng/ml



#26 AR-1254-1


R.T.: 6.535 min
Delta R.T.: 0.002 min
Response: 20578317
Conc: 427.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.623 min
Delta R.T.: -0.002 min
Response: 22116877
Conc: 426.66 ng/ml




#27 AR-1254-2

R.T.: 6.756 min
Delta R.T.: 0.001 min
Response: 29734008
Conc: 433.56 ng/ml



#27 AR-1254-2


R.T.: 5.772 min
Delta R.T.: -0.002 min
Response: 20505560
Conc: 440.34 ng/ml



#28 AR-1254-3


R.T.: 7.127 min
Delta R.T.: 0.001 min
Response: 28666696
Conc: 449.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.182 min
Delta R.T.: -0.001 min
Response: 34833077
Conc: 446.63 ng/ml




#29 AR-1254-4

R.T.: 7.415 min
Delta R.T.: 0.001 min
Response: 18411391
Conc: 487.02 ng/ml



#29 AR-1254-4


R.T.: 6.411 min
Delta R.T.: 0.000 min
Response: 19911436
Conc: 466.56 ng/ml



#30 AR-1254-5

R.T.: 7.839 min
Delta R.T.: 0.000 min
Response: 21926996
Conc: 512.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.833 min
Delta R.T.: -0.002 min
Response: 34705500
Conc: 479.60 ng/ml