Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO113021\
Data File : PO083298.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Dec 2021 8:57
Operator : AJ\MA
Sample : M4850-11 10X
Misc :
ALS Vial : 62 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
BC274271LOK-END-1-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 01 13:04:02 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111921.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Nov 20 02:49:51 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.947	3.949	116862	45502	1.332	1.493
2)	SA Decachlor...	11.034	9.256	109781	34330	1.670	1.142 #
	Target Compounds						
26)	L6 AR-1254-1	7.197	6.078	2926242	1685330	1051.500	866.740
27)	L6 AR-1254-2	7.427	6.239	4185120	1458602	1005.996	866.903
28)	L6 AR-1254-3	7.812	6.659	3910831	1977178	910.539	773.387
29)	L6 AR-1254-4	8.109	6.900	2792247	1285307	902.719	704.554
30)	L6 AR-1254-5	8.541	7.329	3167890	1802892	988.161	774.798

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO113021\
Data File : PO083298.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Dec 2021 8:57
Operator : AJ\MA
Sample : M4850-11 10X
Misc :
ALS Vial : 62 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
BC274271LOK-END-1-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 01 13:04:02 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111921.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Nov 20 02:49:51 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm