

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO113023\
 Data File : PO100174.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Nov 2023 14:59
 Operator : YP/AJ
 Sample : 05578-02 10X
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 15:22:34 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 15:58:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.479	3.703	2509937	2405104	1.471	1.604
2)	SA Decachlor...	10.360	8.797	1428871	1998762	1.967	1.786
Target Compounds							
3)	L1 AR-1016-1	5.667	4.812	47583	199711	1.106	6.325 #
4)	L1 AR-1016-2	5.687	4.812	21327	199711	0.329	4.619 #
5)	L1 AR-1016-3	5.750	5.010	72518	38584	1.664	1.664
6)	L1 AR-1016-4	5.857	5.053	133988	45180	4.059	2.204 #
7)	L1 AR-1016-5	6.140	5.285	89957	29002	2.865	1.137 #
8)	L2 AR-1221-1	4.712	3.935	69949	230003	3.393	12.099 #
9)	L2 AR-1221-2	4.794	3.988	328293	38086	21.735	2.923 #
10)	L2 AR-1221-3	4.849	4.085	37947	39996	0.846	1.113 #
11)	L3 AR-1232-1	4.849	4.085	37947	39996	1.064	1.363 #
12)	L3 AR-1232-2	5.390	4.812	50225	199711	2.886	10.071 #
13)	L3 AR-1232-3	5.687	5.010	21327	38584	0.753	3.723 #
14)	L3 AR-1232-4	5.857	5.072	133988	20278	9.645	2.040 #
15)	L3 AR-1232-5	5.944	5.285	117283	29002	9.557	2.685 #
16)	L4 AR-1242-1	5.667	4.812	47583	199711	1.374	7.599 #
17)	L4 AR-1242-2	5.687	4.812	21327	199711	0.406	5.680 #
18)	L4 AR-1242-3	5.750	5.010	72518	38584	2.084	2.015
19)	L4 AR-1242-4	5.857	5.072	133988	20278	5.096	1.007 #
20)	L4 AR-1242-5	6.599	5.624	56474	65253	2.526	2.793
21)	L5 AR-1248-1	5.667	4.812	47583	199711	1.841	9.990 #
22)	L5 AR-1248-2	5.944	5.053	117283	45180	2.868	1.615 #
23)	L5 AR-1248-3	6.140	5.072	89957	20278	2.100	0.702 #
24)	L5 AR-1248-4	6.559	5.285	50872	29002	1.350	0.884 #
25)	L5 AR-1248-5	6.599	5.663	56474	97689	1.529	3.270 #
26)	L6 AR-1254-1	6.530	5.624	65216	65253	1.354	1.259
27)	L6 AR-1254-2	6.756	5.778	84201	62489	1.228	1.342
28)	L6 AR-1254-3	7.127	6.173	147289	324920	2.312	4.166 #
29)	L6 AR-1254-4	7.414	6.428	190492	136944	5.039	3.209 #
30)	L6 AR-1254-5	7.832	6.839	60406	121060	1.411	1.673
31)	L7 AR-1260-1	7.291	6.302	31456	181939	0.591	3.149 #
32)	L7 AR-1260-2	7.548	6.503	73905	47030	1.397	0.708 #
33)	L7 AR-1260-3	7.914	6.647	69044	88238	1.936	1.379 #
34)	L7 AR-1260-4	8.146	7.177f	125490	395947	3.032	7.852 #
35)	L7 AR-1260-5	8.471	7.400	209681	398836	3.139	3.558
36)	L8 AR-1262-1	7.832	6.839	60406	121060	1.670	3.167 #
37)	L8 AR-1262-2	8.471	7.177f	209681	395947	2.470	5.249 #
38)	L8 AR-1262-3	8.783	7.670	9974	17552	0.166	0.301 #
39)	L8 AR-1262-4	8.885	7.760	44271	59743	0.918	0.581 #
40)	L8 AR-1262-5	9.565	8.232	11706	72863	0.448	1.672 #
41)	L9 AR-1268-1	8.783	7.670	9974	17552	0.090	0.106

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO113023\
 Data File : PO100174.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Nov 2023 14:59
 Operator : YP/AJ
 Sample : 05578-02 10X
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 15:22:34 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 15:58:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.885	7.760	44271	59743	0.460	0.391
43) L9	AR-1268-3	9.122	7.934	50091	216536	0.555	1.511 #
44) L9	AR-1268-4	9.565	8.232	11706	72863	0.412	1.551 #
45) L9	AR-1268-5	10.000	8.538	28035	61360	0.111	0.154 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO113023\
Data File : PO100174.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Nov 2023 14:59
Operator : YP/AJ
Sample : 05578-02 10X
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 30 15:22:34 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 15:58:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.479 min
Delta R.T.: -0.001 min
Response: 2509937
Conc: 1.47 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.703 min
Delta R.T.: -0.003 min
Response: 2405104
Conc: 1.60 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.360 min
Delta R.T.: 0.003 min
Response: 1428871
Conc: 1.97 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.797 min
Delta R.T.: -0.004 min
Response: 1998762
Conc: 1.79 ng/ml



#3 AR-1016-1


R.T.: 5.667 min
Delta R.T.: 0.000 min
Response: 47583
Conc: 1.11 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.812 min
Delta R.T.: 0.000 min
Response: 199711
Conc: 6.32 ng/ml




#4 AR-1016-2

R.T.: 5.687 min
Delta R.T.: -0.003 min
Response: 21327
Conc: 0.33 ng/ml



#4 AR-1016-2


R.T.: 4.812 min
Delta R.T.: -0.019 min
Response: 199711
Conc: 4.62 ng/ml



#5 AR-1016-3


R.T.: 5.750 min
Delta R.T.: -0.002 min
Response: 72518
Conc: 1.66 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 5.010 min
Delta R.T.: 0.000 min
Response: 38584
Conc: 1.66 ng/ml




#6 AR-1016-4

R.T.: 5.857 min
Delta R.T.: 0.005 min
Response: 133988
Conc: 4.06 ng/ml



#6 AR-1016-4


R.T.: 5.053 min
Delta R.T.: 0.001 min
Response: 45180
Conc: 2.20 ng/ml



#7 AR-1016-5


R.T.: 6.140 min
Delta R.T.: -0.011 min
Response: 89957
Conc: 2.87 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.285 min
Delta R.T.: 0.016 min
Response: 29002
Conc: 1.14 ng/ml




#8 AR-1221-1

R.T.: 4.712 min
Delta R.T.: 0.021 min
Response: 69949
Conc: 3.39 ng/ml



#8 AR-1221-1


R.T.: 3.935 min
Delta R.T.: 0.012 min
Response: 230003
Conc: 12.10 ng/ml



#9 AR-1221-2


R.T.: 4.794 min
Delta R.T.: 0.014 min
Response: 328293
Conc: 21.74 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.988 min
Delta R.T.: -0.023 min
Response: 38086
Conc: 2.92 ng/ml




#10 AR-1221-3

R.T.: 4.849 min
Delta R.T.: -0.008 min
Response: 37947
Conc: 0.85 ng/ml



#10 AR-1221-3


R.T.: 4.085 min
Delta R.T.: -0.003 min
Response: 39996
Conc: 1.11 ng/ml



#11 AR-1232-1


R.T.: 4.849 min
Delta R.T.: -0.007 min
Response: 37947
Conc: 1.06 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 4.085 min
Delta R.T.: -0.003 min
Response: 39996
Conc: 1.36 ng/ml




#12 AR-1232-2

R.T.: 5.390 min
Delta R.T.: -0.005 min
Response: 50225
Conc: 2.89 ng/ml



#12 AR-1232-2


R.T.: 4.812 min
Delta R.T.: -0.019 min
Response: 199711
Conc: 10.07 ng/ml



#13 AR-1232-3


R.T.: 5.687 min
Delta R.T.: -0.002 min
Response: 21327
Conc: 0.75 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 5.010 min
Delta R.T.: 0.000 min
Response: 38584
Conc: 3.72 ng/ml




#14 AR-1232-4

R.T.: 5.857 min
Delta R.T.: 0.006 min
Response: 133988
Conc: 9.65 ng/ml



#14 AR-1232-4


R.T.: 5.072 min
Delta R.T.: -0.022 min
Response: 20278
Conc: 2.04 ng/ml



#15 AR-1232-5


R.T.: 5.944 min
Delta R.T.: 0.000 min
Response: 117283
Conc: 9.56 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.285 min
Delta R.T.: 0.017 min
Response: 29002
Conc: 2.69 ng/ml




#16 AR-1242-1

R.T.: 5.667 min
Delta R.T.: 0.000 min
Response: 47583
Conc: 1.37 ng/ml



#16 AR-1242-1


R.T.: 4.812 min
Delta R.T.: 0.000 min
Response: 199711
Conc: 7.60 ng/ml



#17 AR-1242-2


R.T.: 5.687 min
Delta R.T.: -0.003 min
Response: 21327
Conc: 0.41 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.812 min
Delta R.T.: -0.020 min
Response: 199711
Conc: 5.68 ng/ml




#18 AR-1242-3

R.T.: 5.750 min
Delta R.T.: -0.003 min
Response: 72518
Conc: 2.08 ng/ml



#18 AR-1242-3


R.T.: 5.010 min
Delta R.T.: -0.001 min
Response: 38584
Conc: 2.01 ng/ml



#19 AR-1242-4


R.T.: 5.857 min
Delta R.T.: 0.004 min
Response: 133988
Conc: 5.10 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 5.072 min
Delta R.T.: -0.023 min
Response: 20278
Conc: 1.01 ng/ml




#20 AR-1242-5

R.T.: 6.599 min
Delta R.T.: 0.000 min
Response: 56474
Conc: 2.53 ng/ml



#20 AR-1242-5


R.T.: 5.624 min
Delta R.T.: -0.001 min
Response: 65253
Conc: 2.79 ng/ml



#21 AR-1248-1


R.T.: 5.667 min
Delta R.T.: 0.001 min
Response: 47583
Conc: 1.84 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.812 min
Delta R.T.: 0.002 min
Response: 199711
Conc: 9.99 ng/ml




#22 AR-1248-2

R.T.: 5.944 min
Delta R.T.: 0.000 min
Response: 117283
Conc: 2.87 ng/ml



#22 AR-1248-2


R.T.: 5.053 min
Delta R.T.: 0.002 min
Response: 45180
Conc: 1.61 ng/ml



#23 AR-1248-3


R.T.: 6.140 min
Delta R.T.: -0.011 min
Response: 89957
Conc: 2.10 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 5.072 min
Delta R.T.: -0.022 min
Response: 20278
Conc: 0.70 ng/ml




#24 AR-1248-4

R.T.: 6.559 min
Delta R.T.: 0.000 min
Response: 50872
Conc: 1.35 ng/ml



#24 AR-1248-4


R.T.: 5.285 min
Delta R.T.: 0.017 min
Response: 29002
Conc: 0.88 ng/ml



#25 AR-1248-5


R.T.: 6.599 min
Delta R.T.: 0.000 min
Response: 56474
Conc: 1.53 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.663 min
Delta R.T.: -0.003 min
Response: 97689
Conc: 3.27 ng/ml




#26 AR-1254-1

R.T.: 6.530 min
Delta R.T.: -0.003 min
Response: 65216
Conc: 1.35 ng/ml



#26 AR-1254-1


R.T.: 5.624 min
Delta R.T.: -0.002 min
Response: 65253
Conc: 1.26 ng/ml



#27 AR-1254-2


R.T.: 6.756 min
Delta R.T.: 0.001 min
Response: 84201
Conc: 1.23 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.778 min
Delta R.T.: 0.004 min
Response: 62489
Conc: 1.34 ng/ml




#28 AR-1254-3

R.T.: 7.127 min
Delta R.T.: 0.002 min
Response: 147289
Conc: 2.31 ng/ml



#28 AR-1254-3


R.T.: 6.173 min
Delta R.T.: -0.010 min
Response: 324920
Conc: 4.17 ng/ml



#29 AR-1254-4


R.T.: 7.414 min
Delta R.T.: 0.000 min
Response: 190492
Conc: 5.04 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.428 min
Delta R.T.: 0.015 min
Response: 136944
Conc: 3.21 ng/ml




#30 AR-1254-5

R.T.: 7.832 min
Delta R.T.: -0.007 min
Response: 60406
Conc: 1.41 ng/ml



#30 AR-1254-5


R.T.: 6.839 min
Delta R.T.: 0.004 min
Response: 121060
Conc: 1.67 ng/ml



#31 AR-1260-1


R.T.: 7.291 min
Delta R.T.: 0.000 min
Response: 31456
Conc: 0.59 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.302 min
Delta R.T.: -0.013 min
Response: 181939
Conc: 3.15 ng/ml




#32 AR-1260-2

R.T.: 7.548 min
Delta R.T.: -0.003 min
Response: 73905
Conc: 1.40 ng/ml



#32 AR-1260-2


R.T.: 6.503 min
Delta R.T.: 0.000 min
Response: 47030
Conc: 0.71 ng/ml



#33 AR-1260-3


R.T.: 7.914 min
Delta R.T.: 0.000 min
Response: 69044
Conc: 1.94 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.647 min
Delta R.T.: -0.013 min
Response: 88238
Conc: 1.38 ng/ml




#34 AR-1260-4

R.T.: 8.146 min
Delta R.T.: 0.003 min
Response: 125490
Conc: 3.03 ng/ml



#34 AR-1260-4


R.T.: 7.177 min
Delta R.T.: 0.041 min
Response: 395947
Conc: 7.85 ng/ml



#35 AR-1260-5


R.T.: 8.471 min
Delta R.T.: -0.001 min
Response: 209681
Conc: 3.14 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.400 min
Delta R.T.: 0.022 min
Response: 398836
Conc: 3.56 ng/ml




#36 AR-1262-1

R.T.: 7.832 min
Delta R.T.: -0.003 min
Response: 60406
Conc: 1.67 ng/ml



#36 AR-1262-1


R.T.: 6.839 min
Delta R.T.: 0.005 min
Response: 121060
Conc: 3.17 ng/ml



#37 AR-1262-2


R.T.: 8.471 min
Delta R.T.: -0.001 min
Response: 209681
Conc: 2.47 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 7.177 min
Delta R.T.: 0.042 min
Response: 395947
Conc: 5.25 ng/ml




#38 AR-1262-3

R.T.: 8.783 min
Delta R.T.: -0.013 min
Response: 9974
Conc: 0.17 ng/ml



#38 AR-1262-3


R.T.: 7.670 min
Delta R.T.: 0.007 min
Response: 17552
Conc: 0.30 ng/ml



#39 AR-1262-4


R.T.: 8.885 min
Delta R.T.: -0.004 min
Response: 44271
Conc: 0.92 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.760 min
Delta R.T.: 0.032 min
Response: 59743
Conc: 0.58 ng/ml




#40 AR-1262-5

R.T.: 9.565 min
Delta R.T.: -0.004 min
Response: 11706
Conc: 0.45 ng/ml



#40 AR-1262-5


R.T.: 8.232 min
Delta R.T.: 0.002 min
Response: 72863
Conc: 1.67 ng/ml



#41 AR-1268-1


R.T.: 8.783 min
Delta R.T.: -0.011 min
Response: 9974
Conc: 0.09 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.670 min
Delta R.T.: 0.007 min
Response: 17552
Conc: 0.11 ng/ml




#42 AR-1268-2

R.T.: 8.885 min
Delta R.T.: -0.006 min
Response: 44271
Conc: 0.46 ng/ml



#42 AR-1268-2


R.T.: 7.760 min
Delta R.T.: 0.032 min
Response: 59743
Conc: 0.39 ng/ml



#43 AR-1268-3


R.T.: 9.122 min
Delta R.T.: -0.007 min
Response: 50091
Conc: 0.56 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.934 min
Delta R.T.: -0.004 min
Response: 216536
Conc: 1.51 ng/ml




#44 AR-1268-4

R.T.: 9.565 min
Delta R.T.: -0.005 min
Response: 11706
Conc: 0.41 ng/ml



#44 AR-1268-4


R.T.: 8.232 min
Delta R.T.: 0.002 min
Response: 72863
Conc: 1.55 ng/ml



#45 AR-1268-5

R.T.: 10.000 min
Delta R.T.: -0.004 min
Response: 28035
Conc: 0.11 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.538 min
Delta R.T.: 0.005 min
Response: 61360
Conc: 0.15 ng/ml