

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO113023\
 Data File : PO100206.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Dec 2023 02:03
 Operator : YP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 01 03:23:11 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 15:58:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.478	3.703	28596268	26398924	16.754	17.600
2)	SA Decachlor...	10.354	8.794	15667441	22944008	21.570	20.504
Target Compounds							
3)	L1 AR-1016-1	5.677	4.813	99582	191216	2.314	6.056 #
4)	L1 AR-1016-2	5.701	4.813	61748	191216	0.954	4.422 #
5)	L1 AR-1016-3	5.742	5.025	195931	78813	4.495	3.398
6)	L1 AR-1016-4	5.868	5.066	53243	130503	1.613	6.366 #
7)	L1 AR-1016-5	6.151	5.327f	12467	295819	0.397	11.600 #
8)	L2 AR-1221-1	4.699	3.933	17746	328932	0.861	17.303 #
9)	L2 AR-1221-2	4.792	4.008	626022	496858	41.447	38.128
10)	L2 AR-1221-3	4.848	4.087	41664	57584	0.928	1.603 #
11)	L3 AR-1232-1	4.848	4.087	41664	57584	1.168	1.962 #
12)	L3 AR-1232-2	5.389	4.813	37751	191216	2.169	9.642 #
13)	L3 AR-1232-3	5.677	5.025	99582	78813	3.514	7.605 #
14)	L3 AR-1232-4	5.868	5.066	53243	130503	3.833	13.129 #
15)	L3 AR-1232-5	5.946	5.327f	82100	295819	6.690	27.390 #
16)	L4 AR-1242-1	5.677	4.813	99582	191216	2.875	7.276 #
17)	L4 AR-1242-2	5.701	4.813	61748	191216	1.177	5.438 #
18)	L4 AR-1242-3	5.763	5.025	52206	78813	1.500	4.116 #
19)	L4 AR-1242-4	5.868	5.066	53243	130503	2.025	6.479 #
20)	L4 AR-1242-5	6.585	5.645	55066	339843	2.463	14.544 #
21)	L5 AR-1248-1	5.677	4.813	99582	191216	3.852	9.565 #
22)	L5 AR-1248-2	5.946	5.066	82100	130503	2.007	4.664 #
23)	L5 AR-1248-3	6.151	5.066	12467	130503	0.291	4.519 #
24)	L5 AR-1248-4	6.569	5.327f	29740	295819	0.789	9.022 #
25)	L5 AR-1248-5	6.585	5.645	55066	339843	1.491	11.377 #
26)	L6 AR-1254-1	6.537	5.645	175546	339843	3.646	6.556 #
27)	L6 AR-1254-2	6.754	5.782	33215	400478	0.484	8.600 #
28)	L6 AR-1254-3	7.125	6.203	559568	604721	8.783	7.754
29)	L6 AR-1254-4	7.427	6.432	62569	825633	1.655	19.346 #
30)	L6 AR-1254-5	7.837	0.000	84624	0	1.977	N.D. #
31)	L7 AR-1260-1	7.288	0.000	52083	0	0.978	N.D. #
32)	L7 AR-1260-2	7.550	6.485	21983	563848	0.416	8.490 #
33)	L7 AR-1260-3	7.913	0.000	45692	0	1.281	N.D. #
34)	L7 AR-1260-4	8.145	0.000	61973	0	1.497	N.D. #
35)	L7 AR-1260-5	8.471	0.000	118572	0	1.775	N.D. #
36)	L8 AR-1262-1	7.837	0.000	84624	0	2.339	N.D. #
37)	L8 AR-1262-2	8.471	0.000	118572	0	1.397	N.D. #
38)	L8 AR-1262-3	8.779	0.000	36609	0	0.608	N.D. #
39)	L8 AR-1262-4	8.884	0.000	35298	0	0.732	N.D. #
40)	L8 AR-1262-5	9.577	0.000	21895	0	0.839	N.D. #
41)	L9 AR-1268-1	8.779	0.000	36609	0	0.332	N.D. #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO113023\
 Data File : PO100206.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Dec 2023 02:03
 Operator : YP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 01 03:23:11 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 15:58:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.884	0.000	35298	0	0.367	N.D. #
43) L9	AR-1268-3	9.136	7.954	35040	294897	0.389	2.058 #
44) L9	AR-1268-4	9.577	0.000	21895	0	0.770	N.D. #
45) L9	AR-1268-5	9.999	8.528	40226	229880	0.160	0.577 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO113023\
Data File : PO100206.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Dec 2023 02:03
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 01 03:23:11 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 15:58:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.478 min
Delta R.T.: -0.002 min
Response: 28596268
Conc: 16.75 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.703 min
Delta R.T.: -0.003 min
Response: 26398924
Conc: 17.60 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.354 min
Delta R.T.: -0.003 min
Response: 15667441
Conc: 21.57 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.794 min
Delta R.T.: -0.007 min
Response: 22944008
Conc: 20.50 ng/ml



#3 AR-1016-1


R.T.: 5.677 min
Delta R.T.: 0.010 min
Response: 99582
Conc: 2.31 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.813 min
Delta R.T.: 0.001 min
Response: 191216
Conc: 6.06 ng/ml




#4 AR-1016-2

R.T.: 5.701 min
Delta R.T.: 0.011 min
Response: 61748
Conc: 0.95 ng/ml



#4 AR-1016-2


R.T.: 4.813 min
Delta R.T.: -0.019 min
Response: 191216
Conc: 4.42 ng/ml



#5 AR-1016-3


R.T.: 5.742 min
Delta R.T.: -0.010 min
Response: 195931
Conc: 4.50 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 5.025 min
Delta R.T.: 0.015 min
Response: 78813
Conc: 3.40 ng/ml




#6 AR-1016-4

R.T.: 5.868 min
Delta R.T.: 0.017 min
Response: 53243
Conc: 1.61 ng/ml



#6 AR-1016-4


R.T.: 5.066 min
Delta R.T.: 0.014 min
Response: 130503
Conc: 6.37 ng/ml



#7 AR-1016-5


R.T.: 6.151 min
Delta R.T.: 0.000 min
Response: 12467
Conc: 0.40 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.327 min
Delta R.T.: 0.058 min
Response: 295819
Conc: 11.60 ng/ml




#8 AR-1221-1

R.T.: 4.699 min
Delta R.T.: 0.007 min
Response: 17746
Conc: 0.86 ng/ml



#8 AR-1221-1


R.T.: 3.933 min
Delta R.T.: 0.010 min
Response: 328932
Conc: 17.30 ng/ml



#9 AR-1221-2


R.T.: 4.792 min
Delta R.T.: 0.012 min
Response: 626022
Conc: 41.45 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 4.008 min
Delta R.T.: -0.003 min
Response: 496858
Conc: 38.13 ng/ml




#10 AR-1221-3

R.T.: 4.848 min
Delta R.T.: -0.009 min
Response: 41664
Conc: 0.93 ng/ml



#10 AR-1221-3


R.T.: 4.087 min
Delta R.T.: -0.002 min
Response: 57584
Conc: 1.60 ng/ml



#11 AR-1232-1


R.T.: 4.848 min
Delta R.T.: -0.008 min
Response: 41664
Conc: 1.17 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 4.087 min
Delta R.T.: -0.001 min
Response: 57584
Conc: 1.96 ng/ml




#12 AR-1232-2

R.T.: 5.389 min
Delta R.T.: -0.005 min
Response: 37751
Conc: 2.17 ng/ml



#12 AR-1232-2


R.T.: 4.813 min
Delta R.T.: -0.018 min
Response: 191216
Conc: 9.64 ng/ml



#13 AR-1232-3


R.T.: 5.677 min
Delta R.T.: -0.012 min
Response: 99582
Conc: 3.51 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 5.025 min
Delta R.T.: 0.015 min
Response: 78813
Conc: 7.60 ng/ml




#14 AR-1232-4

R.T.: 5.868 min
Delta R.T.: 0.017 min
Response: 53243
Conc: 3.83 ng/ml



#14 AR-1232-4


R.T.: 5.066 min
Delta R.T.: -0.029 min
Response: 130503
Conc: 13.13 ng/ml



#15 AR-1232-5


R.T.: 5.946 min
Delta R.T.: 0.002 min
Response: 82100
Conc: 6.69 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.327 min
Delta R.T.: 0.059 min
Response: 295819
Conc: 27.39 ng/ml




#16 AR-1242-1

R.T.: 5.677 min
Delta R.T.: 0.010 min
Response: 99582
Conc: 2.87 ng/ml



#16 AR-1242-1


R.T.: 4.813 min
Delta R.T.: 0.000 min
Response: 191216
Conc: 7.28 ng/ml



#17 AR-1242-2


R.T.: 5.701 min
Delta R.T.: 0.011 min
Response: 61748
Conc: 1.18 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.813 min
Delta R.T.: -0.019 min
Response: 191216
Conc: 5.44 ng/ml




#18 AR-1242-3

R.T.: 5.763 min
Delta R.T.: 0.010 min
Response: 52206
Conc: 1.50 ng/ml



#18 AR-1242-3


R.T.: 5.025 min
Delta R.T.: 0.014 min
Response: 78813
Conc: 4.12 ng/ml



#19 AR-1242-4


R.T.: 5.868 min
Delta R.T.: 0.015 min
Response: 53243
Conc: 2.03 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 5.066 min
Delta R.T.: -0.029 min
Response: 130503
Conc: 6.48 ng/ml




#20 AR-1242-5

R.T.: 6.585 min
Delta R.T.: -0.015 min
Response: 55066
Conc: 2.46 ng/ml



#20 AR-1242-5


R.T.: 5.645 min
Delta R.T.: 0.019 min
Response: 339843
Conc: 14.54 ng/ml



#21 AR-1248-1


R.T.: 5.677 min
Delta R.T.: 0.012 min
Response: 99582
Conc: 3.85 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.813 min
Delta R.T.: 0.002 min
Response: 191216
Conc: 9.56 ng/ml




#22 AR-1248-2

R.T.: 5.946 min
Delta R.T.: 0.002 min
Response: 82100
Conc: 2.01 ng/ml



#22 AR-1248-2


R.T.: 5.066 min
Delta R.T.: 0.014 min
Response: 130503
Conc: 4.66 ng/ml



#23 AR-1248-3


R.T.: 6.151 min
Delta R.T.: 0.000 min
Response: 12467
Conc: 0.29 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 5.066 min
Delta R.T.: -0.028 min
Response: 130503
Conc: 4.52 ng/ml




#24 AR-1248-4

R.T.: 6.569 min
Delta R.T.: 0.010 min
Response: 29740
Conc: 0.79 ng/ml



#24 AR-1248-4


R.T.: 5.327 min
Delta R.T.: 0.059 min
Response: 295819
Conc: 9.02 ng/ml



#25 AR-1248-5


R.T.: 6.585 min
Delta R.T.: -0.014 min
Response: 55066
Conc: 1.49 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.645 min
Delta R.T.: -0.021 min
Response: 339843
Conc: 11.38 ng/ml




#26 AR-1254-1

R.T.: 6.537 min
Delta R.T.: 0.003 min
Response: 175546
Conc: 3.65 ng/ml



#26 AR-1254-1


R.T.: 5.645 min
Delta R.T.: 0.019 min
Response: 339843
Conc: 6.56 ng/ml



#27 AR-1254-2


R.T.: 6.754 min
Delta R.T.: -0.001 min
Response: 33215
Conc: 0.48 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.782 min
Delta R.T.: 0.008 min
Response: 400478
Conc: 8.60 ng/ml




#28 AR-1254-3

R.T.: 7.125 min
Delta R.T.: 0.000 min
Response: 559568
Conc: 8.78 ng/ml



#28 AR-1254-3


R.T.: 6.203 min
Delta R.T.: 0.020 min
Response: 604721
Conc: 7.75 ng/ml



#29 AR-1254-4


R.T.: 7.427 min
Delta R.T.: 0.013 min
Response: 62569
Conc: 1.66 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.432 min
Delta R.T.: 0.019 min
Response: 825633
Conc: 19.35 ng/ml




#30 AR-1254-5

R.T.: 7.837 min
Delta R.T.: -0.002 min
Response: 84624
Conc: 1.98 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T.: 6.835 min
Response: 0
Conc: N.D.



#31 AR-1260-1


R.T.: 7.288 min
Delta R.T.: -0.003 min
Response: 52083
Conc: 0.98 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 0.000 min
Exp R.T. : 6.315 min
Response: 0
Conc: N.D.




#32 AR-1260-2

R.T.: 7.550 min
Delta R.T.: -0.002 min
Response: 21983
Conc: 0.42 ng/ml



#32 AR-1260-2


R.T.: 6.485 min
Delta R.T.: -0.018 min
Response: 563848
Conc: 8.49 ng/ml



#33 AR-1260-3


R.T.: 7.913 min
Delta R.T.: -0.002 min
Response: 45692
Conc: 1.28 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 6.660 min
Response: 0
Conc: N.D.




#34 AR-1260-4

R.T.: 8.145 min
Delta R.T.: 0.002 min
Response: 61973
Conc: 1.50 ng/ml



#34 AR-1260-4


R.T.: 0.000 min
Exp R.T. : 7.136 min
Response: 0
Conc: N.D.



#35 AR-1260-5


R.T.: 8.471 min
Delta R.T.: -0.002 min
Response: 118572
Conc: 1.78 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 0.000 min
Exp R.T. : 7.378 min
Response: 0
Conc: N.D.




#36 AR-1262-1

R.T.: 7.837 min
Delta R.T.: 0.002 min
Response: 84624
Conc: 2.34 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T. : 6.834 min
Response: 0
Conc: N.D.



#37 AR-1262-2


R.T.: 8.471 min
Delta R.T.: -0.002 min
Response: 118572
Conc: 1.40 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 7.135 min
Response: 0
Conc: N.D.




#38 AR-1262-3

R.T.: 8.779 min
Delta R.T.: -0.017 min
Response: 36609
Conc: 0.61 ng/ml



#38 AR-1262-3


R.T.: 0.000 min
Exp R.T. : 7.664 min
Response: 0
Conc: N.D.



#39 AR-1262-4


R.T.: 8.884 min
Delta R.T.: -0.005 min
Response: 35298
Conc: 0.73 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 0.000 min
Exp R.T. : 7.728 min
Response: 0
Conc: N.D.




#40 AR-1262-5

R.T.: 9.577 min
Delta R.T.: 0.008 min
Response: 21895
Conc: 0.84 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T. : 8.230 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 8.779 min
Delta R.T.: -0.014 min
Response: 36609
Conc: 0.33 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 0.000 min
Exp R.T. : 7.664 min
Response: 0
Conc: N.D.




#42 AR-1268-2

R.T.: 8.884 min
Delta R.T.: -0.007 min
Response: 35298
Conc: 0.37 ng/ml



#42 AR-1268-2


R.T.: 0.000 min
Exp R.T. : 7.728 min
Response: 0
Conc: N.D.



#43 AR-1268-3


R.T.: 9.136 min
Delta R.T.: 0.006 min
Response: 35040
Conc: 0.39 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.954 min
Delta R.T.: 0.016 min
Response: 294897
Conc: 2.06 ng/ml




#44 AR-1268-4

R.T.: 9.577 min
Delta R.T.: 0.007 min
Response: 21895
Conc: 0.77 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T. : 8.229 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 9.999 min
Delta R.T.: -0.005 min
Response: 40226
Conc: 0.16 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.528 min
Delta R.T.: -0.005 min
Response: 229880
Conc: 0.58 ng/ml