

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120123\
 Data File : PO100224.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Dec 2023 18:23
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 01 21:29:15 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 15:58:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.478	3.703	73373110	67170398	42.988	44.783
2)	SA Decachlor...	10.356	8.793	39738250	57910416	54.710	51.752
Target Compounds							
16)	L4 AR-1242-1	5.665	4.808	15290908	11773183	441.385	447.973
17)	L4 AR-1242-2	5.688	4.827	22903308	15725369	436.454	447.222
18)	L4 AR-1242-3	5.750	5.007	15102942	8800833	433.941	459.573
19)	L4 AR-1242-4	5.850	5.090	11533871	9276956	438.711	460.576
20)	L4 AR-1242-5	6.597	5.620	10391305	11297053	464.750	483.487



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120123\
Data File : PO100224.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Dec 2023 18:23
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 01 21:29:15 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 15:58:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.478 min
Delta R.T.: -0.002 min
Response: 73373110
Conc: 42.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.703 min
Delta R.T.: -0.002 min
Response: 67170398
Conc: 44.78 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.356 min
Delta R.T.: -0.001 min
Response: 39738250
Conc: 54.71 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.793 min
Delta R.T.: -0.008 min
Response: 57910416
Conc: 51.75 ng/ml



#16 AR-1242-1


R.T.: 5.665 min
Delta R.T.: -0.002 min
Response: 15290908
Conc: 441.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.808 min
Delta R.T.: -0.004 min
Response: 11773183
Conc: 447.97 ng/ml




#17 AR-1242-2

R.T.: 5.688 min
Delta R.T.: -0.002 min
Response: 22903308
Conc: 436.45 ng/ml



#17 AR-1242-2


R.T.: 4.827 min
Delta R.T.: -0.004 min
Response: 15725369
Conc: 447.22 ng/ml



#18 AR-1242-3


R.T.: 5.750 min
Delta R.T.: -0.003 min
Response: 15102942
Conc: 433.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.007 min
Delta R.T.: -0.005 min
Response: 8800833
Conc: 459.57 ng/ml




#19 AR-1242-4

R.T.: 5.850 min
Delta R.T.: -0.003 min
Response: 11533871
Conc: 438.71 ng/ml



#19 AR-1242-4


R.T.: 5.090 min
Delta R.T.: -0.004 min
Response: 9276956
Conc: 460.58 ng/ml



#20 AR-1242-5

R.T.: 6.597 min
Delta R.T.: -0.003 min
Response: 10391305
Conc: 464.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.620 min
Delta R.T.: -0.005 min
Response: 11297053
Conc: 483.49 ng/ml