

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0120219\
 Data File : P0064282.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Dec 2019 23:48
 Operator : SM/AJ
 Sample : K5676-35
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SU-01-112719

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 03:39:18 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0112719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 28 01:28:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.312	3.560	807767	688013	27.275	28.926
2)	SA Decachlor...	9.940	8.524	675496	603768	16.482	14.780
Target Compounds							
3)	L1 AR-1016-1	5.495	4.636	38477	13882	29.640	13.259 #
4)	L1 AR-1016-2	5.495	4.663	38477	19507	21.105	13.829 #
5)	L1 AR-1016-3	5.495	4.829	38477	22054	33.524	28.634
6)	L1 AR-1016-4	5.629	4.877	19727	46992	20.976	72.238 #
7)	L1 AR-1016-5	0.000	5.084	0	10478	N.D.	12.586 #
8)	L2 AR-1221-1	0.000	3.754	0	64838	N.D.	247.837 #
9)	L2 AR-1221-2	4.617	3.858	61006	90138	220.648	451.018 #
10)	L2 AR-1221-3	4.696	3.937	39088	38290	46.551	59.213 #
11)	L3 AR-1232-1	4.696	3.937	39088	38290	30.488	37.809
12)	L3 AR-1232-2	5.217	4.663	30561	19507	43.501	17.893 #
13)	L3 AR-1232-3	5.495	4.829	38477	22054	27.785	37.936 #
14)	L3 AR-1232-4	5.629	4.932	19727	42666	27.726	80.368 #
15)	L3 AR-1232-5	0.000	5.084	0	10478	N.D.	17.896 #
16)	L4 AR-1242-1	5.495	4.636	38477	13882	40.409	18.129 #
17)	L4 AR-1242-2	5.495	4.663	38477	19507	28.489	18.881 #
18)	L4 AR-1242-3	5.495	4.829	38477	22054	44.839	39.353
19)	L4 AR-1242-4	5.629	4.932	19727	42666	28.213	75.226 #
20)	L4 AR-1242-5	0.000	5.441	0	84361	N.D.	107.494 #
21)	L5 AR-1248-1	5.495	4.636	38477	13882	68.598	30.915 #
22)	L5 AR-1248-2	0.000	4.877	0	46992	N.D.	74.734 #
23)	L5 AR-1248-3	0.000	4.932	0	42666	N.D.	66.126 #
24)	L5 AR-1248-4	6.296	5.084	2215220	10478	1998.583	13.293 #
25)	L5 AR-1248-5	0.000	5.441	0	84361	N.D.	69.630 #
26)	L6 AR-1254-1	6.296	5.441	2215220	84361	1357.261	42.924 #
27)	L6 AR-1254-2	6.548	5.579	67347	59244	25.835	33.081 #
28)	L6 AR-1254-3	6.912	5.978	67901	22901	23.525	7.400 #
29)	L6 AR-1254-4	7.188	6.204	27857	30687	12.164	14.873
30)	L6 AR-1254-5	7.608	6.622	13158	17729	5.149	5.878
31)	L7 AR-1260-1	7.067	6.114	12790	31632	6.777	18.022 #
32)	L7 AR-1260-2	7.319	6.293	86417	24220	36.951	10.887 #
33)	L7 AR-1260-3	7.631	6.449	6061	14183	3.274	6.976 #
34)	L7 AR-1260-4	7.923	6.923	61015	12443	27.841	6.709 #
35)	L7 AR-1260-5	8.215	7.160	18486	13665	4.187	2.823 #
36)	L8 AR-1262-1	7.631	0.000	6061	0	2.030	N.D. #
37)	L8 AR-1262-2	8.215	7.160	18486	13665	3.353	2.291 #
38)	L8 AR-1262-3	0.000	7.460	0	7430	N.D.	3.187 #
39)	L8 AR-1262-4	8.658	7.503	58050	43806	19.858	9.837 #
40)	L8 AR-1262-5	9.182	8.057	33650	14470	16.032	6.851 #
41)	L9 AR-1268-1	0.000	7.460	0	7430	N.D.	1.064 #
42)	L9 AR-1268-2	8.658	7.503	58050	43806	9.608	6.842 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0120219\
Data File : P0064282.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Dec 2019 23:48
Operator : SM/AJ
Sample : K5676-35
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SU-01-112719

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 03:39:18 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0112719.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 28 01:28:08 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
43)	L9 AR-1268-3	0.000	7.731	0	14317	N.D.	2.734 #
44)	L9 AR-1268-4	9.182	8.057	33650	14470	14.382	6.194 #
45)	L9 AR-1268-5	0.000	8.279	0	10296	N.D.	0.623 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120219\
Data File : PO064282.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Dec 2019 23:48
Operator : SM/AJ
Sample : K5676-35
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SU-01-112719

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 03:39:18 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112719.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 28 01:28:08 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.312 min
Delta R.T.: 0.000 min
Response: 807767
Conc: 27.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
SU-01-112719




#1 Tetrachloro-m-xylene

R.T.: 3.560 min
Delta R.T.: -0.001 min
Response: 688013
Conc: 28.93 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.940 min
Delta R.T.: -0.001 min
Response: 675496
Conc: 16.48 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.524 min
Delta R.T.: -0.002 min
Response: 603768
Conc: 14.78 ng/ml



#3 AR-1016-1


R.T.: 5.495 min
Delta R.T.: 0.020 min
Response: 38477
Conc: 29.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
SU-01-112719




#3 AR-1016-1

R.T.: 4.636 min
Delta R.T.: -0.003 min
Response: 13882
Conc: 13.26 ng/ml




#4 AR-1016-2

R.T.: 5.495 min
Delta R.T.: -0.002 min
Response: 38477
Conc: 21.11 ng/ml



#4 AR-1016-2


R.T.: 4.663 min
Delta R.T.: 0.005 min
Response: 19507
Conc: 13.83 ng/ml



#5 AR-1016-3


R.T.: 5.495 min
Delta R.T.: -0.063 min
Response: 38477
Conc: 33.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
SU-01-112719




#5 AR-1016-3

R.T.: 4.829 min
Delta R.T.: -0.003 min
Response: 22054
Conc: 28.63 ng/ml




#6 AR-1016-4

R.T.: 5.629 min
Delta R.T.: -0.027 min
Response: 19727
Conc: 20.98 ng/ml



#6 AR-1016-4


R.T.: 4.877 min
Delta R.T.: 0.005 min
Response: 46992
Conc: 72.24 ng/ml



#7 AR-1016-5


R.T.: 0.000 min
Exp R.T. : 5.948 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
SU-01-112719




#7 AR-1016-5

R.T.: 5.084 min
Delta R.T.: 0.000 min
Response: 10478
Conc: 12.59 ng/ml




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.521 min
Response: 0
Conc: N.D.



#8 AR-1221-1


R.T.: 3.754 min
Delta R.T.: -0.021 min
Response: 64838
Conc: 247.84 ng/ml



#9 AR-1221-2


R.T.: 4.617 min
Delta R.T.: 0.009 min
Response: 61006
Conc: 220.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
SU-01-112719




#9 AR-1221-2

R.T.: 3.858 min
Delta R.T.: -0.002 min
Response: 90138
Conc: 451.02 ng/ml




#10 AR-1221-3

R.T.: 4.696 min
Delta R.T.: 0.012 min
Response: 39088
Conc: 46.55 ng/ml



#10 AR-1221-3


R.T.: 3.937 min
Delta R.T.: 0.003 min
Response: 38290
Conc: 59.21 ng/ml



#11 AR-1232-1


R.T.: 4.696 min
Delta R.T.: 0.011 min
Response: 39088
Conc: 30.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
SU-01-112719




#11 AR-1232-1

R.T.: 3.937 min
Delta R.T.: 0.003 min
Response: 38290
Conc: 37.81 ng/ml




#12 AR-1232-2

R.T.: 5.217 min
Delta R.T.: 0.008 min
Response: 30561
Conc: 43.50 ng/ml



#12 AR-1232-2


R.T.: 4.663 min
Delta R.T.: 0.006 min
Response: 19507
Conc: 17.89 ng/ml



#13 AR-1232-3


R.T.: 5.495 min
Delta R.T.: -0.003 min
Response: 38477
Conc: 27.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
SU-01-112719




#13 AR-1232-3

R.T.: 4.829 min
Delta R.T.: -0.003 min
Response: 22054
Conc: 37.94 ng/ml




#14 AR-1232-4

R.T.: 5.629 min
Delta R.T.: -0.028 min
Response: 19727
Conc: 27.73 ng/ml



#14 AR-1232-4


R.T.: 4.932 min
Delta R.T.: 0.017 min
Response: 42666
Conc: 80.37 ng/ml



#15 AR-1232-5


R.T.: 0.000 min
Exp R.T.: 5.747 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
SU-01-112719




#15 AR-1232-5

R.T.: 5.084 min
Delta R.T.: 0.000 min
Response: 10478
Conc: 17.90 ng/ml




#16 AR-1242-1

R.T.: 5.495 min
Delta R.T.: 0.019 min
Response: 38477
Conc: 40.41 ng/ml



#16 AR-1242-1


R.T.: 4.636 min
Delta R.T.: -0.003 min
Response: 13882
Conc: 18.13 ng/ml



#17 AR-1242-2


R.T.: 5.495 min
Delta R.T.: -0.003 min
Response: 38477
Conc: 28.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
SU-01-112719




#17 AR-1242-2

R.T.: 4.663 min
Delta R.T.: 0.005 min
Response: 19507
Conc: 18.88 ng/ml




#18 AR-1242-3

R.T.: 5.495 min
Delta R.T.: -0.064 min
Response: 38477
Conc: 44.84 ng/ml



#18 AR-1242-3


R.T.: 4.829 min
Delta R.T.: -0.004 min
Response: 22054
Conc: 39.35 ng/ml



#19 AR-1242-4


R.T.: 5.629 min
Delta R.T.: -0.028 min
Response: 19727
Conc: 28.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
SU-01-112719




#19 AR-1242-4

R.T.: 4.932 min
Delta R.T.: 0.017 min
Response: 42666
Conc: 75.23 ng/ml




#20 AR-1242-5

R.T.: 0.000 min
Exp R.T. : 6.390 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 5.441 min
Delta R.T.: 0.007 min
Response: 84361
Conc: 107.49 ng/ml



#21 AR-1248-1


R.T.: 5.495 min
Delta R.T.: 0.020 min
Response: 38477
Conc: 68.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
SU-01-112719




#21 AR-1248-1

R.T.: 4.636 min
Delta R.T.: -0.002 min
Response: 13882
Conc: 30.91 ng/ml




#22 AR-1248-2

R.T.: 0.000 min
Exp R.T. : 5.746 min
Response: 0
Conc: N.D.



#22 AR-1248-2


R.T.: 4.877 min
Delta R.T.: 0.005 min
Response: 46992
Conc: 74.73 ng/ml



#23 AR-1248-3


R.T.: 0.000 min
Exp R.T.: 5.950 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
SU-01-112719




#23 AR-1248-3

R.T.: 4.932 min
Delta R.T.: 0.018 min
Response: 42666
Conc: 66.13 ng/ml




#24 AR-1248-4

R.T.: 6.296 min
Delta R.T.: -0.055 min
Response: 2215220
Conc: 1998.58 ng/ml



#24 AR-1248-4


R.T.: 5.084 min
Delta R.T.: 0.000 min
Response: 10478
Conc: 13.29 ng/ml



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T.: 6.389 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
SU-01-112719




#25 AR-1248-5

R.T.: 5.441 min
Delta R.T.: 0.007 min
Response: 84361
Conc: 69.63 ng/ml




#26 AR-1254-1

R.T.: 6.296 min
Delta R.T.: -0.029 min
Response: 2215220
Conc: 1357.26 ng/ml



#26 AR-1254-1


R.T.: 5.441 min
Delta R.T.: 0.006 min
Response: 84361
Conc: 42.92 ng/ml



#27 AR-1254-2


R.T.: 6.548 min
Delta R.T.: 0.005 min
Response: 67347
Conc: 25.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
SU-01-112719




#27 AR-1254-2

R.T.: 5.579 min
Delta R.T.: -0.001 min
Response: 59244
Conc: 33.08 ng/ml




#28 AR-1254-3

R.T.: 6.912 min
Delta R.T.: 0.004 min
Response: 67901
Conc: 23.53 ng/ml



#28 AR-1254-3


R.T.: 5.978 min
Delta R.T.: -0.005 min
Response: 22901
Conc: 7.40 ng/ml



#29 AR-1254-4


R.T.: 7.188 min
Delta R.T.: -0.004 min
Response: 27857
Conc: 12.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
SU-01-112719




#29 AR-1254-4

R.T.: 6.204 min
Delta R.T.: -0.006 min
Response: 30687
Conc: 14.87 ng/ml




#30 AR-1254-5

R.T.: 7.608 min
Delta R.T.: 0.000 min
Response: 13158
Conc: 5.15 ng/ml



#30 AR-1254-5


R.T.: 6.622 min
Delta R.T.: -0.004 min
Response: 17729
Conc: 5.88 ng/ml



#31 AR-1260-1


R.T.: 7.067 min
Delta R.T.: -0.002 min
Response: 12790
Conc: 6.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
SU-01-112719




#31 AR-1260-1

R.T.: 6.114 min
Delta R.T.: 0.002 min
Response: 31632
Conc: 18.02 ng/ml




#32 AR-1260-2

R.T.: 7.319 min
Delta R.T.: -0.007 min
Response: 86417
Conc: 36.95 ng/ml



#32 AR-1260-2


R.T.: 6.293 min
Delta R.T.: -0.004 min
Response: 24220
Conc: 10.89 ng/ml



#33 AR-1260-3


R.T.: 7.631 min
Delta R.T.: -0.052 min
Response: 6061
Conc: 3.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
SU-01-112719




#33 AR-1260-3

R.T.: 6.449 min
Delta R.T.: -0.002 min
Response: 14183
Conc: 6.98 ng/ml




#34 AR-1260-4

R.T.: 7.923 min
Delta R.T.: 0.016 min
Response: 61015
Conc: 27.84 ng/ml



#34 AR-1260-4


R.T.: 6.923 min
Delta R.T.: 0.001 min
Response: 12443
Conc: 6.71 ng/ml



#35 AR-1260-5


R.T.: 8.215 min
Delta R.T.: -0.004 min
Response: 18486
Conc: 4.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
SU-01-112719




#35 AR-1260-5

R.T.: 7.160 min
Delta R.T.: -0.001 min
Response: 13665
Conc: 2.82 ng/ml




#36 AR-1262-1

R.T.: 7.631 min
Delta R.T.: -0.052 min
Response: 6061
Conc: 2.03 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T. : 6.716 min
Response: 0
Conc: N.D.



#37 AR-1262-2


R.T.: 8.215 min
Delta R.T.: -0.005 min
Response: 18486
Conc: 3.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
SU-01-112719




#37 AR-1262-2

R.T.: 7.160 min
Delta R.T.: -0.003 min
Response: 13665
Conc: 2.29 ng/ml




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T. : 8.518 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 7.460 min
Delta R.T.: 0.015 min
Response: 7430
Conc: 3.19 ng/ml



#39 AR-1262-4


R.T.: 8.658 min
Delta R.T.: 0.055 min
Response: 58050
Conc: 19.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
SU-01-112719




#39 AR-1262-4

R.T.: 7.503 min
Delta R.T.: -0.006 min
Response: 43806
Conc: 9.84 ng/ml




#40 AR-1262-5

R.T.: 9.182 min
Delta R.T.: -0.049 min
Response: 33650
Conc: 16.03 ng/ml



#40 AR-1262-5


R.T.: 8.057 min
Delta R.T.: 0.056 min
Response: 14470
Conc: 6.85 ng/ml



#41 AR-1268-1


R.T.: 0.000 min
Exp R.T.: 8.516 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
SU-01-112719




#41 AR-1268-1

R.T.: 7.460 min
Delta R.T.: 0.015 min
Response: 7430
Conc: 1.06 ng/ml




#42 AR-1268-2

R.T.: 8.658 min
Delta R.T.: 0.052 min
Response: 58050
Conc: 9.61 ng/ml



#42 AR-1268-2


R.T.: 7.503 min
Delta R.T.: -0.007 min
Response: 43806
Conc: 6.84 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 8.823 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
SU-01-112719




#43 AR-1268-3

R.T.: 7.731 min
Delta R.T.: 0.015 min
Response: 14317
Conc: 2.73 ng/ml




#44 AR-1268-4

R.T.: 9.182 min
Delta R.T.: -0.051 min
Response: 33650
Conc: 14.38 ng/ml



#44 AR-1268-4


R.T.: 8.057 min
Delta R.T.: 0.056 min
Response: 14470
Conc: 6.19 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T. : 9.624 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
SU-01-112719



#45 AR-1268-5

R.T.: 8.279 min
Delta R.T.: -0.002 min
Response: 10296
Conc: 0.62 ng/ml