

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120221\
 Data File : PO083381.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Dec 2021 13:46
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 02 14:13:07 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Nov 20 02:49:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.948	3.949	4518281	1498198	51.485	49.150
2)	SA Decachlor...	11.038	9.255	3297646	1495923	50.150	49.762
Target Compounds							
26)	L6 AR-1254-1	7.198	6.078	1579684	923154	567.635	474.765
27)	L6 AR-1254-2	7.429	6.239	2287687	794251	549.901	472.054
28)	L6 AR-1254-3	7.814	6.659	2345626	1280671	546.120	500.943
29)	L6 AR-1254-4	8.112	6.900	1618814	896744	523.354	491.559
30)	L6 AR-1254-5	8.543	7.330	1685021	1114919	525.609	479.140



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120221\
Data File : PO083381.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Dec 2021 13:46
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 02 14:13:07 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111921.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Nov 20 02:49:51 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.948 min
Delta R.T.: -0.004 min
Response: 4518281
Conc: 51.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.949 min
Delta R.T.: -0.003 min
Response: 1498198
Conc: 49.15 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.038 min
Delta R.T.: -0.004 min
Response: 3297646
Conc: 50.15 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.255 min
Delta R.T.: -0.002 min
Response: 1495923
Conc: 49.76 ng/ml



#26 AR-1254-1


R.T.: 7.198 min
Delta R.T.: -0.002 min
Response: 1579684
Conc: 567.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.078 min
Delta R.T.: -0.003 min
Response: 923154
Conc: 474.76 ng/ml




#27 AR-1254-2

R.T.: 7.429 min
Delta R.T.: -0.002 min
Response: 2287687
Conc: 549.90 ng/ml



#27 AR-1254-2


R.T.: 6.239 min
Delta R.T.: -0.002 min
Response: 794251
Conc: 472.05 ng/ml



#28 AR-1254-3


R.T.: 7.814 min
Delta R.T.: -0.001 min
Response: 2345626
Conc: 546.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.659 min
Delta R.T.: -0.002 min
Response: 1280671
Conc: 500.94 ng/ml




#29 AR-1254-4

R.T.: 8.112 min
Delta R.T.: -0.001 min
Response: 1618814
Conc: 523.35 ng/ml



#29 AR-1254-4


R.T.: 6.900 min
Delta R.T.: -0.002 min
Response: 896744
Conc: 491.56 ng/ml



#30 AR-1254-5

R.T.: 8.543 min
Delta R.T.: -0.001 min
Response: 1685021
Conc: 525.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.330 min
Delta R.T.: -0.002 min
Response: 1114919
Conc: 479.14 ng/ml