

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120222\
 Data File : PO090992.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Dec 2022 18:56
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 02 21:51:29 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Nov 28 11:25:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.298	3.475	152.4E6	50406826	45.067	48.027
2) SA Decachlor...	10.043	8.477	118.6E6	43369522	44.636	44.671
Target Compounds						
16) L4 AR-1242-1	5.473	4.560	38837623	13157492	435.120	453.134
17) L4 AR-1242-2	5.496	4.579	55191005	18143484	434.682	450.097
18) L4 AR-1242-3	5.558	4.754	35168653	9847559	421.887	470.889
19) L4 AR-1242-4	5.657	4.839	27630937	10680017	434.956	467.721
20) L4 AR-1242-5	6.399	5.363	29383316	13175323	430.014	466.959



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120222\
Data File : PO090992.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Dec 2022 18:56
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 02 21:51:29 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Nov 28 11:25:59 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.298 min
Delta R.T.: 0.000 min
Response: 152377461
Conc: 45.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.475 min
Delta R.T.: -0.001 min
Response: 50406826
Conc: 48.03 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.043 min
Delta R.T.: -0.007 min
Response: 118560616
Conc: 44.64 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.477 min
Delta R.T.: -0.005 min
Response: 43369522
Conc: 44.67 ng/ml



#16 AR-1242-1


R.T.: 5.473 min
Delta R.T.: -0.003 min
Response: 38837623
Conc: 435.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.560 min
Delta R.T.: -0.002 min
Response: 13157492
Conc: 453.13 ng/ml




#17 AR-1242-2

R.T.: 5.496 min
Delta R.T.: -0.002 min
Response: 55191005
Conc: 434.68 ng/ml



#17 AR-1242-2


R.T.: 4.579 min
Delta R.T.: -0.002 min
Response: 18143484
Conc: 450.10 ng/ml



#18 AR-1242-3


R.T.: 5.558 min
Delta R.T.: -0.003 min
Response: 35168653
Conc: 421.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.754 min
Delta R.T.: -0.002 min
Response: 9847559
Conc: 470.89 ng/ml




#19 AR-1242-4

R.T.: 5.657 min
Delta R.T.: -0.003 min
Response: 27630937
Conc: 434.96 ng/ml



#19 AR-1242-4


R.T.: 4.839 min
Delta R.T.: -0.002 min
Response: 10680017
Conc: 467.72 ng/ml



#20 AR-1242-5

R.T.: 6.399 min
Delta R.T.: -0.003 min
Response: 29383316
Conc: 430.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.363 min
Delta R.T.: -0.003 min
Response: 13175323
Conc: 466.96 ng/ml