

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120321\
 Data File : PO083410.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Dec 2021 9:46
 Operator : AJ\MA
 Sample : M4899-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 01-A-01-B-01-C

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 03 16:03:55 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Nov 20 02:49:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.945	3.947	2634323	861508	30.018	28.263
2) SA Decachlor...	11.031	9.252	2008969	822239	30.552	27.352

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120321\
Data File : PO083410.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 Dec 2021 9:46
Operator : AJ\MA
Sample : M4899-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
01-A-01-B-01-C

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 03 16:03:55 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO111921.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Nov 20 02:49:51 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.945 min
Delta R.T.: -0.007 min
Response: 2634323
Conc: 30.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
01-A-01-B-01-C




#1 Tetrachloro-m-xylene

R.T.: 3.947 min
Delta R.T.: -0.005 min
Response: 861508
Conc: 28.26 ng/ml



#2 Decachlorobiphenyl

R.T.: 11.031 min
Delta R.T.: -0.011 min
Response: 2008969
Conc: 30.55 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.252 min
Delta R.T.: -0.006 min
Response: 822239
Conc: 27.35 ng/ml