

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO120419\
 Data File : PO064328.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Dec 2019 16:05
 Operator : SM/AJ
 Sample : K6100-03MSD
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 04 16:23:39 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO112719.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 28 01:28:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.310	3.558	665782	590412	22.481	24.823
2)	SA Decachlor...	9.938	8.522	821354	946705	20.041	23.175
Target Compounds							
3)	L1 AR-1016-1	5.472	4.635	350274	310690	269.826	296.740
4)	L1 AR-1016-2	5.494	4.653	494904	420362	271.461	297.989
5)	L1 AR-1016-3	5.555	4.829	307155	221421	267.617	287.476
6)	L1 AR-1016-4	5.653	4.869	245643	190330	261.204	292.581
7)	L1 AR-1016-5	5.946	5.081	258382	234900	265.521	282.156
8)	L2 AR-1221-1	4.513	3.769	29548	25807	83.174	98.646
9)	L2 AR-1221-2	4.605	3.856	46764	40198	169.136	201.135
10)	L2 AR-1221-3	4.681	3.931	164393	132440	195.780	204.812
11)	L3 AR-1232-1	4.681	3.931	164393	132440	128.224	130.778
12)	L3 AR-1232-2	5.205	4.653	241879	420362	344.297	385.582
13)	L3 AR-1232-3	5.494	4.829	494904	221421	357.377	380.869
14)	L3 AR-1232-4	5.653	4.910	245643	230548	345.255	434.272 #
15)	L3 AR-1232-5	5.743	5.081	232134	234900	412.160	401.183
16)	L4 AR-1242-1	5.472	4.635	350274	310690	367.860	405.748
17)	L4 AR-1242-2	5.494	4.653	494904	420362	366.434	406.867
18)	L4 AR-1242-3	5.555	4.829	307155	221421	357.938	395.099
19)	L4 AR-1242-4	5.653	4.910	245643	230548	351.314	406.485
20)	L4 AR-1242-5	6.386	5.430	51248	187351	60.662	238.724 #
21)	L5 AR-1248-1	5.472	4.635	350274	310690	624.480	691.898
22)	L5 AR-1248-2	5.743	4.869	232134	190330	280.118	302.691
23)	L5 AR-1248-3	5.946	4.910	258382	230548	280.889	357.313 #
24)	L5 AR-1248-4	6.348	5.081	59895	234900	54.038	297.999 #
25)	L5 AR-1248-5	6.386	5.430	51248	187351	47.763	154.636 #
26)	L6 AR-1254-1	6.320	5.430	189222	187351	115.936	95.326
27)	L6 AR-1254-2	6.537	5.576	213376	179831	81.852	100.413
28)	L6 AR-1254-3	6.904	5.993	132849	363341	46.028	117.403 #
29)	L6 AR-1254-4	7.188	6.206	98689	61097	43.093	29.612 #
30)	L6 AR-1254-5	7.605	6.621	703837	710193	275.447	235.460
31)	L7 AR-1260-1	7.066	6.108	486606	504690	257.850	287.537
32)	L7 AR-1260-2	7.323	6.294	598081	643190	255.732	289.103
33)	L7 AR-1260-3	7.680	6.448	462252	581761	249.690	286.128
34)	L7 AR-1260-4	7.904	6.918	538488	518514	245.707	279.559
35)	L7 AR-1260-5	8.215	7.159	1010517	1273829	228.884	263.111
36)	L8 AR-1262-1	7.680	6.712	462252	251066	154.825	199.809 #
37)	L8 AR-1262-2	8.215	7.159	1010517	1273829	183.273	213.540
38)	L8 AR-1262-3	8.529	7.441	685395	292415	177.139	125.433 #
39)	L8 AR-1262-4	8.582	7.504	439473	936610	150.336	210.328 #
40)	L8 AR-1262-5	9.228	7.997	334454	337839	159.348	159.959
41)	L9 AR-1268-1	8.529	7.441	685395	292415	105.124	41.882 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120419\
Data File : PO064328.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Dec 2019 16:05
Operator : SM/AJ
Sample : K6100-03MSD
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 04 16:23:39 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112719.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 28 01:28:08 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.582	7.504	439473	936610	72.742	146.287 #
43)	L9 AR-1268-3	8.818	7.713	26982	19160	5.434	3.658 #
44)	L9 AR-1268-4	9.228	7.997	334454	337839	142.949	144.611
45)	L9 AR-1268-5	9.620	8.277	97059	91463	6.318	5.538



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120419\
Data File : P0064328.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Dec 2019 16:05
Operator : SM/AJ
Sample : K6100-03MSD
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 04 16:23:39 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112719.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 28 01:28:08 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.310 min
Delta R.T.: -0.003 min
Response: 665782
Conc: 22.48 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.558 min
Delta R.T.: -0.003 min
Response: 590412
Conc: 24.82 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.938 min
Delta R.T.: -0.003 min
Response: 821354
Conc: 20.04 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.522 min
Delta R.T.: -0.005 min
Response: 946705
Conc: 23.17 ng/ml



#3 AR-1016-1


R.T.: 5.472 min
Delta R.T.: -0.003 min
Response: 350274
Conc: 269.83 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.635 min
Delta R.T.: -0.004 min
Response: 310690
Conc: 296.74 ng/ml




#4 AR-1016-2

R.T.: 5.494 min
Delta R.T.: -0.003 min
Response: 494904
Conc: 271.46 ng/ml



#4 AR-1016-2


R.T.: 4.653 min
Delta R.T.: -0.004 min
Response: 420362
Conc: 297.99 ng/ml



#5 AR-1016-3


R.T.: 5.555 min
Delta R.T.: -0.003 min
Response: 307155
Conc: 267.62 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.829 min
Delta R.T.: -0.003 min
Response: 221421
Conc: 287.48 ng/ml




#6 AR-1016-4

R.T.: 5.653 min
Delta R.T.: -0.003 min
Response: 245643
Conc: 261.20 ng/ml



#6 AR-1016-4


R.T.: 4.869 min
Delta R.T.: -0.003 min
Response: 190330
Conc: 292.58 ng/ml



#7 AR-1016-5


R.T.: 5.946 min
Delta R.T.: -0.002 min
Response: 258382
Conc: 265.52 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.081 min
Delta R.T.: -0.003 min
Response: 234900
Conc: 282.16 ng/ml




#8 AR-1221-1

R.T.: 4.513 min
Delta R.T.: -0.008 min
Response: 29548
Conc: 83.17 ng/ml



#8 AR-1221-1


R.T.: 3.769 min
Delta R.T.: -0.005 min
Response: 25807
Conc: 98.65 ng/ml



#9 AR-1221-2


R.T.: 4.605 min
Delta R.T.: -0.003 min
Response: 46764
Conc: 169.14 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.856 min
Delta R.T.: -0.003 min
Response: 40198
Conc: 201.13 ng/ml




#10 AR-1221-3

R.T.: 4.681 min
Delta R.T.: -0.003 min
Response: 164393
Conc: 195.78 ng/ml



#10 AR-1221-3


R.T.: 3.931 min
Delta R.T.: -0.004 min
Response: 132440
Conc: 204.81 ng/ml



#11 AR-1232-1


R.T.: 4.681 min
Delta R.T.: -0.004 min
Response: 164393
Conc: 128.22 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 3.931 min
Delta R.T.: -0.004 min
Response: 132440
Conc: 130.78 ng/ml




#12 AR-1232-2

R.T.: 5.205 min
Delta R.T.: -0.004 min
Response: 241879
Conc: 344.30 ng/ml



#12 AR-1232-2


R.T.: 4.653 min
Delta R.T.: -0.003 min
Response: 420362
Conc: 385.58 ng/ml



#13 AR-1232-3


R.T.: 5.494 min
Delta R.T.: -0.005 min
Response: 494904
Conc: 357.38 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.829 min
Delta R.T.: -0.003 min
Response: 221421
Conc: 380.87 ng/ml




#14 AR-1232-4

R.T.: 5.653 min
Delta R.T.: -0.004 min
Response: 245643
Conc: 345.26 ng/ml



#14 AR-1232-4


R.T.: 4.910 min
Delta R.T.: -0.004 min
Response: 230548
Conc: 434.27 ng/ml



#15 AR-1232-5


R.T.: 5.743 min
Delta R.T.: -0.004 min
Response: 232134
Conc: 412.16 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.081 min
Delta R.T.: -0.004 min
Response: 234900
Conc: 401.18 ng/ml




#16 AR-1242-1

R.T.: 5.472 min
Delta R.T.: -0.004 min
Response: 350274
Conc: 367.86 ng/ml



#16 AR-1242-1


R.T.: 4.635 min
Delta R.T.: -0.004 min
Response: 310690
Conc: 405.75 ng/ml



#17 AR-1242-2


R.T.: 5.494 min
Delta R.T.: -0.004 min
Response: 494904
Conc: 366.43 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.653 min
Delta R.T.: -0.004 min
Response: 420362
Conc: 406.87 ng/ml




#18 AR-1242-3

R.T.: 5.555 min
Delta R.T.: -0.004 min
Response: 307155
Conc: 357.94 ng/ml



#18 AR-1242-3


R.T.: 4.829 min
Delta R.T.: -0.004 min
Response: 221421
Conc: 395.10 ng/ml



#19 AR-1242-4


R.T.: 5.653 min
Delta R.T.: -0.005 min
Response: 245643
Conc: 351.31 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 4.910 min
Delta R.T.: -0.004 min
Response: 230548
Conc: 406.49 ng/ml




#20 AR-1242-5

R.T.: 6.386 min
Delta R.T.: -0.003 min
Response: 51248
Conc: 60.66 ng/ml



#20 AR-1242-5


R.T.: 5.430 min
Delta R.T.: -0.004 min
Response: 187351
Conc: 238.72 ng/ml



#21 AR-1248-1


R.T.: 5.472 min
Delta R.T.: -0.004 min
Response: 350274
Conc: 624.48 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.635 min
Delta R.T.: -0.003 min
Response: 310690
Conc: 691.90 ng/ml




#22 AR-1248-2

R.T.: 5.743 min
Delta R.T.: -0.003 min
Response: 232134
Conc: 280.12 ng/ml



#22 AR-1248-2


R.T.: 4.869 min
Delta R.T.: -0.003 min
Response: 190330
Conc: 302.69 ng/ml



#23 AR-1248-3


R.T.: 5.946 min
Delta R.T.: -0.004 min
Response: 258382
Conc: 280.89 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 4.910 min
Delta R.T.: -0.003 min
Response: 230548
Conc: 357.31 ng/ml




#24 AR-1248-4

R.T.: 6.348 min
Delta R.T.: -0.003 min
Response: 59895
Conc: 54.04 ng/ml



#24 AR-1248-4


R.T.: 5.081 min
Delta R.T.: -0.004 min
Response: 234900
Conc: 298.00 ng/ml



#25 AR-1248-5


R.T.: 6.386 min
Delta R.T.: -0.003 min
Response: 51248
Conc: 47.76 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.430 min
Delta R.T.: -0.004 min
Response: 187351
Conc: 154.64 ng/ml




#26 AR-1254-1

R.T.: 6.320 min
Delta R.T.: -0.005 min
Response: 189222
Conc: 115.94 ng/ml



#26 AR-1254-1


R.T.: 5.430 min
Delta R.T.: -0.005 min
Response: 187351
Conc: 95.33 ng/ml



#27 AR-1254-2


R.T.: 6.537 min
Delta R.T.: -0.005 min
Response: 213376
Conc: 81.85 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.576 min
Delta R.T.: -0.004 min
Response: 179831
Conc: 100.41 ng/ml




#28 AR-1254-3

R.T.: 6.904 min
Delta R.T.: -0.004 min
Response: 132849
Conc: 46.03 ng/ml



#28 AR-1254-3


R.T.: 5.993 min
Delta R.T.: 0.010 min
Response: 363341
Conc: 117.40 ng/ml



#29 AR-1254-4


R.T.: 7.188 min
Delta R.T.: -0.005 min
Response: 98689
Conc: 43.09 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.206 min
Delta R.T.: -0.004 min
Response: 61097
Conc: 29.61 ng/ml




#30 AR-1254-5

R.T.: 7.605 min
Delta R.T.: -0.004 min
Response: 703837
Conc: 275.45 ng/ml



#30 AR-1254-5


R.T.: 6.621 min
Delta R.T.: -0.005 min
Response: 710193
Conc: 235.46 ng/ml



#31 AR-1260-1


R.T.: 7.066 min
Delta R.T.: -0.003 min
Response: 486606
Conc: 257.85 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.108 min
Delta R.T.: -0.004 min
Response: 504690
Conc: 287.54 ng/ml




#32 AR-1260-2

R.T.: 7.323 min
Delta R.T.: -0.003 min
Response: 598081
Conc: 255.73 ng/ml



#32 AR-1260-2


R.T.: 6.294 min
Delta R.T.: -0.003 min
Response: 643190
Conc: 289.10 ng/ml



#33 AR-1260-3


R.T.: 7.680 min
Delta R.T.: -0.003 min
Response: 462252
Conc: 249.69 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.448 min
Delta R.T.: -0.003 min
Response: 581761
Conc: 286.13 ng/ml




#34 AR-1260-4

R.T.: 7.904 min
Delta R.T.: -0.003 min
Response: 538488
Conc: 245.71 ng/ml



#34 AR-1260-4


R.T.: 6.918 min
Delta R.T.: -0.004 min
Response: 518514
Conc: 279.56 ng/ml



#35 AR-1260-5


R.T.: 8.215 min
Delta R.T.: -0.004 min
Response: 1010517
Conc: 228.88 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.159 min
Delta R.T.: -0.003 min
Response: 1273829
Conc: 263.11 ng/ml




#36 AR-1262-1

R.T.: 7.680 min
Delta R.T.: -0.003 min
Response: 462252
Conc: 154.82 ng/ml



#36 AR-1262-1


R.T.: 6.712 min
Delta R.T.: -0.004 min
Response: 251066
Conc: 199.81 ng/ml



#37 AR-1262-2


R.T.: 8.215 min
Delta R.T.: -0.004 min
Response: 1010517
Conc: 183.27 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 7.159 min
Delta R.T.: -0.004 min
Response: 1273829
Conc: 213.54 ng/ml




#38 AR-1262-3

R.T.: 8.529 min
Delta R.T.: 0.011 min
Response: 685395
Conc: 177.14 ng/ml



#38 AR-1262-3


R.T.: 7.441 min
Delta R.T.: -0.004 min
Response: 292415
Conc: 125.43 ng/ml



#39 AR-1262-4


R.T.: 8.582 min
Delta R.T.: -0.021 min
Response: 439473
Conc: 150.34 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.504 min
Delta R.T.: -0.005 min
Response: 936610
Conc: 210.33 ng/ml




#40 AR-1262-5

R.T.: 9.228 min
Delta R.T.: -0.004 min
Response: 334454
Conc: 159.35 ng/ml



#40 AR-1262-5


R.T.: 7.997 min
Delta R.T.: -0.004 min
Response: 337839
Conc: 159.96 ng/ml



#41 AR-1268-1


R.T.: 8.529 min
Delta R.T.: 0.013 min
Response: 685395
Conc: 105.12 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.441 min
Delta R.T.: -0.004 min
Response: 292415
Conc: 41.88 ng/ml




#42 AR-1268-2

R.T.: 8.582 min
Delta R.T.: -0.024 min
Response: 439473
Conc: 72.74 ng/ml



#42 AR-1268-2


R.T.: 7.504 min
Delta R.T.: -0.006 min
Response: 936610
Conc: 146.29 ng/ml



#43 AR-1268-3


R.T.: 8.818 min
Delta R.T.: -0.005 min
Response: 26982
Conc: 5.43 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.713 min
Delta R.T.: -0.003 min
Response: 19160
Conc: 3.66 ng/ml




#44 AR-1268-4

R.T.: 9.228 min
Delta R.T.: -0.005 min
Response: 334454
Conc: 142.95 ng/ml



#44 AR-1268-4


R.T.: 7.997 min
Delta R.T.: -0.004 min
Response: 337839
Conc: 144.61 ng/ml



#45 AR-1268-5

R.T.: 9.620 min
Delta R.T.: -0.005 min
Response: 97059
Conc: 6.32 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.277 min
Delta R.T.: -0.004 min
Response: 91463
Conc: 5.54 ng/ml