

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO120423\
 Data File : PO100236.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Dec 2023 09:13
 Operator : YP/AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 04 14:37:36 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO112123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 15:58:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.469	3.733f	30133	21933	0.018	0.015
2)	SA Decachlor...	10.360	0.000	38303	0	0.053	N.D. #
Target Compounds							
3)	L1 AR-1016-1	5.677	4.809	22504	106013	0.523	3.357 #
4)	L1 AR-1016-2	5.685	4.838	64449	21786	0.995	0.504 #
5)	L1 AR-1016-3	5.752	5.015	44395	75822	1.019	3.269 #
6)	L1 AR-1016-4	5.847	5.050	68110	34006	2.063	1.659
7)	L1 AR-1016-5	6.155	5.265	50318	19121	1.603	0.750 #
8)	L2 AR-1221-1	4.706	3.931	30847	81229	1.496	4.273 #
9)	L2 AR-1221-2	4.786	4.028	80587	18408	5.335	1.413 #
10)	L2 AR-1221-3	4.845	4.087	118576	21940	2.642	0.611 #
11)	L3 AR-1232-1	4.845	4.087	118576	21940	3.323	0.748 #
12)	L3 AR-1232-2	5.401	4.838	38204	21786	2.195	1.099 #
13)	L3 AR-1232-3	5.685	5.015	64449	75822	2.274	7.316 #
14)	L3 AR-1232-4	5.847	5.106	68110	47748	4.903	4.804
15)	L3 AR-1232-5	5.934	5.265	77976	19121	6.354	1.770 #
16)	L4 AR-1242-1	5.677	4.809	22504	106013	0.650	4.034 #
17)	L4 AR-1242-2	5.685	4.838	64449	21786	1.228	0.620 #
18)	L4 AR-1242-3	5.752	5.015	44395	75822	1.276	3.959 #
19)	L4 AR-1242-4	5.847	5.106	68110	47748	2.591	2.371
20)	L4 AR-1242-5	6.598	5.594	63692	48315	2.849	2.068 #
21)	L5 AR-1248-1	5.654	4.809	155861	106013	6.029	5.303
22)	L5 AR-1248-2	5.934	5.050	77976	34006	1.907	1.215 #
23)	L5 AR-1248-3	6.155	5.106	50318	47748	1.175	1.653 #
24)	L5 AR-1248-4	6.563	5.265	100125	19121	2.656	0.583 #
25)	L5 AR-1248-5	6.598	5.715f	63692	508723	1.725	17.030 #
26)	L6 AR-1254-1	6.540	5.594	64777	48315	1.345	0.932 #
27)	L6 AR-1254-2	6.759	5.773	55717	221604	0.812	4.759 #
28)	L6 AR-1254-3	7.129	6.183	333429	911347	5.234	11.685 #
29)	L6 AR-1254-4	7.418	6.411	108022	444876	2.857	10.424 #
30)	L6 AR-1254-5	7.839	6.833	77673	344105	1.815	4.755 #
31)	L7 AR-1260-1	7.297	0.000	223262	0	4.193	N.D. #
32)	L7 AR-1260-2	7.549	6.488	133120	301944	2.517	4.546 #
33)	L7 AR-1260-3	7.937	6.614f	238816	59474	6.695	0.929 #
34)	L7 AR-1260-4	8.139	7.170	96727	437503	2.337	8.676 #
35)	L7 AR-1260-5	8.482	0.000	257403	0	3.853	N.D. #
36)	L8 AR-1262-1	7.839	6.833	77673	344105	2.147	9.003 #
37)	L8 AR-1262-2	8.482	7.170	257403	437503	3.032	5.800 #
38)	L8 AR-1262-3	8.807	7.656	127262	52848	2.113	0.907 #
39)	L8 AR-1262-4	8.898	0.000	82645	0	1.714	N.D. #
40)	L8 AR-1262-5	9.581	8.297f	61179	35699	2.343	0.819 #
41)	L9 AR-1268-1	8.807	7.656	127262	52848	1.153	0.318 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120423\
 Data File : PO100236.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Dec 2023 09:13
 Operator : YP/AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 04 14:37:36 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 15:58:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.898	0.000	82645	0	0.858	N.D. #
43) L9	AR-1268-3	9.142	7.956	12600	476848	0.140	3.327 #
44) L9	AR-1268-4	9.581	8.297f	61179	35699	2.151	0.760 #
45) L9	AR-1268-5	10.008	8.488f	79853	65978	0.317	0.166 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120423\
Data File : PO100236.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Dec 2023 09:13
Operator : YP/AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 04 14:37:36 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 15:58:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.469 min
Delta R.T.: -0.011 min
Response: 30133
Conc: 0.02 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.733 min
Delta R.T.: 0.027 min
Response: 21933
Conc: 0.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.360 min
Delta R.T.: 0.003 min
Response: 38303
Conc: 0.05 ng/ml



#2 Decachlorobiphenyl


R.T.: 0.000 min
Exp R.T.: 8.801 min
Response: 0
Conc: N.D.



#3 AR-1016-1


R.T.: 5.677 min
Delta R.T.: 0.010 min
Response: 22504
Conc: 0.52 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.809 min
Delta R.T.: -0.003 min
Response: 106013
Conc: 3.36 ng/ml




#4 AR-1016-2

R.T.: 5.685 min
Delta R.T.: -0.005 min
Response: 64449
Conc: 1.00 ng/ml



#4 AR-1016-2


R.T.: 4.838 min
Delta R.T.: 0.007 min
Response: 21786
Conc: 0.50 ng/ml



#5 AR-1016-3


R.T.: 5.752 min
Delta R.T.: -0.001 min
Response: 44395
Conc: 1.02 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 5.015 min
Delta R.T.: 0.005 min
Response: 75822
Conc: 3.27 ng/ml




#6 AR-1016-4

R.T.: 5.847 min
Delta R.T.: -0.005 min
Response: 68110
Conc: 2.06 ng/ml



#6 AR-1016-4


R.T.: 5.050 min
Delta R.T.: -0.001 min
Response: 34006
Conc: 1.66 ng/ml



#7 AR-1016-5


R.T.: 6.155 min
Delta R.T.: 0.004 min
Response: 50318
Conc: 1.60 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.265 min
Delta R.T.: -0.004 min
Response: 19121
Conc: 0.75 ng/ml




#8 AR-1221-1

R.T.: 4.706 min
Delta R.T.: 0.014 min
Response: 30847
Conc: 1.50 ng/ml



#8 AR-1221-1


R.T.: 3.931 min
Delta R.T.: 0.008 min
Response: 81229
Conc: 4.27 ng/ml



#9 AR-1221-2


R.T.: 4.786 min
Delta R.T.: 0.006 min
Response: 80587
Conc: 5.34 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 4.028 min
Delta R.T.: 0.017 min
Response: 18408
Conc: 1.41 ng/ml




#10 AR-1221-3

R.T.: 4.845 min
Delta R.T.: -0.012 min
Response: 118576
Conc: 2.64 ng/ml



#10 AR-1221-3


R.T.: 4.087 min
Delta R.T.: -0.001 min
Response: 21940
Conc: 0.61 ng/ml



#11 AR-1232-1


R.T.: 4.845 min
Delta R.T.: -0.012 min
Response: 118576
Conc: 3.32 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 4.087 min
Delta R.T.: 0.000 min
Response: 21940
Conc: 0.75 ng/ml




#12 AR-1232-2

R.T.: 5.401 min
Delta R.T.: 0.006 min
Response: 38204
Conc: 2.20 ng/ml



#12 AR-1232-2


R.T.: 4.838 min
Delta R.T.: 0.007 min
Response: 21786
Conc: 1.10 ng/ml



#13 AR-1232-3


R.T.: 5.685 min
Delta R.T.: -0.004 min
Response: 64449
Conc: 2.27 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 5.015 min
Delta R.T.: 0.005 min
Response: 75822
Conc: 7.32 ng/ml




#14 AR-1232-4

R.T.: 5.847 min
Delta R.T.: -0.004 min
Response: 68110
Conc: 4.90 ng/ml



#14 AR-1232-4


R.T.: 5.106 min
Delta R.T.: 0.012 min
Response: 47748
Conc: 4.80 ng/ml



#15 AR-1232-5


R.T.: 5.934 min
Delta R.T.: -0.010 min
Response: 77976
Conc: 6.35 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.265 min
Delta R.T.: -0.004 min
Response: 19121
Conc: 1.77 ng/ml




#16 AR-1242-1

R.T.: 5.677 min
Delta R.T.: 0.010 min
Response: 22504
Conc: 0.65 ng/ml



#16 AR-1242-1


R.T.: 4.809 min
Delta R.T.: -0.003 min
Response: 106013
Conc: 4.03 ng/ml



#17 AR-1242-2


R.T.: 5.685 min
Delta R.T.: -0.005 min
Response: 64449
Conc: 1.23 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.838 min
Delta R.T.: 0.006 min
Response: 21786
Conc: 0.62 ng/ml




#18 AR-1242-3

R.T.: 5.752 min
Delta R.T.: -0.002 min
Response: 44395
Conc: 1.28 ng/ml



#18 AR-1242-3


R.T.: 5.015 min
Delta R.T.: 0.004 min
Response: 75822
Conc: 3.96 ng/ml



#19 AR-1242-4


R.T.: 5.847 min
Delta R.T.: -0.006 min
Response: 68110
Conc: 2.59 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 5.106 min
Delta R.T.: 0.011 min
Response: 47748
Conc: 2.37 ng/ml




#20 AR-1242-5

R.T.: 6.598 min
Delta R.T.: -0.002 min
Response: 63692
Conc: 2.85 ng/ml



#20 AR-1242-5


R.T.: 5.594 min
Delta R.T.: -0.031 min
Response: 48315
Conc: 2.07 ng/ml



#21 AR-1248-1


R.T.: 5.654 min
Delta R.T.: -0.011 min
Response: 155861
Conc: 6.03 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.809 min
Delta R.T.: -0.002 min
Response: 106013
Conc: 5.30 ng/ml




#22 AR-1248-2

R.T.: 5.934 min
Delta R.T.: -0.010 min
Response: 77976
Conc: 1.91 ng/ml



#22 AR-1248-2


R.T.: 5.050 min
Delta R.T.: 0.000 min
Response: 34006
Conc: 1.22 ng/ml



#23 AR-1248-3


R.T.: 6.155 min
Delta R.T.: 0.004 min
Response: 50318
Conc: 1.17 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 5.106 min
Delta R.T.: 0.012 min
Response: 47748
Conc: 1.65 ng/ml




#24 AR-1248-4

R.T.: 6.563 min
Delta R.T.: 0.003 min
Response: 100125
Conc: 2.66 ng/ml



#24 AR-1248-4


R.T.: 5.265 min
Delta R.T.: -0.004 min
Response: 19121
Conc: 0.58 ng/ml



#25 AR-1248-5


R.T.: 6.598 min
Delta R.T.: 0.000 min
Response: 63692
Conc: 1.72 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.715 min
Delta R.T.: 0.049 min
Response: 508723
Conc: 17.03 ng/ml




#26 AR-1254-1

R.T.: 6.540 min
Delta R.T.: 0.007 min
Response: 64777
Conc: 1.35 ng/ml



#26 AR-1254-1


R.T.: 5.594 min
Delta R.T.: -0.032 min
Response: 48315
Conc: 0.93 ng/ml



#27 AR-1254-2


R.T.: 6.759 min
Delta R.T.: 0.004 min
Response: 55717
Conc: 0.81 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.773 min
Delta R.T.: 0.000 min
Response: 221604
Conc: 4.76 ng/ml




#28 AR-1254-3

R.T.: 7.129 min
Delta R.T.: 0.003 min
Response: 333429
Conc: 5.23 ng/ml



#28 AR-1254-3


R.T.: 6.183 min
Delta R.T.: 0.000 min
Response: 911347
Conc: 11.69 ng/ml



#29 AR-1254-4


R.T.: 7.418 min
Delta R.T.: 0.004 min
Response: 108022
Conc: 2.86 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.411 min
Delta R.T.: -0.001 min
Response: 444876
Conc: 10.42 ng/ml




#30 AR-1254-5

R.T.: 7.839 min
Delta R.T.: 0.000 min
Response: 77673
Conc: 1.81 ng/ml



#30 AR-1254-5


R.T.: 6.833 min
Delta R.T.: -0.002 min
Response: 344105
Conc: 4.76 ng/ml



#31 AR-1260-1


R.T.: 7.297 min
Delta R.T.: 0.005 min
Response: 223262
Conc: 4.19 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 0.000 min
Exp R.T. : 6.315 min
Response: 0
Conc: N.D.




#32 AR-1260-2

R.T.: 7.549 min
Delta R.T.: -0.002 min
Response: 133120
Conc: 2.52 ng/ml



#32 AR-1260-2


R.T.: 6.488 min
Delta R.T.: -0.016 min
Response: 301944
Conc: 4.55 ng/ml



#33 AR-1260-3


R.T.: 7.937 min
Delta R.T.: 0.022 min
Response: 238816
Conc: 6.70 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.614 min
Delta R.T.: -0.046 min
Response: 59474
Conc: 0.93 ng/ml




#34 AR-1260-4

R.T.: 8.139 min
Delta R.T.: -0.005 min
Response: 96727
Conc: 2.34 ng/ml



#34 AR-1260-4


R.T.: 7.170 min
Delta R.T.: 0.034 min
Response: 437503
Conc: 8.68 ng/ml



#35 AR-1260-5


R.T.: 8.482 min
Delta R.T.: 0.010 min
Response: 257403
Conc: 3.85 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 0.000 min
Exp R.T. : 7.378 min
Response: 0
Conc: N.D.




#36 AR-1262-1

R.T.: 7.839 min
Delta R.T.: 0.005 min
Response: 77673
Conc: 2.15 ng/ml



#36 AR-1262-1


R.T.: 6.833 min
Delta R.T.: 0.000 min
Response: 344105
Conc: 9.00 ng/ml



#37 AR-1262-2


R.T.: 8.482 min
Delta R.T.: 0.010 min
Response: 257403
Conc: 3.03 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 7.170 min
Delta R.T.: 0.035 min
Response: 437503
Conc: 5.80 ng/ml




#38 AR-1262-3

R.T.: 8.807 min
Delta R.T.: 0.011 min
Response: 127262
Conc: 2.11 ng/ml



#38 AR-1262-3


R.T.: 7.656 min
Delta R.T.: -0.008 min
Response: 52848
Conc: 0.91 ng/ml



#39 AR-1262-4


R.T.: 8.898 min
Delta R.T.: 0.008 min
Response: 82645
Conc: 1.71 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 0.000 min
Exp R.T. : 7.728 min
Response: 0
Conc: N.D.




#40 AR-1262-5

R.T.: 9.581 min
Delta R.T.: 0.012 min
Response: 61179
Conc: 2.34 ng/ml



#40 AR-1262-5


R.T.: 8.297 min
Delta R.T.: 0.068 min
Response: 35699
Conc: 0.82 ng/ml



#41 AR-1268-1


R.T.: 8.807 min
Delta R.T.: 0.013 min
Response: 127262
Conc: 1.15 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.656 min
Delta R.T.: -0.008 min
Response: 52848
Conc: 0.32 ng/ml




#42 AR-1268-2

R.T.: 8.898 min
Delta R.T.: 0.006 min
Response: 82645
Conc: 0.86 ng/ml



#42 AR-1268-2


R.T.: 0.000 min
Exp R.T.: 7.728 min
Response: 0
Conc: N.D.



#43 AR-1268-3


R.T.: 9.142 min
Delta R.T.: 0.013 min
Response: 12600
Conc: 0.14 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.956 min
Delta R.T.: 0.018 min
Response: 476848
Conc: 3.33 ng/ml




#44 AR-1268-4

R.T.: 9.581 min
Delta R.T.: 0.012 min
Response: 61179
Conc: 2.15 ng/ml



#44 AR-1268-4


R.T.: 8.297 min
Delta R.T.: 0.068 min
Response: 35699
Conc: 0.76 ng/ml



#45 AR-1268-5

R.T.: 10.008 min
Delta R.T.: 0.004 min
Response: 79853
Conc: 0.32 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.488 min
Delta R.T.: -0.045 min
Response: 65978
Conc: 0.17 ng/ml