

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120423\
 Data File : PO100250.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Dec 2023 15:25
 Operator : YP/AJ
 Sample : 05599-11
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 04 16:48:03 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 15:58:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.480	3.703	35319455	32546488	20.693	21.699
2)	SA Decachlor...	10.362	8.795	15679278	23424917	21.586	20.934
Target Compounds							
3)	L1 AR-1016-1	5.671	4.809	84850	61653	1.972	1.953
4)	L1 AR-1016-2	5.691	4.835	182185	45441	2.814	1.051 #
5)	L1 AR-1016-3	5.749	5.007	156921	275305	3.600	11.871 #
6)	L1 AR-1016-4	5.862	5.046	60221	68517	1.824	3.342 #
7)	L1 AR-1016-5	6.149	5.254	27957	142924	0.890	5.604 #
8)	L2 AR-1221-1	4.689	3.926	1950256	744699	94.602	39.174 #
9)	L2 AR-1221-2	4.791	3.991	592509	2963400	39.228	227.405 #
10)	L2 AR-1221-3	4.857	4.094	90526	544703	2.017	15.164 #
11)	L3 AR-1232-1	4.857	4.094	90526	544703	2.537	18.562 #
12)	L3 AR-1232-2	5.419	4.835	347546	45441	19.969	2.291 #
13)	L3 AR-1232-3	5.691	5.007	182185	275305	6.429	26.565 #
14)	L3 AR-1232-4	5.862	5.098	60221	114784	4.335	11.548 #
15)	L3 AR-1232-5	5.943	5.254	66514	142924	5.420	13.234 #
16)	L4 AR-1242-1	5.671	4.809	84850	61653	2.449	2.346
17)	L4 AR-1242-2	5.691	4.835	182185	45441	3.472	1.292 #
18)	L4 AR-1242-3	5.749	5.007	156921	275305	4.509	14.376 #
19)	L4 AR-1242-4	5.862	5.098	60221	114784	2.291	5.699 #
20)	L4 AR-1242-5	6.588	5.609	119159	330016	5.329	14.124 #
21)	L5 AR-1248-1	5.671	4.809	84850	61653	3.282	3.084
22)	L5 AR-1248-2	5.943	5.046	66514	68517	1.626	2.449 #
23)	L5 AR-1248-3	6.149	5.098	27957	114784	0.653	3.975 #
24)	L5 AR-1248-4	6.535	5.254	321021	142924	8.517	4.359 #
25)	L5 AR-1248-5	6.588	5.681	119159	102117	3.226	3.419
26)	L6 AR-1254-1	6.535	5.609	321021	330016	6.667	6.366
27)	L6 AR-1254-2	6.756	5.774	75686	86390	1.104	1.855 #
28)	L6 AR-1254-3	7.128	6.193	166524	356784	2.614	4.575 #
29)	L6 AR-1254-4	7.415	6.415	40607	45066	1.074	1.056
30)	L6 AR-1254-5	7.844	6.840	86861	130039	2.029	1.797
31)	L7 AR-1260-1	7.294	6.347	49497	607119	0.930	10.507 #
32)	L7 AR-1260-2	7.540	6.484	82376	338242	1.557	5.093 #
33)	L7 AR-1260-3	7.911	6.655	32993	19499	0.925	0.305 #
34)	L7 AR-1260-4	8.130	7.106	184222	175068	4.451	3.472
35)	L7 AR-1260-5	8.472	7.375	39174	31877	0.586	0.284 #
36)	L8 AR-1262-1	7.830	6.840	42024	130039	1.162	3.402 #
37)	L8 AR-1262-2	8.472	7.106	39174	175068	0.461	2.321 #
38)	L8 AR-1262-3	8.815	7.706f	77662	698703	1.289	11.996 #
39)	L8 AR-1262-4	8.903	7.706	197006	698703	4.086	6.798 #
40)	L8 AR-1262-5	9.574	8.231	44879	119264	1.719	2.737 #
41)	L9 AR-1268-1	8.776	7.706f	-2544	698703	N.D.	4.207 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120423\
 Data File : PO100250.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Dec 2023 15:25
 Operator : YP/AJ
 Sample : 05599-11
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 04 16:48:03 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 15:58:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.903	7.706	197006	698703	2.046	4.577 #
43) L9	AR-1268-3	9.126	7.945	32589	199494	0.361	1.392 #
44) L9	AR-1268-4	9.574	8.231	44879	119264	1.578	2.538 #
45) L9	AR-1268-5	10.002	8.531	133811	160450	0.532	0.403



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120423\
Data File : PO100250.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Dec 2023 15:25
Operator : YP/AJ
Sample : 05599-11
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 04 16:48:03 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 15:58:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.480 min
Delta R.T.: 0.000 min
Response: 35319455
Conc: 20.69 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.703 min
Delta R.T.: -0.003 min
Response: 32546488
Conc: 21.70 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.362 min
Delta R.T.: 0.005 min
Response: 15679278
Conc: 21.59 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.795 min
Delta R.T.: -0.006 min
Response: 23424917
Conc: 20.93 ng/ml



#3 AR-1016-1


R.T.: 5.671 min
Delta R.T.: 0.004 min
Response: 84850
Conc: 1.97 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.809 min
Delta R.T.: -0.003 min
Response: 61653
Conc: 1.95 ng/ml




#4 AR-1016-2

R.T.: 5.691 min
Delta R.T.: 0.001 min
Response: 182185
Conc: 2.81 ng/ml



#4 AR-1016-2


R.T.: 4.835 min
Delta R.T.: 0.003 min
Response: 45441
Conc: 1.05 ng/ml



#5 AR-1016-3


R.T.: 5.749 min
Delta R.T.: -0.004 min
Response: 156921
Conc: 3.60 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 5.007 min
Delta R.T.: -0.004 min
Response: 275305
Conc: 11.87 ng/ml




#6 AR-1016-4

R.T.: 5.862 min
Delta R.T.: 0.010 min
Response: 60221
Conc: 1.82 ng/ml



#6 AR-1016-4


R.T.: 5.046 min
Delta R.T.: -0.006 min
Response: 68517
Conc: 3.34 ng/ml



#7 AR-1016-5


R.T.: 6.149 min
Delta R.T.: -0.002 min
Response: 27957
Conc: 0.89 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.254 min
Delta R.T.: -0.015 min
Response: 142924
Conc: 5.60 ng/ml




#8 AR-1221-1

R.T.: 4.689 min
Delta R.T.: -0.003 min
Response: 1950256
Conc: 94.60 ng/ml



#8 AR-1221-1


R.T.: 3.926 min
Delta R.T.: 0.003 min
Response: 744699
Conc: 39.17 ng/ml



#9 AR-1221-2


R.T.: 4.791 min
Delta R.T.: 0.011 min
Response: 592509
Conc: 39.23 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.991 min
Delta R.T.: -0.019 min
Response: 2963400
Conc: 227.41 ng/ml




#10 AR-1221-3

R.T.: 4.857 min
Delta R.T.: 0.000 min
Response: 90526
Conc: 2.02 ng/ml



#10 AR-1221-3


R.T.: 4.094 min
Delta R.T.: 0.005 min
Response: 544703
Conc: 15.16 ng/ml



#11 AR-1232-1


R.T.: 4.857 min
Delta R.T.: 0.000 min
Response: 90526
Conc: 2.54 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 4.094 min
Delta R.T.: 0.006 min
Response: 544703
Conc: 18.56 ng/ml




#12 AR-1232-2

R.T.: 5.419 min
Delta R.T.: 0.025 min
Response: 347546
Conc: 19.97 ng/ml



#12 AR-1232-2


R.T.: 4.835 min
Delta R.T.: 0.004 min
Response: 45441
Conc: 2.29 ng/ml



#13 AR-1232-3


R.T.: 5.691 min
Delta R.T.: 0.002 min
Response: 182185
Conc: 6.43 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 5.007 min
Delta R.T.: -0.003 min
Response: 275305
Conc: 26.57 ng/ml




#14 AR-1232-4

R.T.: 5.862 min
Delta R.T.: 0.010 min
Response: 60221
Conc: 4.34 ng/ml



#14 AR-1232-4


R.T.: 5.098 min
Delta R.T.: 0.003 min
Response: 114784
Conc: 11.55 ng/ml



#15 AR-1232-5


R.T.: 5.943 min
Delta R.T.: 0.000 min
Response: 66514
Conc: 5.42 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.254 min
Delta R.T.: -0.014 min
Response: 142924
Conc: 13.23 ng/ml




#16 AR-1242-1

R.T.: 5.671 min
Delta R.T.: 0.004 min
Response: 84850
Conc: 2.45 ng/ml



#16 AR-1242-1


R.T.: 4.809 min
Delta R.T.: -0.003 min
Response: 61653
Conc: 2.35 ng/ml



#17 AR-1242-2


R.T.: 5.691 min
Delta R.T.: 0.000 min
Response: 182185
Conc: 3.47 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.835 min
Delta R.T.: 0.003 min
Response: 45441
Conc: 1.29 ng/ml




#18 AR-1242-3

R.T.: 5.749 min
Delta R.T.: -0.005 min
Response: 156921
Conc: 4.51 ng/ml



#18 AR-1242-3


R.T.: 5.007 min
Delta R.T.: -0.004 min
Response: 275305
Conc: 14.38 ng/ml



#19 AR-1242-4


R.T.: 5.862 min
Delta R.T.: 0.009 min
Response: 60221
Conc: 2.29 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 5.098 min
Delta R.T.: 0.003 min
Response: 114784
Conc: 5.70 ng/ml




#20 AR-1242-5

R.T.: 6.588 min
Delta R.T.: -0.012 min
Response: 119159
Conc: 5.33 ng/ml



#20 AR-1242-5


R.T.: 5.609 min
Delta R.T.: -0.016 min
Response: 330016
Conc: 14.12 ng/ml



#21 AR-1248-1


R.T.: 5.671 min
Delta R.T.: 0.005 min
Response: 84850
Conc: 3.28 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.809 min
Delta R.T.: -0.002 min
Response: 61653
Conc: 3.08 ng/ml




#22 AR-1248-2

R.T.: 5.943 min
Delta R.T.: 0.000 min
Response: 66514
Conc: 1.63 ng/ml



#22 AR-1248-2


R.T.: 5.046 min
Delta R.T.: -0.005 min
Response: 68517
Conc: 2.45 ng/ml



#23 AR-1248-3


R.T.: 6.149 min
Delta R.T.: -0.002 min
Response: 27957
Conc: 0.65 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 5.098 min
Delta R.T.: 0.004 min
Response: 114784
Conc: 3.97 ng/ml




#24 AR-1248-4

R.T.: 6.535 min
Delta R.T.: -0.025 min
Response: 321021
Conc: 8.52 ng/ml



#24 AR-1248-4


R.T.: 5.254 min
Delta R.T.: -0.014 min
Response: 142924
Conc: 4.36 ng/ml



#25 AR-1248-5


R.T.: 6.588 min
Delta R.T.: -0.010 min
Response: 119159
Conc: 3.23 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.681 min
Delta R.T.: 0.015 min
Response: 102117
Conc: 3.42 ng/ml




#26 AR-1254-1

R.T.: 6.535 min
Delta R.T.: 0.002 min
Response: 321021
Conc: 6.67 ng/ml



#26 AR-1254-1


R.T.: 5.609 min
Delta R.T.: -0.017 min
Response: 330016
Conc: 6.37 ng/ml



#27 AR-1254-2


R.T.: 6.756 min
Delta R.T.: 0.000 min
Response: 75686
Conc: 1.10 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.774 min
Delta R.T.: 0.000 min
Response: 86390
Conc: 1.86 ng/ml




#28 AR-1254-3

R.T.: 7.128 min
Delta R.T.: 0.002 min
Response: 166524
Conc: 2.61 ng/ml



#28 AR-1254-3


R.T.: 6.193 min
Delta R.T.: 0.010 min
Response: 356784
Conc: 4.57 ng/ml



#29 AR-1254-4


R.T.: 7.415 min
Delta R.T.: 0.001 min
Response: 40607
Conc: 1.07 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.415 min
Delta R.T.: 0.003 min
Response: 45066
Conc: 1.06 ng/ml




#30 AR-1254-5

R.T.: 7.844 min
Delta R.T.: 0.005 min
Response: 86861
Conc: 2.03 ng/ml



#30 AR-1254-5


R.T.: 6.840 min
Delta R.T.: 0.005 min
Response: 130039
Conc: 1.80 ng/ml



#31 AR-1260-1


R.T.: 7.294 min
Delta R.T.: 0.003 min
Response: 49497
Conc: 0.93 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.347 min
Delta R.T.: 0.032 min
Response: 607119
Conc: 10.51 ng/ml




#32 AR-1260-2

R.T.: 7.540 min
Delta R.T.: -0.011 min
Response: 82376
Conc: 1.56 ng/ml



#32 AR-1260-2


R.T.: 6.484 min
Delta R.T.: -0.019 min
Response: 338242
Conc: 5.09 ng/ml



#33 AR-1260-3


R.T.: 7.911 min
Delta R.T.: -0.004 min
Response: 32993
Conc: 0.92 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.655 min
Delta R.T.: -0.005 min
Response: 19499
Conc: 0.30 ng/ml




#34 AR-1260-4

R.T.: 8.130 min
Delta R.T.: -0.013 min
Response: 184222
Conc: 4.45 ng/ml



#34 AR-1260-4


R.T.: 7.106 min
Delta R.T.: -0.029 min
Response: 175068
Conc: 3.47 ng/ml



#35 AR-1260-5


R.T.: 8.472 min
Delta R.T.: -0.001 min
Response: 39174
Conc: 0.59 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.375 min
Delta R.T.: -0.003 min
Response: 31877
Conc: 0.28 ng/ml




#36 AR-1262-1

R.T.: 7.830 min
Delta R.T.: -0.005 min
Response: 42024
Conc: 1.16 ng/ml



#36 AR-1262-1


R.T.: 6.840 min
Delta R.T.: 0.006 min
Response: 130039
Conc: 3.40 ng/ml



#37 AR-1262-2


R.T.: 8.472 min
Delta R.T.: 0.000 min
Response: 39174
Conc: 0.46 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 7.106 min
Delta R.T.: -0.029 min
Response: 175068
Conc: 2.32 ng/ml




#38 AR-1262-3

R.T.: 8.815 min
Delta R.T.: 0.019 min
Response: 77662
Conc: 1.29 ng/ml



#38 AR-1262-3


R.T.: 7.706 min
Delta R.T.: 0.042 min
Response: 698703
Conc: 12.00 ng/ml



#39 AR-1262-4


R.T.: 8.903 min
Delta R.T.: 0.013 min
Response: 197006
Conc: 4.09 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.706 min
Delta R.T.: -0.022 min
Response: 698703
Conc: 6.80 ng/ml




#40 AR-1262-5

R.T.: 9.574 min
Delta R.T.: 0.005 min
Response: 44879
Conc: 1.72 ng/ml



#40 AR-1262-5


R.T.: 8.231 min
Delta R.T.: 0.001 min
Response: 119264
Conc: 2.74 ng/ml



#41 AR-1268-1


R.T.: 8.776 min
Delta R.T.: -0.017 min
Response: -2544
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.706 min
Delta R.T.: 0.042 min
Response: 698703
Conc: 4.21 ng/ml




#42 AR-1268-2

R.T.: 8.903 min
Delta R.T.: 0.011 min
Response: 197006
Conc: 2.05 ng/ml



#42 AR-1268-2


R.T.: 7.706 min
Delta R.T.: -0.022 min
Response: 698703
Conc: 4.58 ng/ml



#43 AR-1268-3


R.T.: 9.126 min
Delta R.T.: -0.003 min
Response: 32589
Conc: 0.36 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.945 min
Delta R.T.: 0.007 min
Response: 199494
Conc: 1.39 ng/ml




#44 AR-1268-4

R.T.: 9.574 min
Delta R.T.: 0.004 min
Response: 44879
Conc: 1.58 ng/ml



#44 AR-1268-4


R.T.: 8.231 min
Delta R.T.: 0.002 min
Response: 119264
Conc: 2.54 ng/ml



#45 AR-1268-5

R.T.: 10.002 min
Delta R.T.: -0.002 min
Response: 133811
Conc: 0.53 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.531 min
Delta R.T.: -0.002 min
Response: 160450
Conc: 0.40 ng/ml