

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO120425\
 Data File : PO115598-1.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Dec 2025 17:39
 Operator : YP/AJ
 Sample : AR1248ICC50
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 05 00:41:41 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO120425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 05 00:36:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.335	3.637	32904422	21052074	4.397	4.456
2) SA Decachlor...	9.910	8.620	23193521	20016647	4.453	4.898
Target Compounds						
21) L5 AR-1248-1	5.472	4.713	6803477	5724201	43.735	50.928
22) L5 AR-1248-2	5.744	4.948	8593662	7963493	40.838	52.359 #
23) L5 AR-1248-3	5.945	4.990	14141493	8347640	59.225	51.886
24) L5 AR-1248-4	6.343	5.161	17069181	9641303	59.992	51.515
25) L5 AR-1248-5	6.379	5.551	12140324	10870079	45.421	56.037



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120425\
Data File : PO115598-1.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Dec 2025 17:39
Operator : YP/AJ
Sample : AR1248ICC50
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 05 00:41:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120425.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 05 00:36:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.335 min
Delta R.T.: 0.000 min
Response: 32904422
Conc: 4.40 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.637 min
Delta R.T.: 0.000 min
Response: 21052074
Conc: 4.46 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.910 min
Delta R.T.: -0.001 min
Response: 23193521
Conc: 4.45 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.620 min
Delta R.T.: -0.002 min
Response: 20016647
Conc: 4.90 ng/ml



#21 AR-1248-1


R.T.: 5.472 min
Delta R.T.: -0.002 min
Response: 6803477
Conc: 43.73 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.713 min
Delta R.T.: 0.000 min
Response: 5724201
Conc: 50.93 ng/ml




#22 AR-1248-2

R.T.: 5.744 min
Delta R.T.: -0.001 min
Response: 8593662
Conc: 40.84 ng/ml



#22 AR-1248-2


R.T.: 4.948 min
Delta R.T.: 0.000 min
Response: 7963493
Conc: 52.36 ng/ml



#23 AR-1248-3


R.T.: 5.945 min
Delta R.T.: 0.000 min
Response: 14141493
Conc: 59.23 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 4.990 min
Delta R.T.: 0.000 min
Response: 8347640
Conc: 51.89 ng/ml




#24 AR-1248-4

R.T.: 6.343 min
Delta R.T.: 0.002 min
Response: 17069181
Conc: 59.99 ng/ml



#24 AR-1248-4


R.T.: 5.161 min
Delta R.T.: 0.000 min
Response: 9641303
Conc: 51.52 ng/ml



#25 AR-1248-5

R.T.: 6.379 min
Delta R.T.: 0.000 min
Response: 12140324
Conc: 45.42 ng/ml

Instrument :
ECD_O
ClientSampleId :



#25 AR-1248-5

R.T.: 5.551 min
Delta R.T.: 0.000 min
Response: 10870079
Conc: 56.04 ng/ml