

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120425\
 Data File : PO115605.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Dec 2025 20:05
 Operator : YP/AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 05 01:30:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 05 01:29:15 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.336	3.638	714.8E6	446.3E6	87.345	87.323
2) SA Decachlor...	9.910	8.620	857.7E6	643.8E6	87.551	86.453
Target Compounds						
41) L9 AR-1268-1	8.496	7.524	912.5E6	795.6E6	878.224	872.110
42) L9 AR-1268-2	8.585	7.588	790.2E6	677.2E6	881.900	876.610
43) L9 AR-1268-3	8.804	7.795	685.6E6	577.7E6	880.155	875.450
44) L9 AR-1268-4	9.198	8.079	285.0E6	233.7E6	871.991	869.801
45) L9 AR-1268-5	9.590	8.367	2031.0E6	1545.4E6	888.725	883.942



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120425\
Data File : PO115605.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Dec 2025 20:05
Operator : YP/AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 05 01:30:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120425.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 05 01:29:15 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.336 min
Delta R.T.: 0.000 min
Response: 714807195
Conc: 87.34 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.638 min
Delta R.T.: 0.000 min
Response: 446321199
Conc: 87.32 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.910 min
Delta R.T.: 0.000 min
Response: 857658507
Conc: 87.55 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.620 min
Delta R.T.: 0.000 min
Response: 643830457
Conc: 86.45 ng/ml



#41 AR-1268-1


R.T.: 8.496 min
Delta R.T.: 0.001 min
Response: 912508253
Conc: 878.22 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.524 min
Delta R.T.: 0.000 min
Response: 795605721
Conc: 872.11 ng/ml




#42 AR-1268-2

R.T.: 8.585 min
Delta R.T.: 0.001 min
Response: 790178939
Conc: 881.90 ng/ml



#42 AR-1268-2


R.T.: 7.588 min
Delta R.T.: 0.000 min
Response: 677182657
Conc: 876.61 ng/ml



#43 AR-1268-3


R.T.: 8.804 min
Delta R.T.: 0.001 min
Response: 685629606
Conc: 880.15 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.795 min
Delta R.T.: 0.000 min
Response: 577726676
Conc: 875.45 ng/ml




#44 AR-1268-4

R.T.: 9.198 min
Delta R.T.: 0.001 min
Response: 285018478
Conc: 871.99 ng/ml



#44 AR-1268-4


R.T.: 8.079 min
Delta R.T.: 0.000 min
Response: 233669633
Conc: 869.80 ng/ml



#45 AR-1268-5

R.T.: 9.590 min
Delta R.T.: 0.000 min
Response: 2030988326
Conc: 888.73 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.367 min
Delta R.T.: -0.001 min
Response: 1545366536
Conc: 883.94 ng/ml