

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120425\
 Data File : PO115599.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Dec 2025 17:57
 Operator : YP/AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 05 01:07:14 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 05 01:04:29 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.335	3.637	722.4E6	449.1E6	94.476	94.071
2)	SA Decachlor...	9.910	8.621	499.1E6	381.9E6	94.610	92.989
Target Compounds							
26)	L6 AR-1254-1	6.316	5.510	240.9E6	252.8E6	933.900	906.586
27)	L6 AR-1254-2	6.533	5.658	372.6E6	218.6E6	932.342	909.462
28)	L6 AR-1254-3	6.893	6.059	381.0E6	354.7E6	945.839	923.024
29)	L6 AR-1254-4	7.174	6.287	325.5E6	237.8E6	939.743	909.944
30)	L6 AR-1254-5	7.590	6.704	304.1E6	322.8E6	943.917	927.389



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120425\
Data File : PO115599.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Dec 2025 17:57
Operator : YP/AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 05 01:07:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120425.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 05 01:04:29 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.335 min
Delta R.T.: 0.002 min
Response: 722390756
Conc: 94.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.637 min
Delta R.T.: 0.000 min
Response: 449147255
Conc: 94.07 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.910 min
Delta R.T.: 0.000 min
Response: 499053440
Conc: 94.61 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.621 min
Delta R.T.: 0.001 min
Response: 381899484
Conc: 92.99 ng/ml



#26 AR-1254-1


R.T.: 6.316 min
Delta R.T.: 0.000 min
Response: 240930759
Conc: 933.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000




#26 AR-1254-1

R.T.: 5.510 min
Delta R.T.: 0.001 min
Response: 252780245
Conc: 906.59 ng/ml




#27 AR-1254-2

R.T.: 6.533 min
Delta R.T.: 0.001 min
Response: 372595639
Conc: 932.34 ng/ml



#27 AR-1254-2


R.T.: 5.658 min
Delta R.T.: 0.000 min
Response: 218597114
Conc: 909.46 ng/ml



#28 AR-1254-3


R.T.: 6.893 min
Delta R.T.: 0.000 min
Response: 381021457
Conc: 945.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000




#28 AR-1254-3

R.T.: 6.059 min
Delta R.T.: 0.001 min
Response: 354655855
Conc: 923.02 ng/ml




#29 AR-1254-4

R.T.: 7.174 min
Delta R.T.: 0.001 min
Response: 325539364
Conc: 939.74 ng/ml



#29 AR-1254-4


R.T.: 6.287 min
Delta R.T.: 0.001 min
Response: 237848426
Conc: 909.94 ng/ml



#30 AR-1254-5

R.T.: 7.590 min
Delta R.T.: 0.001 min
Response: 304124469
Conc: 943.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000



#30 AR-1254-5

R.T.: 6.704 min
Delta R.T.: 0.001 min
Response: 322835753
Conc: 927.39 ng/ml