

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120522\
 Data File : PO091079.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Dec 2022 08:17
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 06 09:41:38 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Nov 28 11:25:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.298	3.475	176.9E6	59498294	52.322	56.689
2) SA Decachlor...	10.043	8.477	122.9E6	51369786	46.282	52.911
Target Compounds						
16) L4 AR-1242-1	5.473	4.561	44715798	15314777	500.976	527.430
17) L4 AR-1242-2	5.495	4.579	62905550	21242306	495.441	526.972
18) L4 AR-1242-3	5.558	4.754	40216220	11681754	482.438	558.596
19) L4 AR-1242-4	5.657	4.839	32016421	12541666	503.991	549.250
20) L4 AR-1242-5	6.398	5.364	36754193	15727466	537.884	557.412



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120522\
Data File : PO091079.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Dec 2022 08:17
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 06 09:41:38 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Nov 28 11:25:59 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.298 min
Delta R.T.: 0.000 min
Response: 176906497
Conc: 52.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.475 min
Delta R.T.: -0.001 min
Response: 59498294
Conc: 56.69 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.043 min
Delta R.T.: -0.007 min
Response: 122932697
Conc: 46.28 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.477 min
Delta R.T.: -0.005 min
Response: 51369786
Conc: 52.91 ng/ml



#16 AR-1242-1


R.T.: 5.473 min
Delta R.T.: -0.003 min
Response: 44715798
Conc: 500.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.561 min
Delta R.T.: -0.002 min
Response: 15314777
Conc: 527.43 ng/ml




#17 AR-1242-2

R.T.: 5.495 min
Delta R.T.: -0.003 min
Response: 62905550
Conc: 495.44 ng/ml



#17 AR-1242-2


R.T.: 4.579 min
Delta R.T.: -0.002 min
Response: 21242306
Conc: 526.97 ng/ml



#18 AR-1242-3


R.T.: 5.558 min
Delta R.T.: -0.003 min
Response: 40216220
Conc: 482.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.754 min
Delta R.T.: -0.002 min
Response: 11681754
Conc: 558.60 ng/ml




#19 AR-1242-4

R.T.: 5.657 min
Delta R.T.: -0.003 min
Response: 32016421
Conc: 503.99 ng/ml



#19 AR-1242-4


R.T.: 4.839 min
Delta R.T.: -0.002 min
Response: 12541666
Conc: 549.25 ng/ml



#20 AR-1242-5

R.T.: 6.398 min
Delta R.T.: -0.004 min
Response: 36754193
Conc: 537.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.364 min
Delta R.T.: -0.003 min
Response: 15727466
Conc: 557.41 ng/ml