

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120523\
 Data File : PO100287.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Dec 2023 12:43
 Operator : YP/AJ
 Sample : 05608-01 10X
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CLI-2023-15A

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 05 14:01:37 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 15:58:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.477	3.701	2647698	2543551	1.551	1.696
2) SA Decachlor...	10.359	8.791	1478390	2380684	2.035	2.128

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120523\
Data File : PO100287.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Dec 2023 12:43
Operator : YP/AJ
Sample : 05608-01 10X
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
CLI-2023-15A

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 05 14:01:37 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 15:58:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.477 min
Delta R.T.: -0.003 min
Response: 2647698
Conc: 1.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
CLI-2023-15A




#1 Tetrachloro-m-xylene

R.T.: 3.701 min
Delta R.T.: -0.005 min
Response: 2543551
Conc: 1.70 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.359 min
Delta R.T.: 0.002 min
Response: 1478390
Conc: 2.04 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.791 min
Delta R.T.: -0.010 min
Response: 2380684
Conc: 2.13 ng/ml