

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120525\
 Data File : PO115626.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Dec 2025 15:24
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 12/08/2025
 Supervised By :Yogesh Patel 12/08/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 06 00:19:01 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120425.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 05 02:11:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.334	3.637	410.4E6	250.7E6	52.839	51.447
2) SA Decachlor...	9.909	8.619	283.7E6	212.4E6	53.924	49.389m
Target Compounds						
21) L5 AR-1248-1	5.473	4.712	86229399	59492940	568.170	531.999
22) L5 AR-1248-2	5.743	4.947	115.9E6	71956711	577.275	472.499
23) L5 AR-1248-3	5.944	4.990	136.1E6	84037450	551.898	523.510
24) L5 AR-1248-4	6.340	5.160	155.4E6	106.5E6	527.025m	569.104
25) L5 AR-1248-5	6.377	5.550	144.4E6	108.4E6	557.649m	547.691

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120525\
Data File : PO115626.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Dec 2025 15:24
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 12/08/2025
Supervised By :Yogesh Patel 12/08/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 06 00:19:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120425.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 05 02:11:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.334 min
Delta R.T.: -0.002 min
Response: 410412631
Conc: 52.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 12/08/2025
Supervised By :Yogesh Patel 12/08/2025




#1 Tetrachloro-m-xylene

R.T.: 3.637 min
Delta R.T.: -0.001 min
Response: 250717112
Conc: 51.45 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.909 min
Delta R.T.: -0.003 min
Response: 283699787
Conc: 53.92 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.619 min
Delta R.T.: -0.003 min
Response: 212385235
Conc: 49.39 ng/ml m




#21 AR-1248-1

R.T.: 5.473 min
Delta R.T.: -0.002 min
Response: 86229399
Conc: 568.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 12/08/2025
Supervised By :Yogesh Patel 12/08/2025




#21 AR-1248-1

R.T.: 4.712 min
Delta R.T.: -0.001 min
Response: 59492940
Conc: 532.00 ng/ml




#22 AR-1248-2

R.T.: 5.743 min
Delta R.T.: -0.002 min
Response: 115925512
Conc: 577.28 ng/ml



#22 AR-1248-2

R.T.: 4.947 min
Delta R.T.: -0.002 min
Response: 71956711
Conc: 472.50 ng/ml




#23 AR-1248-3

R.T.: 5.944 min
Delta R.T.: -0.001 min
Response: 136097673
Conc: 551.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 12/08/2025
Supervised By :Yogesh Patel 12/08/2025




#23 AR-1248-3

R.T.: 4.990 min
Delta R.T.: -0.001 min
Response: 84037450
Conc: 523.51 ng/ml




#24 AR-1248-4

R.T.: 6.340 min
Delta R.T.: -0.002 min
Response: 155395830
Conc: 527.02 ng/ml m



#24 AR-1248-4

R.T.: 5.160 min
Delta R.T.: -0.001 min
Response: 106505506
Conc: 569.10 ng/ml




#25 AR-1248-5

R.T.: 6.377 min
Delta R.T.: -0.002 min
Response: 144389794
Conc: 557.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 12/08/2025
Supervised By :Yogesh Patel 12/08/2025



#25 AR-1248-5

R.T.: 5.550 min
Delta R.T.: -0.001 min
Response: 108418338
Conc: 547.69 ng/ml