

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120622\
 Data File : PO091112.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Dec 2022 19:29
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 06 23:52:05 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Nov 28 11:25:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.298	3.476	178.9E6	61635899	52.897	58.725
2) SA Decachlor...	10.043	8.477	124.2E6	54245965	46.751	55.873
Target Compounds						
26) L6 AR-1254-1	6.336	5.365	60613297	35788900	482.623	610.661 #
27) L6 AR-1254-2	6.555	5.514	97309660	30213395	516.074	566.109
28) L6 AR-1254-3	6.923	5.918	98028774	44755523	522.995	549.339
29) L6 AR-1254-4	7.210	6.148	72171420	27051462	522.452	552.891
30) L6 AR-1254-5	7.632	6.567	76975673	40702925	501.703	546.022



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120622\
Data File : PO091112.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Dec 2022 19:29
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 06 23:52:05 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Nov 28 11:25:59 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.298 min
Delta R.T.: 0.000 min
Response: 178851457
Conc: 52.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.476 min
Delta R.T.: 0.000 min
Response: 61635899
Conc: 58.73 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.043 min
Delta R.T.: -0.007 min
Response: 124180300
Conc: 46.75 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.477 min
Delta R.T.: -0.005 min
Response: 54245965
Conc: 55.87 ng/ml



#26 AR-1254-1


R.T.: 6.336 min
Delta R.T.: -0.003 min
Response: 60613297
Conc: 482.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.365 min
Delta R.T.: -0.003 min
Response: 35788900
Conc: 610.66 ng/ml




#27 AR-1254-2

R.T.: 6.555 min
Delta R.T.: -0.004 min
Response: 97309660
Conc: 516.07 ng/ml



#27 AR-1254-2


R.T.: 5.514 min
Delta R.T.: -0.003 min
Response: 30213395
Conc: 566.11 ng/ml



#28 AR-1254-3


R.T.: 6.923 min
Delta R.T.: -0.004 min
Response: 98028774
Conc: 523.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.918 min
Delta R.T.: -0.004 min
Response: 44755523
Conc: 549.34 ng/ml




#29 AR-1254-4

R.T.: 7.210 min
Delta R.T.: -0.006 min
Response: 72171420
Conc: 522.45 ng/ml



#29 AR-1254-4


R.T.: 6.148 min
Delta R.T.: -0.004 min
Response: 27051462
Conc: 552.89 ng/ml



#30 AR-1254-5

R.T.: 7.632 min
Delta R.T.: -0.006 min
Response: 76975673
Conc: 501.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.567 min
Delta R.T.: -0.005 min
Response: 40702925
Conc: 546.02 ng/ml