

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120622\
 Data File : PO091119.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Dec 2022 22:11
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 06 23:55:47 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Nov 28 11:25:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.298	3.476	182.0E6	62791569	53.821	59.827
2) SA Decachlor...	10.043	8.476	125.9E6	54162789	47.411	55.788
Target Compounds						
26) L6 AR-1254-1	6.335	5.364	60842826	35923389	484.451	612.956 #
27) L6 AR-1254-2	6.554	5.514	98113535	30479412	520.338	571.094
28) L6 AR-1254-3	6.922	5.917	99152829	45171896	528.992	554.449
29) L6 AR-1254-4	7.210	6.148	72581297	27413112	525.419	560.283
30) L6 AR-1254-5	7.631	6.567	78105867	41326637	509.070	554.389



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120622\
Data File : PO091119.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Dec 2022 22:11
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 06 23:55:47 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Nov 28 11:25:59 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.298 min
Delta R.T.: 0.000 min
Response: 181975920
Conc: 53.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.476 min
Delta R.T.: 0.000 min
Response: 62791569
Conc: 59.83 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.043 min
Delta R.T.: -0.007 min
Response: 125932490
Conc: 47.41 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.476 min
Delta R.T.: -0.005 min
Response: 54162789
Conc: 55.79 ng/ml



#26 AR-1254-1


R.T.: 6.335 min
Delta R.T.: -0.003 min
Response: 60842826
Conc: 484.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.364 min
Delta R.T.: -0.004 min
Response: 35923389
Conc: 612.96 ng/ml




#27 AR-1254-2

R.T.: 6.554 min
Delta R.T.: -0.005 min
Response: 98113535
Conc: 520.34 ng/ml



#27 AR-1254-2


R.T.: 5.514 min
Delta R.T.: -0.003 min
Response: 30479412
Conc: 571.09 ng/ml



#28 AR-1254-3


R.T.: 6.922 min
Delta R.T.: -0.005 min
Response: 99152829
Conc: 528.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.917 min
Delta R.T.: -0.004 min
Response: 45171896
Conc: 554.45 ng/ml




#29 AR-1254-4

R.T.: 7.210 min
Delta R.T.: -0.006 min
Response: 72581297
Conc: 525.42 ng/ml



#29 AR-1254-4


R.T.: 6.148 min
Delta R.T.: -0.004 min
Response: 27413112
Conc: 560.28 ng/ml



#30 AR-1254-5

R.T.: 7.631 min
Delta R.T.: -0.006 min
Response: 78105867
Conc: 509.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.567 min
Delta R.T.: -0.005 min
Response: 41326637
Conc: 554.39 ng/ml