

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120623\
 Data File : PO100328.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Dec 2023 16:02
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 07 00:23:25 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 15:58:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.477	3.702	67300840	64793176	39.430	43.198
2)	SA Decachlor...	10.356	8.792	37983723	55911208	52.294	49.965
Target Compounds							
21)	L5 AR-1248-1	5.665	4.806	10646551	8818433	411.817	441.104
22)	L5 AR-1248-2	5.942	5.046	16571484	12236888	405.202	437.332
23)	L5 AR-1248-3	6.149	5.089	17721457	12928566	413.733	447.697
24)	L5 AR-1248-4	6.558	5.264	17327173	14922069	459.712	455.085
25)	L5 AR-1248-5	6.597	5.661	16650486	14380005	450.831	481.393



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120623\
Data File : PO100328.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Dec 2023 16:02
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 07 00:23:25 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 15:58:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.477 min
Delta R.T.: -0.003 min
Response: 67300840
Conc: 39.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.702 min
Delta R.T.: -0.004 min
Response: 64793176
Conc: 43.20 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.356 min
Delta R.T.: 0.000 min
Response: 37983723
Conc: 52.29 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.792 min
Delta R.T.: -0.009 min
Response: 55911208
Conc: 49.97 ng/ml



#21 AR-1248-1


R.T.: 5.665 min
Delta R.T.: -0.001 min
Response: 10646551
Conc: 411.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.806 min
Delta R.T.: -0.004 min
Response: 8818433
Conc: 441.10 ng/ml




#22 AR-1248-2

R.T.: 5.942 min
Delta R.T.: -0.002 min
Response: 16571484
Conc: 405.20 ng/ml



#22 AR-1248-2


R.T.: 5.046 min
Delta R.T.: -0.005 min
Response: 12236888
Conc: 437.33 ng/ml



#23 AR-1248-3


R.T.: 6.149 min
Delta R.T.: -0.002 min
Response: 17721457
Conc: 413.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.089 min
Delta R.T.: -0.005 min
Response: 12928566
Conc: 447.70 ng/ml




#24 AR-1248-4

R.T.: 6.558 min
Delta R.T.: -0.002 min
Response: 17327173
Conc: 459.71 ng/ml



#24 AR-1248-4


R.T.: 5.264 min
Delta R.T.: -0.005 min
Response: 14922069
Conc: 455.08 ng/ml



#25 AR-1248-5

R.T.: 6.597 min
Delta R.T.: -0.002 min
Response: 16650486
Conc: 450.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.661 min
Delta R.T.: -0.005 min
Response: 14380005
Conc: 481.39 ng/ml