

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120623\
 Data File : PO100329.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Dec 2023 16:19
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 07 00:23:44 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 15:58:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.476	3.702	70349694	67715991	41.217	45.147
2) SA Decachlor...	10.354	8.792	39737349	57441919	54.708	51.333
Target Compounds						
26) L6 AR-1254-1	6.530	5.619	20003513	23407949	415.416	451.567
27) L6 AR-1254-2	6.751	5.767	29559771	21507583	431.023	461.856
28) L6 AR-1254-3	7.122	6.177	28879912	36671655	453.321	470.205
29) L6 AR-1254-4	7.412	6.407	18779428	21555337	496.760	505.084
30) L6 AR-1254-5	7.835	6.828	21441082	36360537	500.933	502.474



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120623\
Data File : PO100329.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Dec 2023 16:19
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 07 00:23:44 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 15:58:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.476 min
Delta R.T.: -0.004 min
Response: 70349694
Conc: 41.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.702 min
Delta R.T.: -0.004 min
Response: 67715991
Conc: 45.15 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.354 min
Delta R.T.: -0.003 min
Response: 39737349
Conc: 54.71 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.792 min
Delta R.T.: -0.009 min
Response: 57441919
Conc: 51.33 ng/ml



#26 AR-1254-1


R.T.: 6.530 min
Delta R.T.: -0.003 min
Response: 20003513
Conc: 415.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.619 min
Delta R.T.: -0.006 min
Response: 23407949
Conc: 451.57 ng/ml




#27 AR-1254-2

R.T.: 6.751 min
Delta R.T.: -0.004 min
Response: 29559771
Conc: 431.02 ng/ml



#27 AR-1254-2


R.T.: 5.767 min
Delta R.T.: -0.006 min
Response: 21507583
Conc: 461.86 ng/ml



#28 AR-1254-3


R.T.: 7.122 min
Delta R.T.: -0.003 min
Response: 28879912
Conc: 453.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.177 min
Delta R.T.: -0.006 min
Response: 36671655
Conc: 470.21 ng/ml




#29 AR-1254-4

R.T.: 7.412 min
Delta R.T.: -0.002 min
Response: 18779428
Conc: 496.76 ng/ml



#29 AR-1254-4


R.T.: 6.407 min
Delta R.T.: -0.006 min
Response: 21555337
Conc: 505.08 ng/ml



#30 AR-1254-5

R.T.: 7.835 min
Delta R.T.: -0.003 min
Response: 21441082
Conc: 500.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.828 min
Delta R.T.: -0.007 min
Response: 36360537
Conc: 502.47 ng/ml