

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120624\
 Data File : PO108393.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Dec 2024 23:48
 Operator : YP/AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 07 05:09:17 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 07 05:08:30 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.709	3.707	496.6E6	280.8E6	54.077	54.172
2)	SA Decachlor...	8.788	8.740	411.8E6	224.2E6	53.122	53.431
Target Compounds							
26)	L6 AR-1254-1	5.620	5.601	273.4E6	144.3E6	532.286	532.859
27)	L6 AR-1254-2	5.768	5.749	242.1E6	127.8E6	531.623	532.309
28)	L6 AR-1254-3	6.175	6.153	387.5E6	203.0E6	538.152	540.509
29)	L6 AR-1254-4	6.405	6.381	237.3E6	117.8E6	540.738	541.351
30)	L6 AR-1254-5	6.827	6.800	339.7E6	174.8E6	534.857	541.443



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120624\
Data File : PO108393.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Dec 2024 23:48
Operator : YP/AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 07 05:09:17 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120624.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 07 05:08:30 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.709 min
Delta R.T.: 0.000 min
Response: 496582945
Conc: 54.08 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.707 min
Delta R.T.: 0.000 min
Response: 280770967
Conc: 54.17 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.788 min
Delta R.T.: 0.002 min
Response: 411835318
Conc: 53.12 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.740 min
Delta R.T.: 0.001 min
Response: 224175723
Conc: 53.43 ng/ml



#26 AR-1254-1


R.T.: 5.620 min
Delta R.T.: 0.002 min
Response: 273403576
Conc: 532.29 ng/ml

Instrument :
ECD_O
ClientSampleId :




#26 AR-1254-1

R.T.: 5.601 min
Delta R.T.: 0.000 min
Response: 144315218
Conc: 532.86 ng/ml




#27 AR-1254-2

R.T.: 5.768 min
Delta R.T.: 0.001 min
Response: 242131954
Conc: 531.62 ng/ml



#27 AR-1254-2


R.T.: 5.749 min
Delta R.T.: 0.000 min
Response: 127765050
Conc: 532.31 ng/ml



#28 AR-1254-3


R.T.: 6.175 min
Delta R.T.: 0.001 min
Response: 387538706
Conc: 538.15 ng/ml

Instrument :
ECD_O
ClientSampleId :




#28 AR-1254-3

R.T.: 6.153 min
Delta R.T.: 0.000 min
Response: 202966969
Conc: 540.51 ng/ml




#29 AR-1254-4

R.T.: 6.405 min
Delta R.T.: 0.001 min
Response: 237282578
Conc: 540.74 ng/ml



#29 AR-1254-4


R.T.: 6.381 min
Delta R.T.: 0.000 min
Response: 117839788
Conc: 541.35 ng/ml



#30 AR-1254-5

R.T.: 6.827 min
Delta R.T.: 0.000 min
Response: 339699576
Conc: 534.86 ng/ml

Instrument :
ECD_O
ClientSampleId :



#30 AR-1254-5

R.T.: 6.800 min
Delta R.T.: 0.000 min
Response: 174756076
Conc: 541.44 ng/ml