

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120722\
 Data File : PO091139.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Dec 2022 17:34
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 08 00:11:15 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Nov 28 11:25:59 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.299	3.476	168.6E6	58594380	49.877	55.828
2) SA Decachlor...	10.046	8.478	119.3E6	51555685	44.913	53.102
Target Compounds						
21) L5 AR-1248-1	5.475	4.561	33716801	12169005	473.330	517.262
22) L5 AR-1248-2	5.751	4.799	46154372	17156896	435.381	524.036
23) L5 AR-1248-3	5.956	4.840	52363174	17793512	444.957	524.266
24) L5 AR-1248-4	6.362	5.011	54795430	20865609	448.085	522.917
25) L5 AR-1248-5	6.400	5.405	53549946	19107805	466.884	533.067



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120722\
Data File : PO091139.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Dec 2022 17:34
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 08 00:11:15 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO112522.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Nov 28 11:25:59 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.299 min
Delta R.T.: 0.000 min
Response: 168640156
Conc: 49.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.476 min
Delta R.T.: 0.000 min
Response: 58594380
Conc: 55.83 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.046 min
Delta R.T.: -0.004 min
Response: 119297287
Conc: 44.91 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.478 min
Delta R.T.: -0.004 min
Response: 51555685
Conc: 53.10 ng/ml



#21 AR-1248-1


R.T.: 5.475 min
Delta R.T.: 0.000 min
Response: 33716801
Conc: 473.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.561 min
Delta R.T.: 0.000 min
Response: 12169005
Conc: 517.26 ng/ml




#22 AR-1248-2

R.T.: 5.751 min
Delta R.T.: 0.000 min
Response: 46154372
Conc: 435.38 ng/ml



#22 AR-1248-2


R.T.: 4.799 min
Delta R.T.: 0.000 min
Response: 17156896
Conc: 524.04 ng/ml



#23 AR-1248-3


R.T.: 5.956 min
Delta R.T.: -0.001 min
Response: 52363174
Conc: 444.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.840 min
Delta R.T.: 0.000 min
Response: 17793512
Conc: 524.27 ng/ml




#24 AR-1248-4

R.T.: 6.362 min
Delta R.T.: 0.000 min
Response: 54795430
Conc: 448.09 ng/ml



#24 AR-1248-4


R.T.: 5.011 min
Delta R.T.: 0.000 min
Response: 20865609
Conc: 522.92 ng/ml



#25 AR-1248-5

R.T.: 6.400 min
Delta R.T.: -0.002 min
Response: 53549946
Conc: 466.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.405 min
Delta R.T.: -0.001 min
Response: 19107805
Conc: 533.07 ng/ml